BS ISO 20637:2015

BSI Standards Publication

Infant formula and adult nutritionals — Determination of myo-inositol by liquid chromatography and pulsed amperometry

bsi.

...making excellence a habit.™

Document Sharing Hub

Full document details: BS ISO 20637:2015

National foreword

This British Standard is the UK implementation of ISO 20637:2015.

The UK participation in its preparation was entrusted to Technical Committee AW/-/2, Food Technical Committee Chairmen.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2015.
Published by BSI Standards Limited 2015

ISBN 978 0 580 90420 2

ICS 67.050

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 November 2015.

Amendments/corrigenda issued since publication

Date	Text affected

Documents
Document Sharing Hub

Full document details: BS ISO 20637:2015

INTERNATIONAL STANDARD

ISO 20637

First edition
2015-11-01

Infant formula and adult nutritionals — Determination of myo-inositol by liquid chromatography and pulsed amperometry

Formules infantiles et produits nutritionnels pour adultes — Détermination de la teneur en myo-inositol par chromatographie liquide et ampérométrie pulsée

Reference number
ISO 20637:2015(E)

Documents
Document Sharing Hub

© ISO 2015

Full document details: BS ISO 20637:2015

COPYRIGHT PROTECTED DOCUMENT

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

© ISO 2015 – All rights reserved

Documents
Document Sharing Hub

Full document details: BS ISO 20637:2015

Contents

Page

Foreword ... **iv**

1 Scope ... **1**

2 Terms and definitions ... **1**

3 Principle ... **1**

4 Reagents and materials ... **1**

5 Apparatus ... **3**

6 Procedure ... **4**
- 6.1 Free myo-inositol ... 4
  - 6.1.1 Sample preparation ... 4
  - 6.1.2 Extraction ... 5
- 6.2 Myo-inositol bound as phosphatidylinositol ... 5
  - 6.2.1 Sample preparation ... 5
  - 6.2.2 Extraction ... 5
  - 6.2.3 Cleanup ... 6
  - 6.2.4 Hydrolysis ... 6
- 6.3 HPLC analysis ... 6
  - 6.3.1 Instrument operating conditions ... 6
  - 6.3.2 PAD settings with gold electrode ... 8
  - 6.3.3 Instrument startup ... 8
  - 6.3.4 Standard and sample analysis ... 8
  - 6.3.5 System shutdown ... 8

7 Calculations ... **8**
- 7.1 General ... 8
- 7.2 Concentration of calibration standards ... 9
- 7.3 Preparation of standard curve ... 9
- 7.4 Calculation of free or free plus bound myo-inositol in samples ... 9
  - 7.4.1 Calculation of free myo-inositol ... 9
  - 7.4.2 Calculation of bound myo-inositol ... 10
  - 7.4.3 Calculation of free plus bound myo-inositol ... 10

Annex A (informative) **Examples of chromatograms** ... **11**

Annex B (informative) **Precision data** ... **12**

Bibliography ... **15**

© ISO 2015 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 20637:2015

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2. www.iso.org/directives

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received. www.iso.org/patents

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

The committee responsible for this document is ISO/TC 34, *Food products* in collaboration with AOAC INTERNATIONAL. It is being published by ISO and separately by AOAC INTERNATIONAL. The method described in this International Standard is equivalent to the AOAC Official Method 2011.18: *Myo-inositol (free and bound as phosphatidyl inositol) in infant and pediatric formula and adult nutritional*.

© ISO 2015 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS ISO 20637:2015

INTERNATIONAL STANDARD ISO 20637:2015(E)

Infant formula and adult nutritionals — Determination of myo-inositol by liquid chromatography and pulsed amperometry

WARNING — The use of this International Standard can involve hazardous materials, operations and equipment. This International Standard does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this International Standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

1 Scope

This International Standard specifies a method for the determination of myo-inositol (free or free plus bound as phosphatidylinositol) in infant formula and adult nutritionals using liquid chromatography and pulsed amperometry with column switching.

2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

2.1
adult nutritional
nutritionally complete, specially formulated food, consumed in liquid form, which may constitute the sole source of nourishment, made from any combination of milk, soy, rice, whey, hydrolysed protein, starch and amino acids, with and without intact protein

2.2
infant formula
breast-milk substitute specially manufactured to satisfy, by itself, the nutritional requirements of infants during the first months of life up to the introduction of appropriate complementary feeding

[SOURCE: Codex Standard 72-1981]

3 Principle

Free myo-inositol and phosphatidyl bound myo-inositol are extracted using two different sample preparation procedures. Free myo-inositol is extracted from samples with dilute hydrochloric acid and water. Phosphatidylinositol is extracted from samples with chloroform and separated from other fats with silica solid phase extraction cartridges. Myo-inositol is then released from the glycerol backbone with concentrated acetic and hydrochloric acid at 120°C. The ion chromatographic method uses a combination of two different ion exchange columns with column switching and pulsed amperometric detection (PAD). The concentration of myo-inositol is calculated by comparison with external standards of known concentration.

4 Reagents and materials

During the analysis, unless otherwise stated, use only reagents of recognized analytical grade and distilled or demineralized water or water of equivalent purity.

© ISO 2015 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 20637:2015

4.1 Chemicals and solvents

4.1.1 Acetic acid, glacial, ACS.

4.1.2 Chloroform, high-purity, HPLC grade.

4.1.3 Diethyl ether, anhydrous, HPLC grade.

4.1.4 Drierite, (desiccant), anhydrous calcium sulfate, 8 mesh.

4.1.5 Helium, zero grade or equivalent.

4.1.6 Hexane, HPLC grade.

4.1.7 Hydrochloric acid, concentrated (36 % to 38 %), ACS.

4.1.8 Metaphosphoric acid, ACS.

4.1.9 Methanol, HPLC grade.

4.1.10 Myo-inositol, primary reference standard, official lot, store desiccated. See standard label for purity.

4.1.11 Sodium chloride, ACS.

4.1.12 Sodium hydroxide, 50 % (m/m), low carbonate form.

4.2 Preparation of reagents and standard solutions

4.2.1 General. All solutions can be scaled up or down for convenience provided good laboratory practices are observed. Solutions can be stored at refrigerated or at ambient temperature in tight, inert containers unless otherwise specified.

4.2.2 Myo-inositol stock standard solution (approximately 2 000 mg/l). Accurately weigh approximately 0,100 g myo-inositol and quantitatively transfer to a 50 ml volumetric flask. Dilute to volume with water. Mix well. Store refrigerated. Expiration: 3 months.

4.2.3 Myo-inositol intermediate standard solution (approximately 200 mg/l). Dilute 10,0 ml stock standard (4.2.2) to 100 ml with water and mix well. Discard after use.

4.2.4 Preparation of calibration standard solutions

4.2.4.1 Myo-inositol calibration standard solutions high, (approximately 4 mg/l, 2 mg/l, 1 mg/l, 0,5 mg/l).

Into separate volumetric flasks, dilute 2,0 ml, 1,0 ml and 0,5 ml myo-inositol intermediate standard (4.2.3) to 100 ml with water. Dilute 0,5 ml myo-inositol intermediate standard (4.2.3) to 200 ml with water. Expiration: 2 weeks

4.2.4.2 Myo-inositol calibration standard solutions low, (approximately 0,2 mg/l and 0,05 mg/l).

Into separate volumetric flasks, dilute 4 ml and 1 ml of the 0,5 mg/l myo-inositol calibration standard to 10 ml with water. Expiration: 2 weeks.

© ISO 2015 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS ISO 20637:2015