BS ISO 24166-1:2022

BSI Standards Publication

Snap-on bottles for metering pumps

Part 1: Tubular glass

Documents
Document Sharing Hub

Full document details: BS ISO 24166-1:2022

National foreword

This British Standard is the UK implementation of ISO 24166-1:2022.

The UK participation in its preparation was entrusted to Technical Committee CH/212, IVDs.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2022
Published by BSI Standards Limited 2022

ISBN 978 0 539 05154 4

ICS 11.120.99

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 August 2022.

Amendments/corrigenda issued since publication

Date	Text affected

Documents
Document Sharing Hub

Full document details: BS ISO 24166-1:2022

INTERNATIONAL STANDARD

ISO 24166-1

First edition
2022-08

Snap-on bottles for metering pumps —

Part 1: Tubular glass

Flacons encliquetables pour pompes doseuses —

Partie 1: Verre étiré

Reference number
ISO 24166-1:2022(E)

Documents
Document Sharing Hub

© ISO 2022

Full document details: BS ISO 24166-1:2022

COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub

Full document details: BS ISO 24166-1:2022

Contents

Page

Foreword ..... **iv**
Introduction ..... **v**
1 Scope **1**
2 Normative references **1**
3 Terms and definitions **1**
4 Symbols **3**
5 Dimensions **4**
5.1 General 4
5.2 Vials without sealing ring 4
5.3 Vials with sealing ring 7
6 Materials **10**
7 Requirements **10**
7.1 General requirements 10
7.2 Physical requirements 10
7.2.1 Vertical force 10
7.2.2 Particulate contaminations 10
7.3 Chemical requirements 11
7.4 Sterilization 11
8 Marking of the vial **11**
9 Packaging options **11**
9.1 General packaging requirements 11
9.2 Packaging for nonsterile glass vials 11
9.3 Ready to sterilize or sterile glass vials 11
Annex A (informative) **Snap-on considerations** **13**
Annex B (informative) **Tightness test** **15**
Annex C (informative) **Inner depth measuring** **16**
Annex D (normative) **Vertical force measurement test method** **17**
Annex E (informative) **Example of a packaging configuration** **19**
Bibliography **20**

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 24166-1:2022

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use*.

A list of all parts in the ISO 24166 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS ISO 24166-1:2022

Introduction

This document is part of a series of International Standards for containers made from different materials used in combination with metering pumps for medicinal applications.

Glass containers are mainly used for that purpose. Plastic containers can be used as an alternative.

This document can be used for the development of standardized filling and assembling equipment.

Based on the dimensions of the containers, appropriate components, such as metering pumps and other closure systems can also be standardized. As such, this document provides important inputs for developing entire packaging systems for medicinal applications.

Primary packaging materials are an integral part of medicinal products. Thus, depending on the jurisdiction, the principles of the current Good Manufacturing Practices (cGMP) can apply to the manufacturing of these components (e.g. ISO 15378).

The containers referred to as bottles in the title of this document are usually referred to as “vials” if they are made of tubular glass. For this reason, the text of this document refers to “vials” rather than “bottles” in order to differentiate them from containers made of moulded glass and plastic.

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 24166-1:2022

Documents
Document Sharing Hub

Snap-on bottles for metering pumps —

Part 1:
Tubular glass

1 Scope

This document specifies the shape, dimensions, fill capacities and performance requirements of tubular glass vials for metering pumps. It also specifies the material for the manufacturing of such containers as well as the secondary packaging.

This document provides also requirements for packaging of the tubular glass vials and addresses nonsterile, ready to sterilize or sterile as three possible options.

This document is applicable to colourless or amber containers made of tubular glass and intended to be used for packaging, storage or transportation of products intended for medicinal use.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 719, *Glass — Hydrolytic resistance of glass grains at 98 °C — Method of test and classification*

ISO 720, *Glass — Hydrolytic resistance of glass grains at 121 °C — Method of test and classification*

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes*

ISO 4802-1, *Glassware — Hydrolytic resistance of the interior surfaces of glass containers — Part 1: Determination by titration method and classification*

ISO 4802-2, *Glassware — Hydrolytic resistance of the interior surfaces of glass containers — Part 2: Determination by flame spectrometry and classification*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at https://www.iso.org/obp

— IEC Electropedia: available at https://www.electropedia.org/

3.1
base foil
packaging component for protection of the pallet load from contamination

© ISO 2022 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 24166-1:2022

3.2
bottle
vial
container made of a suitable material, which is intended for packaging, storage and application of liquid medicinal products

EXAMPLE Tubular glass, moulded glass or plastic.

3.3
brimful capacity
volume of water required to fill a container, placed on a flat, horizontal surface

[SOURCE: ISO 4802-1:2016, 3.3]

3.4
chemical indicator
test system that reveals change in one or more pre-defined process variables based on a chemical change resulting from exposure to a process

[SOURCE: ISO 11139:2018, 3.43, modified — "pre-specified" was replaced with "pre-defined", "or physical" was deleted.]

3.5
customer
business entity that purchases bottles for metering pumps and conducts further processing or filling as appropriate

[SOURCE: ISO 21882:2019, 3.1 modified — “sterilized ready for filling vials” was replaced with “bottles for metering pumps”.]

3.6
dip-tube
conduit, part of the metering pump, which delivers the product from the vial to the metering mechanism

3.7
foil bag
gas permeable or non-permeable bag

Note 1 to entry: Foil bags can be used as *sterile barrier system* (3.16) or as *protective packaging* (3.13), depending on final usage.

3.8
label
written, printed or graphic information appearing on the item itself, on the packaging of each item or on the packaging of multiple items

[SOURCE: IMDRF/GRRP WG/N52:2019, 3.17, modified – replaced "unit" and "devices" with "item".]

3.9
manufacturer
business entity that manufactures or is otherwise responsible for the manufacturing of the bottles ready to be filled by the *customer* (3.5)

[SOURCE: ISO 21882:2019, 3.4, modified — “performs” was replaced with “manufactures”, “vials” was replaced with “bottles”.]

3.10
metering pump
device actuated by the user to deliver a specific dose of liquid

© ISO 2022 – All rights reserved
Documents
Document Sharing Hub
Full document details: BS ISO 24166-1:2022