

Petroleum and liquid petroleum products — Measurement of level and temperature in storage tanks by automatic methods —

**Part 3: Measurement of level in
pressurized storage tanks
(non-refrigerated)**

ICS 75.180.30

National foreword

This British Standard reproduces verbatim ISO 4266-3:2002 and implements it as the UK national standard. Together with BS ISO 4266:2002 Parts 1 and 2 and Parts 4 to 6, it supersedes BS 7812:1995 which is withdrawn.

The UK participation in its preparation was entrusted by Technical Committee PTI/12, Petroleum measurement and sampling, to Subcommittee PTI/12/1, Static and dynamic petroleum measurement, which has the responsibility to:

- aid enquirers to understand the text;
- present to the responsible international/European committee any enquiries on the interpretation, or proposals for change, and keep the UK interests informed;
- monitor related international and European developments and promulgate them in the UK.

A list of organizations represented on this subcommittee can be obtained on request to its secretary.

Cross-references

The British Standards which implement international publications referred to in this document may be found in the *BSI Catalogue* under the section entitled “International Standards Correspondence Index”, or by using the “Search” facility of the *BSI Electronic Catalogue* or of British Standards Online.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

This British Standard, having been prepared under the direction of the Materials and Chemicals Sector Policy and Strategy Committee, was published under the authority of the Standards Policy and Strategy Committee on 9 December 2002

Summary of pages

This document comprises a front cover, an inside front cover, the ISO title page, pages ii to v, a blank page, pages 1 to 15 and a back cover.

The BSI copyright date displayed in this document indicates when the document was last issued.

Amendments issued since publication

Amd. No.	Date	Comments

© BSI 9 December 2002

ISBN 0 580 40919 8

INTERNATIONAL STANDARD

ISO
4266-3

First edition
2002-11-15

Petroleum and liquid petroleum products — Measurement of level and temperature in storage tanks by automatic methods —

Part 3: Measurement of level in pressurized storage tanks (non-refrigerated)

*Pétrole et produits pétroliers liquides — Mesurage du niveau et de la
température dans les réservoirs de stockage par méthodes automatiques —*

*Partie 3: Mesurage du niveau dans les réservoirs de stockage sous
pression (non réfrigérés)*



Reference number
ISO 4266-3:2002(E)

Contents

	Page
1 Scope	1
2 Normative reference	1
3 Terms and definitions	1
4 Precautions	3
4.1 Safety precautions	3
4.2 Equipment precautions	3
4.3 General precautions	3
5 Accuracy	4
5.1 Intrinsic error of ALGs	4
5.2 Calibration prior to installation	4
5.3 Error caused by installation and operating conditions	4
5.4 Overall accuracy	4
5.4.1 General	4
5.4.2 Use of ALGs for fiscal/custody transfer applications	5
6 Installation of ALGs	5
6.1 General	5
6.2 Mounting location	5
6.3 Manufacturer's requirements	5
6.4 Installation	5
6.4.1 Installation of intrusive ullage ALGs (e.g. float-operated and servo-operated type) mounted on still-wells	5
6.4.2 Installation of intrusive ullage ALGs (e.g. float and servo-operated types) using guide wires	5
6.4.3 Installation of non-intrusive ullage ALGs (e.g. microwave or radar) mounted on still-wells	9
6.4.4 Location of ALG	9
6.4.5 Installation of ALGs other than those described in this part of ISO 4266	9
6.5 Still-well design	9
7 Initial setting and initial verification of ALGs in the field	10
7.1 Preparation	10
7.1.1 Check for critical distances	10
7.1.2 Check for free movement of level-sensing element of intrusive ALGs	10
7.1.3 Check for influences of changes in physical and electrical properties of the product (vapour/liquid)	10
7.2 Initial setting	10
7.2.1 General	10
7.2.2 Setting of intrusive ullage ALGs	11
7.2.3 Setting of non-intrusive ALGs	11
7.2.4 Other ALGs not described in this part of ISO 4266	12
7.3 Initial field verification	12
7.3.1 Introduction	12
7.3.2 Verification procedure	12

7.3.3 Tolerance of initial field verification **13**

7.4 Record keeping **13**

8 Subsequent verification of ALGs **13**

8.1 General **13**

8.2 Frequency of subsequent verification **13**

8.3 Procedure **13**

8.4 Tolerance for subsequent verification of ALGs in fiscal/custody transfer application **13**

8.5 Comparison of the ALG readings from current and previous verifications **14**

8.6 Adjustment in subsequent verification **14**

9 Data communication and receiving **14**

Bibliography..... **15**

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 4266 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 4266-3 was prepared by Technical Committee ISO/TC 28, *Petroleum products and lubricants*, Subcommittee SC 3, *Static petroleum measurement*.

ISO 4266-3, together with ISO 4266-1, ISO 4266-2 and ISO 4266-4 to ISO 4266-6, cancels and replaces ISO 4266:1994, which has been technically revised.

ISO 4266 consists of the following parts, under the general title *Petroleum and liquid petroleum products — Measurement of level and temperature in storage tanks by automatic methods*:

- *Part 1: Measurement of level in atmospheric tanks*
- *Part 2: Measurement of level in marine vessels*
- *Part 3: Measurement of level in pressurized storage tanks (non-refrigerated)*
- *Part 4: Measurement of temperature in atmospheric tanks*
- *Part 5: Measurement of temperature in marine vessels*
- *Part 6: Measurement of temperature in pressurized storage tanks (non-refrigerated)*

