BS ISO 4717:2024

BSI Standards Publication

Water quality — Protactinium 231 — Test method using ICP-MS

Documents
Document Sharing Hub

Full document details: BS ISO 4717:2024

National foreword

This British Standard is the UK implementation of ISO 4717:2024.

The UK participation in its preparation was entrusted to Technical Committee EH/3/8, Radioactivity measurements methods.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2024
Published by BSI Standards Limited 2024

ISBN 978 0 539 17243 0

ICS 13.060.60; 17.240

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 December 2024.

Amendments/corrigenda issued since publication

Date	Text affected

Documents
Document Sharing Hub

Full document details: BS ISO 4717:2024

International Standard

ISO 4717

Water quality — Protactinium 231 — Test method using ICP-MS

Qualité de l'eau — Protactinium 231 — Méthode d'essai par ICP-MS

First edition
2024-11

Reference number
ISO 4717:2024(en)

© ISO 2024

Documents
Document Sharing Hub

Full document details: BS ISO 4717:2024

COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Documents
Document Sharing Hub
Full document details: BS ISO 4717:2024

Contents

Page

Foreword ... **iv**
Introduction ... **v**
1 Scope ... **1**
2 Normative references ... **1**
3 Terms and definitions ... **2**
4 Symbols ... **2**
5 Principle ... **3**
6 Sampling and sample storage ... **4**
7 Chemical reagents and apparatus ... **5**
7.1 General ... 5
7.2 Chemical reagents ... 5
7.3 Apparatus ... 5
8 Separation ... **6**
9 Quality control ... **6**
9.1 General ... 6
9.2 Variables that can influence the measurement ... 6
9.3 Instrument verification ... 6
9.4 Method verification ... 7
10 Expression of results ... **7**
10.1 Data analysis ... 7
10.2 Background ... 8
10.3 Internal standard ... 8
10.4 Expression of results using ^{233}Pa as a recovery tracer ... 8
10.4.1 Calculation of activity of the tracer and mass of the analyte ... 8
10.4.2 Chemical recovery ... 9
10.4.3 Measurement bias ... 9
10.4.4 Sample mass concentration ... 10
10.5 Limit of detection ... 10
10.6 Limit of quantification ... 10
10.7 Correcting for ^{231}Pa contamination in the tracer ... 10
10.8 Conversion of mass concentration to mass activity ... 11
10.9 Conversion from mass to volume units ... 11
11 Test report ... **11**
Annex A (informative) **Chemical separation of protactinium by extraction chromatography** ... **13**
Annex B (informative) **Anion exchange resin method** ... **15**
Bibliography ... **17**

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 4717:2024

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 3, *Radioactivity measurements.*

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 4717:2024

Introduction

Radionuclides are present throughout the environment; thus, water bodies (e.g. surface waters, ground waters and sea waters) contain radionuclides, which can be of either natural or anthropogenic origin.

— Naturally-occurring radionuclides, including ^{3}H, ^{14}C, ^{40}K and those originating from the thorium and uranium decay series, in particular ^{210}Pb, ^{210}Po, ^{222}Rn, ^{226}Ra, ^{228}Ra, ^{227}Ac, ^{231}Pa, ^{234}U and ^{238}U, can be found in water bodies due to either natural processes (e.g. desorption from the soil and runoff by rain water) or released from technological processes involving naturally occurring radioactive materials (e.g. mining, mineral processing, oil, gas, and coal production, water treatment, and production and use of phosphate fertilisers).

— Anthropogenic radionuclides, such as ^{55}Fe, ^{59}Ni, ^{63}Ni, ^{90}Sr and ^{99}Tc, transuranic elements (e.g. Np, Pu, Am and Cm), and some gamma emitting radionuclides, such as ^{60}Co and ^{137}Cs, can also be found in natural waters. Small quantities of anthropogenic radionuclides can be discharged from nuclear facilities to the environment as a result of authorized routine releases. The radionuclides present in liquid effluents are usually controlled before being discharged into the environment[1] and water bodies. Anthropogenic radionuclides used in medical and industrial applications can be released to the environment after use. Anthropogenic radionuclides are also found in waters due to contamination from fallout resulting from above-ground nuclear detonations and accidents such as those that have occurred at the Chernobyl and Fukushima nuclear facilities.

Radionuclide activity concentrations in water bodies can vary according to local geological characteristics and climatic conditions and can be locally and temporally enhanced by releases from nuclear facilities during planned, existing and emergency exposure situations[2],[3]. Some drinking water sources can thus contain radionuclides at activity concentrations that can present a human health risk. The World Health Organization (WHO) recommends to routinely monitor radioactivity in drinking waters[4] and to take proper actions when needed to minimize the health risk.

National regulations usually specify the activity concentration limits that are authorized in drinking waters, water bodies and liquid effluents to be discharged to the environment. These limits can vary for planned, existing and emergency exposure situations. As an example, during either a planned or existing situation, the WHO guidance level for ^{231}Pa in drinking water is 0,1 $Bq \cdot l^{-1}$, see NOTES 1 and 2. Compliance with these limits is assessed by measuring radioactivity in water samples and by comparing the results obtained, with their associated uncertainties, as specified by ISO/IEC Guide 98-3 and ISO 5667-20[5].

NOTE 1 If the value is not specified in Annex 6 of Reference [4], the value has been calculated using the formula provided in Reference [4] and the dose coefficient data from References [6] and [7].

NOTE 2 The guidance level calculated in Reference [4] is the activity concentration that, with an intake of 2 $l \cdot d^{-1}$ of drinking water for one year, results in an effective dose of 0,1 $mSv \cdot a^{-1}$ to members of the public. This is an effective dose that represents a very low level of risk to human health and which is not expected to give rise to any detectable adverse health effects[4].

This document contains method(s) to support laboratories, which need to determine ^{231}Pa in water samples.

The method described in this document can be used for various types of waters (see Clause 1). For radiometric methods, minor modifications such as sample volume and counting time can be made if needed to ensure that the decision threshold, limit of detection, and uncertainties are below the required limits. For ICP-MS methods, minor modifications to, for example, the sample pre-concentration volume and the interference separation can be made if needed to ensure that the limit of detection, limit of quantification and uncertainties are below the required limits. This can be done for several reasons such as emergency situations, lower national guidance limits and operational requirements.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 4717:2024

Documents
Document Sharing Hub

Water quality — Protactinium 231 — Test method using ICP-MS

WARNING — Persons using this document should be familiar with normal laboratory practices. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to determine the applicability of any other restrictions.

IMPORTANT — It is essential that tests conducted according to this document be carried out by suitably trained staff.

1 Scope

This document specifies a method to determine ^{231}Pa by inductively coupled plasma mass spectrometry (ICP-MS). The mass concentrations obtained can be converted into activity concentrations.

The method described in this document is applicable to test samples of drinking water, rainwater, surface and ground water, marine water, as well as cooling water, industrial water, domestic and industrial wastewater after proper sampling and handling and test sample preparation.

The limit of detection depends on the sample volume, the instrument used, the background count rate, the detection efficiency and the chemical yield. In this document, the limit of detection of the method using currently available apparatus is approximately 0,1 Bq·l^{-1} (or Bq·kg^{-1}), which is the same as the WHO criteria for safe consumption of drinking water (0,1 Bq·l^{-1})[4].

The method described in this document covers the measurement of ^{231}Pa in water at activity concentrations between 0,1 Bq·l^{-1} and 100 Bq·l^{-1}. Samples with higher activity concentrations than 100 Bq·l^{-1} can be measured if a dilution is performed.

The method described in this document is applicable in the event of an emergency.

Filtration of the test sample is necessary for the method described in this document. The analysis of ^{231}Pa adsorbed to suspended matter is not covered by this method. The analysis of the insoluble fraction requires a mineralization step that is not covered by this document. In this case, the measurement is made on the different phases obtained.

It is the user's responsibility to ensure the validity of this test method for the water samples tested.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Preservation and handling of water samples*

ISO 5667-10, *Water quality — Sampling — Part 10: Guidance on sampling of waste water*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 4717:2024