BS ISO 7383-2:2024

BSI Standards Publication

Fine bubble technology — Evaluation method for determining gas content in fine bubble dispersions in water

Part 2: Hydrogen content

Documents
Document Sharing Hub

Full document details: BS ISO 7383-2:2024

National foreword

This British Standard is the UK implementation of ISO 7383-2:2024.

The UK participation in its preparation was entrusted to Technical Committee LBI/50, Fine Bubble Technology (FBT).

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2024
Published by BSI Standards Limited 2024

ISBN 978 0 539 29005 9

ICS 07.030

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 July 2024.

Amendments/corrigenda issued since publication

Date	Text affected

Documents
Document Sharing Hub

Full document details: BS ISO 7383-2:2024

International Standard

ISO 7383-2

First edition
2024-07

Fine bubble technology — Evaluation method for determining gas content in fine bubble dispersions in water —

Part 2: Hydrogen content

Reference number
ISO 7383-2:2024(en)

© ISO 2024

Documents
Document Sharing Hub

Full document details: BS ISO 7383-2:2024

COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Documents
Document Sharing Hub
Full document details: BS ISO 7383-2:2024

Contents

Page

Foreword ... **v**
Introduction ... **vi**
1 Scope ... **1**
2 Normative references ... **1**
3 Terms and definitions ... **1**
4 Principle and application ... **2**
4.1 General ... 2
4.2 Titration (oxidimetry) method ... 2
4.3 Gas chromatography ... 2
5 Apparatus and materials ... **2**
5.1 Titration (oxidimetry) method ... 2
5.1.1 Reagent ... 2
5.1.2 Dropper or titrator ... 3
5.1.3 Clear vial ... 3
5.2 Gas chromatography ... 3
5.2.1 Vial and clamping machine ... 3
5.2.2 UFBs elimination ... 3
5.2.3 Carrier gases and column ... 3
5.2.4 Gas chromatograph ... 3
5.2.5 Measurement device for UFB size and concentration ... 4
6 Procedure ... **4**
6.1 Titration (oxidimetry) method ... 4
6.2 Gas chromatography ... 4
6.2.1 General ... 4
6.2.2 Sample sealing ... 5
6.2.3 UFBs elimination ... 6
6.2.4 Analysis of gas substitutes ... 6
7 Results and calculation ... **8**
7.1 Titration (oxidimetry) method ... 8
7.2 Gas chromatography ... 8
7.2.1 General ... 8
7.2.2 Calibration curve for hydrogen concentration measurement ... 8
7.2.3 Volume of the water sample in the vial ... 9
7.2.4 Volume of the headspace above the UFB dispersion in the vial ... 9
7.2.5 Pressure of the gas in the vial headspace ... 10
7.2.6 Hydrogen content in the vial head space ... 10
7.2.7 Hydrogen content in the vial water sample ... 10
7.2.8 Hydrogen content in association with each number concentration of UFBs and dissolved hydrogen ... 11
8 Error of measurement ... **12**
8.1 Titration (oxidimetry) method ... 12
8.2 Gas chromatography ... 12
8.3 Relative standard deviation of the gas chromatograph ... 13
9 Test report ... **13**
9.1 Testing results ... 13
9.2 Testing conditions ... 13
Annex A (informative) **Example of test result at testing laboratory by titration (oxidimetry) method** ... **14**
Annex B (informative) **Example of test result at testing laboratory by gas chromatography using the capillary column** ... **17**

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 7383-2:2024

Annex C (informative) **Example of test result at testing laboratory by gas chromatography using the packed column** **20**

Annex D (informative) **Comparison of different vials when testing the hydrogen content by gas chromatography** **23**

Bibliography **25**

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub

Full document details: BS ISO 7383-2:2024

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 281, *Fine bubble technology*.

A list of all parts in the ISO 7383 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 7383-2:2024

Introduction

Hydrogen has recently attracted considerable research attention across several fields, such as agriculture, food, environment, and medicine sciences, owing to its antioxidant capacity and zero-residual and ecofriendly properties. Hydrogen concentration is a crucial parameter to ensure its successful hydrogen application. At present, there are potential limitations regarding the application of hydrogen in living organisms, such as accessibility, availability, and biological intake. For example, hydrogen exhibits low solubility, with a saturation concentration of only 1,6 mg/l. Fine bubble and ultrafine bubble (UFB) technology have been used to address the low solubility of hydrogen in recent years. UFBs have a large surface area, high internal pressure, and negatively charged surfaces, which accelerate the dissolution of gas into liquids and maintain their stability in the liquids for relatively long periods. Therefore, UFBs can potentially expand the application of hydrogen molecules because of their unique characteristics.

An accurate hydrogen concentration measurement is required to develop an impact evaluation standard for the variety of hydrogen production equipment and hydrogen-rich water products. Normal techniques for measuring the hydrogen content in water include the electrochemical probe (membrane-type polarography) method, titration (oxidimetry) method using a methylene blue (MB)–platinum colloid reagent, and gas chromatography.

The electrochemical probe (membrane-type polarography) method is advantageous as it can be used for both in situ and real-time measurements. Detection limits are stated in the instruction manuals of instruments and, in most cases, are approximately 0,01 mg/l. The upper limit depends on the specifications of the instrument, but it is generally lower than 20 mg/l. Electrochemical probe methods can measure only the dissolved hydrogen concentration in hydrogen UFB dispersions under a steady state or flow state and are not suitable for static water samples. Additionally, the presence of UFBs in water can influence measurement results.

The titration (oxidimetry) method is straightforward and economical; however, its lower limit of detection is only 0,1 mg/l, and the titration range with acceptable accuracy is narrow. Moreover, hydrogen evaporation during measurements and the presence of UFBs in water can influence measurements. Thus, this method can be used only to approximate hydrogen concentrations in UFB dispersions.

Gas chromatography is the most accurate method for measuring the hydrogen content of gas. Its lower limit of detection is the lowest among all methods, and there are no upper measurement limits. However, few attempts have been made to apply this method to measuring hydrogen contents in water. Furthermore, appropriate sample preparation and UFB elimination methods have yet to be developed.

Therefore, a standard method for measuring total hydrogen contents in UFB dispersions has been established. For this document, the titration (oxidimetry) method is proposed as a rapid estimation method and gas chromatography is proposed as an accurate measurement method. The following procedure was used for gas chromatography. First, a bubble-elimination pretreatment is performed to drive both dissolved hydrogen and hydrogen in UFBs from the liquid phase into the headspace, followed by gas chromatography combined with theoretical calculations to determine the total hydrogen content in UFB dispersions. The establishment of this standard method will serve various hydrogen production enterprises, the corresponding customers, and research institutions, enabling them to have a common measurement standard when determining hydrogen contents, thereby facilitating comparisons and judgments of the quality and function of hydrogen products.

The standardized evaluation method for hydrogen content provides an important theoretical basis for future applications in several fields, with further potential for industry, commerce, government, consumers, and academic and research bodies. The establishment of standards will enable governments to establish policies, regulate the development of the hydrogen health industry, and promote research and development into products and technologies related to hydrogen agriculture, hydrogen medicine, and hydrogen-based environmental applications.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 7383-2:2024

Fine bubble technology — Evaluation method for determining gas content in fine bubble dispersions in water —

Part 2:
Hydrogen content

1 Scope

This document specifies the evaluation methods for hydrogen content in ultrafine bubble (UFB) dispersions.

The titration (oxidimetry) method can be used as a quick method to estimate the hydrogen content in hydrogen UFB dispersions. The lower limit of detection is 0,1 mg/l, and the range with acceptable accuracy is between 0,2 mg/l and 1,6 mg/l. The existence of oxidizing or reducing substances in dispersions influences measurement accuracy.

The gas chromatographic method features a considerably high accuracy range and lower limit of detection. The existence of UFBs in water does not influence the measurement results. The existence of oxidizing or reducing substances in water does not affect the measurement accuracy either. However, the measurement procedure is time consuming.

NOTE This document only provides a method for determining hydrogen contents in UFB dispersions and does not involve the specific effects of hydrogen UFB dispersion application.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at https://www.iso.org/obp

— IEC Electropedia: available at https://www.electropedia.org/

3.1
dissolved hydrogen
DH
hydrogen molecules, dissolved in a liquid

3.2
hydrogen UFB
hydrogen molecules, dissolved as ultrafine bubble (UFB)

3.3
titration
method or process of determining the concentration of a dissolved substance in terms of the smallest amount of a reagent of known concentration required to bring about a given effect in reaction with a known volume of the test solution

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 7383-2:2024

4 Principle and application

4.1 General

The following two methods can be employed for hydrogen content measurements in UFB dispersions, which are generated by cleaned UFB generation systems utilizing pure water, saline, or buffer solutions, and the gas in which can be pure hydrogen or a gaseous mixture containing hydrogen.

The titration (oxidimetry) method is a quick, economical method with a lower limit of detection of 0,1 mg/l; it is recommended for use only as a rough estimation method.

NOTE The existence of oxidizing or reducing substances in the dispersions influences measurement accuracy.

Gas chromatography is a method with higher accuracy, but it is more expensive and time consuming; it is recommended to be used for more accurate hydrogen content measurements in UFB dispersions and as the reference for the validation titration method.

4.2 Titration (oxidimetry) method

The titration (oxidimetry) method involves a redox reaction of a MB oxidant in the presence of a colloidal platinum catalyst. MB generally reacts with an equimolar amount of hydrogen with platinum or palladium to produce colorless, reduced MB (leucomethylene blue, leucoMB), as shown below:

$$\text{MB (blue)} + 2H^+ + 2e^- \rightarrow \text{leucoMB (colorless)}.$$

The oxidimetry determination of the hydrogen concentration was performed by redox titration. The MB–platinum reagent was added dropwise to a 6 ml sample of a hydrogen UFB dispersion until the solution changed from blue to colourless.

If 6 ml of hydrogen dispersion reduces one drop (20 µl) of the MB–platinum reagent, the concentration of dissolved hydrogen is 0,1 mg/l.

4.3 Gas chromatography

After the hydrogen UFB dispersion is sealed in a vial, the rubber stopper made of butyl rubber and aluminium cap can effectively prevent the diffusion of hydrogen molecules. Through the change of the hydrogen UFB dispersions from a liquid phase to a solid phase by the freezing process, the hydrogen molecules that dissolve in the dispersions are transferred to the gas phase in the vial. Thereafter, the gas components in the vial were separated into electrical signals that were sent to signal processing devices (computer) to obtain peaks corresponding to the separated gas components. Resultantly, the molar concentration of the hydrogen released from the UFB dispersions can be measured. Finally, the total hydrogen content in the hydrogen UFB dispersion can be calculated.

5 Apparatus and materials

5.1 Titration (oxidimetry) method

5.1.1 Reagent

5.1.1.1 MB, $C_{16}H_{18}ClN_3S$ Mass mol: 319,86 g/mol.

Form: brownish-red powder

Grade: analytical pure

5.1.1.2 Ethanol, C_2H_6O · 98 % (GC) Mass mol: 46,07 g/mol.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 7383-2:2024

5.1.1.3 Colloidal platinum, Pt concentration 2 g/kg.

5.1.2 Dropper or titrator

Plastic Pasteur pipette (dropper) commonly comes in 1 ml, 2 ml, 3 ml, and 5 ml which comes with a specific drop size of 10 µl, 20 µl, 25 µl, 35 µl, and 50 µl. For 2 ml plastic Pasteur pipette, one drop of the dropper should be approximately 17 mg or 0,02 ml in the MB–platinum reagent.

To reduce the dosing error, a pipettor covering a volume range of 0 µl to 20 µl can be used instead of a dropper; a commercial titrator can be used as well.

5.1.3 Clear vial

The vial had 6 ml graduated marks, and the colour change of the solution in the vial can be seen clearly.

5.2 Gas chromatography

5.2.1 Vial and clamping machine

5.2.1.1 Vial, made of borosilicate glass; the recommended volume of the vial was 100 ml. However, if the amount of water is less, volumes as low as 15 ml are also acceptable. In Annex D, the comparative results of the same water sample measured in different vials are displayed.

5.2.1.2 Rubber stopper, made of butyl rubber.

5.2.1.3 Cap, made of aluminium.

5.2.1.4 Clamping machine for a flip-top-cap.

5.2.2 UFBs elimination

5.2.2.1 4 °C-refrigerator.

5.2.2.2 –20 °C-freezer.

5.2.3 Carrier gases and column

5.2.3.1 Carrier gas, argon (Ar) or helium (He).

5.2.3.2 Standard hydrogen, can be obtained from a gas company or hydrogen generator (purity > 99,99 %).

5.2.3.3 Hydrogen calibration gases with 0,5 %, 1 %, 5 %, 10 %, 20 %, 50 % and 100 % hydrogen concentrations in air.

5.2.3.4 Gastight plunger syringe, 1 ml in volume.

5.2.3.5 Chromatographic column, molecular sieve 5A capillary/plot columns (Msieve-5A or MolSieve-5A).

5.2.4 Gas chromatograph

A thermal conductivity detector (TCD) should be equipped for gas chromatography for the detection of hydrogen gas present in an air mixture.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: BS ISO 7383-2:2024