

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Electroacoustics – Audiometric equipment –
Part 6: Instruments for the measurement of otoacoustic emissions**

**Électroacoustique – Appareils audiométriques –
Partie 6: Instruments pour la mesure des émissions otoacoustiques**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2022 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 300 terminological entries in English and French, with equivalent terms in 19 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC - webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 300 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 19 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.



IEC 60645-6

Edition 2.0 2022-03

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Electroacoustics – Audiometric equipment –
Part 6: Instruments for the measurement of otoacoustic emissions**

**Électroacoustique – Appareils audiométriques –
Partie 6: Instruments pour la mesure des émissions otoacoustiques**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 17.140.50

ISBN 978-2-8322-1086-6

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

© Registered trademark of the International Electrotechnical Commission
Marque déposée de la Commission Electrotechnique Internationale

Provided by IHS Markit under license with IEC
No reproduction or networking permitted without license from IHS Markit
Full document details: IEC 60645-6-2022

AVA Documents
Document Sharing Hub

CONTENTS

FOREWORD	4
INTRODUCTION	6
1 Scope	7
2 Normative references	7
3 Terms and definitions	8
4 Requirements for specific instruments	9
5 General specifications	9
5.1 Acoustic stimulus system	9
5.1.1 General requirements	9
5.1.2 Stimulus types	9
5.1.3 Stimulus frequency range	10
5.1.4 Stimulus level	10
5.1.5 Intermodulation distortion	11
5.2 Test quality assuring system	11
5.2.1 Stability of acoustic response in the external auditory meatus	11
5.2.2 Test quality assurance	11
5.2.3 Individual stimulus recordings	11
5.3 Measuring system	12
5.3.1 Units of measurement	12
5.3.2 Measurement range	12
5.3.3 Accuracy of measurement	12
5.3.4 Frequency range	12
5.3.5 Noise reduction	12
5.3.6 Response detection	12
5.3.7 Response quality estimates	12
5.3.8 Normative values	12
5.4 Presentation of results	12
6 Demonstration of conformity with specifications	13
6.1 General	13
6.2 Probe signal	13
6.2.1 Probe signal frequency spectrum	13
6.2.2 Probe signal level and harmonic distortion	13
6.2.3 Probe measurement accuracy	13
6.3 Complete system	14
6.4 Maximum permitted expanded uncertainty of measurements $U_{\max}$	14
7 General requirements	15
7.1 Marking	15
7.2 Instruction manual	15
7.3 Safety requirements	15
7.4 Immunity to power and radiofrequency fields	15
7.5 Warm-up time	15
7.6 Voltage supply variation and environmental conditions	15
7.6.1 Mains operation	15
7.6.2 Battery operation	15
7.6.3 Environmental conditions	15

8	Additional characteristics to be specified by the manufacturer	16
9	Periodic calibration	16
	Bibliography	17
	Table 1 – Mandatory functions for otoacoustic emission instruments	9
	Table 2 – Documentation of test conditions, parameters and results	13
	Table 3 – Values of $U_{\max}$ for conformance and periodic calibration measurements	14

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**ELECTROACOUSTICS –
AUDIOMETRIC EQUIPMENT –****Part 6: Instruments for the measurement of otoacoustic emissions****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 60645-6 has been prepared by IEC technical committee 29: Electroacoustics. It is an International Standard.

This second edition cancels and replaces the first edition published in 2009. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) the nominal test frequency used in DPOAE is now defined as the higher of the two frequencies, f_2 ;
- b) the permitted deviation of the stimulus signal for TEOAE has been specified;
- c) the frequency range for DPOAE stimulus signals has been redefined;
- d) the stimulus level requirements for TEOAE have been redefined;
- e) the stimulus level requirements for DPOAE have been redefined;

- f) the harmonic distortion requirements for DPOAE have been redefined;
- g) a minimum measurement range for DPOAE has been added.

The text of this International Standard is based on the following documents:

Draft	Report on voting
29/1109/FDIS	29/1114/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at <http://www.iec.ch/standardsdev/publications>.

A list of all parts in the IEC 60645 series, published under the general title *Electroacoustics – Audiometric equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

Developments in the field of diagnostic hearing measurement have resulted in a number of instruments designed to evaluate the otoacoustic emissions of the human ear. Such emissions may be evoked by acoustic test signals having different spectral and temporal characteristics.

The practical use of such instruments concerns the measurement of sound energy emitted by the inner ear and its separation from sounds emerging from physiological or other sources.

The spontaneous otoacoustic emissions (SOAE) and stimulus frequency otoacoustic emissions (SFOAE), which comprise part of the otoacoustic emissions, are not covered by this document.

Conformance to the performance specification in this document is demonstrated when a measured deviation from a design goal equals or does not exceed the corresponding acceptance limit(s), and the laboratory has demonstrated that the associated uncertainty of measurement equals or does not exceed the maximum permitted uncertainty specified in this document.

ELECTROACOUSTICS – AUDIOMETRIC EQUIPMENT –

Part 6: Instruments for the measurement of otoacoustic emissions

1 Scope

This part of IEC 60645 applies to instruments designed primarily for the measurement of otoacoustic emissions in the human external auditory meatus evoked by acoustic probe stimuli. This document defines the characteristics to be specified by the manufacturer, specifies minimum mandatory functions for two types of instruments and provides performance specifications applicable to both instrument types. This document describes methods to be used to demonstrate conformance with the specifications in this document and guidance on methods for periodic calibration.

The purpose of this document is to ensure that measurements made under comparable test conditions with different instruments complying with this document will be consistent. Instruments can provide a measurement function not specifically within the scope of this document and still comply with the relevant requirements of this document for the functions that are within the scope. This document is not intended to restrict development or incorporation of new features, nor to discourage innovative approaches.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60318-4, *Electroacoustics – Simulators of human head and ear – Part 4: Occluded-ear simulator for the measurement of earphones coupled to the ear by means of ear inserts*

IEC 60318-5, *Electroacoustics – Simulators of human head and ear – Part 5: 2 cm³ coupler for the measurement of hearing aids and earphones coupled to the ear by means of ear inserts*

IEC 60601-1, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1-2, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic disturbances – Requirements and tests*

IEC 60645-1:2017, *Electroacoustics – Audiometric equipment – Part 1: Equipment for pure-tone and speech audiometry*

IEC 60645-3:2020, *Electroacoustics – Audiometric equipment – Part 3: Test signals of short duration*

ISO/IEC Guide 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

otoacoustic emissions

OAE

acoustic signals generated in the inner ear which can be recorded in the external auditory meatus

3.2

transient-evoked otoacoustic emissions

TEOAE

acoustic signals emitted by the inner ear after stimulation with a stimulus of short duration

3.3

distortion product otoacoustic emissions

DPOAE

acoustic signals generated in the inner ear during stimulation with two pure tones

Note 1 to entry: The pure tones are frequencies f_1 and f_2 , f_1 being the lower frequency.

Note 2 to entry: The frequencies of the DPOAE are given by the formulas for intermodulation distortions (IMD), i.e. $2f_1 - f_2$, $2f_2 - f_1$, etc.

3.4

nominal test frequency

frequency for which a DPOAE measurement is reported

3.5

primary tones

pure-tone stimuli used to evoke DPOAE

3.6

probe

part of the instrument, usually containing acoustic transducers, interfacing the instrument to the ear

3.7

ear tip

device used to assist acoustic coupling, to reduce acoustic leakage, to reduce the influence of environmental noise on measurements and to aid retention of the probe in the external auditory meatus

3.8

probe signal

acoustic stimulus signal that is emitted into the external auditory meatus by means of a probe

3.9

peak-to-peak equivalent sound pressure level peSPL

root mean squared (RMS) value of a long-duration sinusoidal sound signal which, when compared under the same test conditions with a short-duration output signal from the transducer under test, has the same peak-to-peak value (i.e., difference between the extreme positive and the extreme negative values) as the short-duration signal

Note 1 to entry: See IEC 60645-3:2020, Figure 2.

4 Requirements for specific instruments

Two different types of otoacoustic emission instruments are specified by the requirements for minimum mandatory functions (see Table 1). Additional functions are not precluded. The two types relate to their presumed primary application (diagnostic/clinical or screening); however, a device of one type is not required to comply with the additional specifications of the other type.

Table 1 – Mandatory functions for otoacoustic emission instruments

	Type	
	1 Diagnostic/clinical	2 Screening
Automatic test	x	x
Manual test	x	
Display of PASS/REFER		x
Display of detailed result in graphical and/or tabular format	x	
Display of stability of acoustic response in the external auditory meatus (see 5.2.1)	x	
Display of response quality estimate (see 5.3.7)	x	
Digital storage of detailed result	x	
Export of full test report	x	
Type 1. This type of devices shall include the ability to manually start the test and to adjust the parameters of the test.		
Type 2. This type of device shall include the ability to automatically start the test.		

5 General specifications

5.1 Acoustic stimulus system

5.1.1 General requirements

Specifications for the acoustic stimulus system are as given in the relevant parts of Clause 6, Clause 8 and Clause 10 of IEC 60645-1:2017 and Clause 5 of IEC 60645-3:2020 with the exceptions specified below.

NOTE If the instrument is designed to also allow the measurement of hearing thresholds, the full text of the relevant clauses of IEC 60645-1:2017 applies.

5.1.2 Stimulus types

5.1.2.1 General

The general properties and temporal characteristics of the acoustic stimulus signals are specified within 5.1.2.2 and 5.1.2.3 depending on the type of OAE being measured.

5.1.2.2 TEOAE

The full characteristics of the short-duration signal used for the measurements of TEOAE shall be specified by the manufacturer (i.e., as specified in IEC 60645-3:2020).

NOTE A series of clicks with different polarity and levels is often used and this is usually referred to as a "non-linear click series". The specifications found in IEC 60645-3 are applicable to each single click in the series.

5.1.2.3 DPOAE

The stimulus signal used for the measurement of DPOAE shall be composed of two primary tones with frequencies f_1 and f_2 . Although the DPOAE of principal interest is at a frequency of $2f_1 - f_2$, the nominal test frequency of the measurement normally refers to f_2 . If f_1 is used as the nominal test frequency, this shall be stated by the manufacturer. If additional test signals are used (such as those used for masking), their full characteristics shall be specified by the manufacturer.

5.1.3 Stimulus frequency range

5.1.3.1 General

The frequency content of the stimulus signal shall, as a minimum, meet the requirements specified in 5.1.3.2 and 5.1.3.3 depending on the type of OAEs being measured.

5.1.3.2 TEOAE

The frequency spectrum of the transient stimulus signal shall at least cover the range from 0,5 kHz to 4 kHz for Type 1 instruments and the range from 1,5 kHz to 3 kHz for Type 2 instruments. The stimulus level frequency spectrum shall be flat within a limit of ± 5 dB as measured in an occluded-ear simulator according to IEC 60318-4 or a 2 cm³ coupler according to IEC 60318-5, using the ear simulator or 2 cm³ coupler microphone, over the frequency range.

5.1.3.3 DPOAE

For the measurement of DPOAE, nominal stimulus frequencies between 0,75 kHz and 8 kHz in at least three steps per octave shall be provided in instruments of Type 1 and at least two frequencies between 1 kHz and 4 kHz for Type 2. The frequency ratio of the two primary tones shall be stated by the manufacturer and shall normally be from 1:1,15 to 1:1,25.

The acceptance limit of the actual frequencies is ± 1 %.

5.1.4 Stimulus level

5.1.4.1 General

The sound pressure level of the stimulus signals shall be variable within the ranges specified in 5.1.4.2 and 5.1.4.3 depending on the type of OAEs. Its actual value within the residual ear-canal volume shall be measured prior to each recording with the probe microphone.

5.1.4.2 TEOAE

For Type 1 instruments, the stimulus level shall be adjustable with a step size no greater than 5 dB and include a range of at least 60 dB peSPL to 85 dB peSPL. For Type 2 instruments, a single fixed level of stimulus is acceptable, and this level shall be stated clearly in the documentation since it impacts on the specificity of screening for a certain level of hearing loss. The stimulus levels stated shall be measured in an occluded-ear simulator according to IEC 60318-4 or a 2 cm³ coupler according to IEC 60318-5, using the occluded-ear simulator or 2 cm³ coupler microphone.