



International
Standard

ISO/IEEE
11073-10471

**Health informatics — Device
interoperability —**

**Part 10471:
Personal health device
communication — Device
specialization — Independent living
activity hub**

*Informatique de santé — Interopérabilité des dispositifs —
Partie 10471: Communication entre dispositifs de santé
personnels — Spécialisation des dispositifs — Concentrateur
d'activités pour une vie autonome*

**Second edition
2024-09**

Reference number
ISO/IEEE 11073-10471:2024(en)

Authorized licensed use limited to: Tsinghua University. Downloaded on February 24, 2026 at 09:17:35 UTC from IEEE Xplore. Restrictions apply.

Full document details: [IEEE 11073-10471-2024](#)

© IEEE 2024

AVA Documents
Document Sharing Hub



COPYRIGHT PROTECTED DOCUMENT

© IEEE 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from IEEE at the address below.

Institute of Electrical and Electronics Engineers, Inc
3 Park Avenue, New York
NY 10016-5997, USA

Email: stds.ipr@ieee.org
Website: www.ieee.org

Published in Switzerland

© IEEE 2024 – All rights reserved

Authorized licensed use limited to: Tsinghua University. Downloaded on February 24, 2026 at 09:17:35 UTC from IEEE Xplore. Restrictions apply.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted (see www.iso.org/directives).

IEEE Standards documents are developed within the IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (IEEE-SA) Standards Board. The IEEE develops its standards through a consensus development process, approved by the American National Standards Institute, which brings together volunteers representing varied viewpoints and interests to achieve the final product. Volunteers are not necessarily members of the Institute and serve without compensation. While the IEEE administers the process and establishes rules to promote fairness in the consensus development process, the IEEE does not independently evaluate, test, or verify the accuracy of any of the information contained in its standards.

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

ISO/IEEE 11073-10471 was prepared by the IEEE 11073 Standards Committee of the IEEE Engineering in Medicine and Biology Society (as IEEE Std 11073-10471) and drafted in accordance with its editorial rules. It was adopted, under the “fast-track procedure” defined in the Partner Standards Development Organization cooperation agreement between ISO and IEEE, by Technical Committee ISO/TC 215, *Health informatics*.

This second edition cancels and replaces the first edition (ISO/IEEE 11073-10471:2010), which has been technically revised.

The main changes are as follows:

- added current value sensors (temperature, humidity, utility usage);
- added humidity alert sensor;
- added support for location sensors;
- added support for Base-Offset-Time;

- added new location identifiers;
- incorporated location identifiers from ZigBee Home Automation and Health Care profiles;
- added new appliance identifiers;
- changed from bit string to simple OIDs to report events;
- updated the example to simple OID;
- changed supplementary types to have upper 11 bits for location and lower 5 bits for number to support more codes within partition;
- updated the version of this device specialization;
- updated the association details based on new version;
- updated the wording in 6.3 regarding the Not Allowed and Observational;
- added systematic names;
- tamper detected flag made a general health status flag for all sensors;
- enumeration codes provided to other systems (ZigBee);
- location RefId discriminator changed from parentheses to underscore ((3)->_3);
- encounter object added.

A list of all parts in the ISO 11073 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Abstract: Within the context of the ISO/IEEE 11073 family of standards for device communication, a normative definition of communication between personal telehealth independent living activity hub devices and compute engines (e.g., cell phones, personal computers, personal health appliances, set top boxes) in a manner that enables plug-and-play interoperability is established in this standard. Appropriate portions of existing standards, including ISO/IEEE 11073 terminology, information models, application profile standards, and transport standards are leveraged. The use of specific term codes, formats, and behaviors in telehealth environments restricting optionality in base frameworks in favor of interoperability are specified. A common core of communication functionality for personal telehealth independent living activity hubs is defined in this standard.

Keywords: IEEE 11073-10471™, independent living activity hub, medical device communication, personal health devices

The Institute of Electrical and Electronics Engineers, Inc.
3 Park Avenue, New York, NY 10016-5997, USA

Copyright © 2023 by The Institute of Electrical and Electronics Engineers, Inc.
All rights reserved. Published 22 June 2023. Printed in the United States of America.

IEEE is a registered trademark in the U.S. Patent & Trademark Office, owned by The Institute of Electrical and Electronics Engineers, Incorporated.

PDF: ISBN 978-1-5044-9825-8 STD26246
Print: ISBN 978-1-5044-9826-5 STDPD26246

IEEE prohibits discrimination, harassment, and bullying.

For more information, visit <https://www.ieee.org/about/corporate/governance/p9-26.html>.

No part of this publication may be reproduced in any form, in an electronic retrieval system or otherwise, without the prior written permission of the publisher.

Important Notices and Disclaimers Concerning IEEE Standards Documents

IEEE Standards documents are made available for use subject to important notices and legal disclaimers. These notices and disclaimers, or a reference to this page (<https://standards.ieee.org/ipr/disclaimers.html>), appear in all standards and may be found under the heading “Important Notices and Disclaimers Concerning IEEE Standards Documents.”

Notice and Disclaimer of Liability Concerning the Use of IEEE Standards Documents

IEEE Standards documents are developed within IEEE Societies and subcommittees of IEEE Standards Association (IEEE SA) Board of Governors. IEEE develops its standards through an accredited consensus development process, which brings together volunteers representing varied viewpoints and interests to achieve the final product. IEEE Standards are documents developed by volunteers with scientific, academic, and industry-based expertise in technical working groups. Volunteers are not necessarily members of IEEE or IEEE SA and participate without compensation from IEEE. While IEEE administers the process and establishes rules to promote fairness in the consensus development process, IEEE does not independently evaluate, test, or verify the accuracy of any of the information or the soundness of any judgments contained in its standards.

IEEE makes no warranties or representations concerning its standards, and expressly disclaims all warranties, express or implied, concerning this standard, including but not limited to the warranties of merchantability, fitness for a particular purpose and non-infringement. In addition, IEEE does not warrant or represent that the use of the material contained in its standards is free from patent infringement. IEEE standards documents are supplied “AS IS” and “WITH ALL FAULTS.”

Use of an IEEE standard is wholly voluntary. The existence of an IEEE Standard does not imply that there are no other ways to produce, test, measure, purchase, market, or provide other goods and services related to the scope of the IEEE standard. Furthermore, the viewpoint expressed at the time a standard is approved and issued is subject to change brought about through developments in the state of the art and comments received from users of the standard.

In publishing and making its standards available, IEEE is not suggesting or rendering professional or other services for, or on behalf of, any person or entity, nor is IEEE undertaking to perform any duty owed by any other person or entity to another. Any person utilizing any IEEE Standards document, should rely upon his or her own independent judgment in the exercise of reasonable care in any given circumstances or, as appropriate, seek the advice of a competent professional in determining the appropriateness of a given IEEE standard.

IN NO EVENT SHALL IEEE BE LIABLE FOR ANY DIRECT, INDIRECT, INCIDENTAL, SPECIAL, EXEMPLARY, OR CONSEQUENTIAL DAMAGES (INCLUDING, BUT NOT LIMITED TO: THE NEED TO PROCURE SUBSTITUTE GOODS OR SERVICES; LOSS OF USE, DATA, OR PROFITS; OR BUSINESS INTERRUPTION) HOWEVER CAUSED AND ON ANY THEORY OF LIABILITY, WHETHER IN CONTRACT, STRICT LIABILITY, OR TORT (INCLUDING NEGLIGENCE OR OTHERWISE) ARISING IN ANY WAY OUT OF THE PUBLICATION, USE OF, OR RELIANCE UPON ANY STANDARD, EVEN IF ADVISED OF THE POSSIBILITY OF SUCH DAMAGE AND REGARDLESS OF WHETHER SUCH DAMAGE WAS FORESEEABLE.

Translations

The IEEE consensus development process involves the review of documents in English only. In the event that an IEEE standard is translated, only the English version published by IEEE is the approved IEEE standard.

Official statements

A statement, written or oral, that is not processed in accordance with the IEEE SA Standards Board Operations Manual shall not be considered or inferred to be the official position of IEEE or any of its committees and shall not be considered to be, nor be relied upon as, a formal position of IEEE. At lectures, symposia, seminars, or educational courses, an individual presenting information on IEEE standards shall make it clear that the presenter’s views should be considered the personal views of that individual rather than the formal position of IEEE, IEEE SA, the Standards Committee, or the Working Group. Statements made by volunteers may not represent the formal position of their employer(s) or affiliation(s).

Comments on standards

Comments for revision of IEEE Standards documents are welcome from any interested party, regardless of membership affiliation with IEEE or IEEE SA. However, **IEEE does not provide interpretations, consulting information, or advice pertaining to IEEE Standards documents.**

Suggestions for changes in documents should be in the form of a proposed change of text, together with appropriate supporting comments. Since IEEE standards represent a consensus of concerned interests, it is important that any responses to comments and questions also receive the concurrence of a balance of interests. For this reason, IEEE and the members of its Societies and subcommittees of the IEEE SA Board of Governors are not able to provide an instant response to comments, or questions except in those cases where the matter has previously been addressed. For the same reason, IEEE does not respond to interpretation requests. Any person who would like to participate in evaluating comments or in revisions to an IEEE standard is welcome to join the relevant IEEE working group. You can indicate interest in a working group using the Interests tab in the Manage Profile & Interests area of the [IEEE SA myProject system](#).¹ An IEEE Account is needed to access the application.

Comments on standards should be submitted using the [Contact Us](#) form.²

Laws and regulations

Users of IEEE Standards documents should consult all applicable laws and regulations. Compliance with the provisions of any IEEE Standards document does not constitute compliance to any applicable regulatory requirements. Implementers of the standard are responsible for observing or referring to the applicable regulatory requirements. IEEE does not, by the publication of its standards, intend to urge action that is not in compliance with applicable laws, and these documents may not be construed as doing so.

Data privacy

Users of IEEE Standards documents should evaluate the standards for considerations of data privacy and data ownership in the context of assessing and using the standards in compliance with applicable laws and regulations.

Copyrights

IEEE draft and approved standards are copyrighted by IEEE under US and international copyright laws. They are made available by IEEE and are adopted for a wide variety of both public and private uses. These include both use, by reference, in laws and regulations, and use in private self-regulation, standardization, and the promotion of engineering practices and methods. By making these documents available for use and adoption by public authorities and private users, neither IEEE nor its licensors waive any rights in copyright to the documents.

Photocopies

Subject to payment of the appropriate licensing fees, IEEE will grant users a limited, non-exclusive license to photocopy portions of any individual standard for company or organizational internal use or individual, non-commercial use only. To arrange for payment of licensing fees, please contact Copyright Clearance Center, Customer Service, 222 Rosewood Drive, Danvers, MA 01923 USA; +1 978 750 8400; <https://www.copyright.com/>. Permission to photocopy portions of any individual standard for educational classroom use can also be obtained through the Copyright Clearance Center.

Updating of IEEE Standards documents

Users of IEEE Standards documents should be aware that these documents may be superseded at any time by the issuance of new editions or may be amended from time to time through the issuance of amendments, corrigenda, or errata. An official IEEE document at any point in time consists of the current edition of the document together with any amendments, corrigenda, or errata then in effect.

¹ Available at: <https://development.standards.ieee.org/myproject-web/public/view.html#landing>.

² Available at: <https://standards.ieee.org/content/ieee-standards/en/about/contact/index.html>.

Every IEEE standard is subjected to review at least every 10 years. When a document is more than 10 years old and has not undergone a revision process, it is reasonable to conclude that its contents, although still of some value, do not wholly reflect the present state of the art. Users are cautioned to check to determine that they have the latest edition of any IEEE standard.

In order to determine whether a given document is the current edition and whether it has been amended through the issuance of amendments, corrigenda, or errata, visit [IEEE Xplore](#) or [contact IEEE](#).³ For more information about the IEEE SA or IEEE's standards development process, visit the IEEE SA Website.

Errata

Errata, if any, for all IEEE standards can be accessed on the [IEEE SA Website](#).⁴ Search for standard number and year of approval to access the web page of the published standard. Errata links are located under the Additional Resources Details section. Errata are also available in [IEEE Xplore](#). Users are encouraged to periodically check for errata.

Patents

IEEE Standards are developed in compliance with the [IEEE SA Patent Policy](#).⁵

Attention is called to the possibility that implementation of this standard may require use of subject matter covered by patent rights. By publication of this standard, no position is taken by the IEEE with respect to the existence or validity of any patent rights in connection therewith. If a patent holder or patent applicant has filed a statement of assurance via an Accepted Letter of Assurance, then the statement is listed on the IEEE SA Website at <https://standards.ieee.org/about/sasb/patcom/patents.html>. Letters of Assurance may indicate whether the Submitter is willing or unwilling to grant licenses under patent rights without compensation or under reasonable rates, with reasonable terms and conditions that are demonstrably free of any unfair discrimination to applicants desiring to obtain such licenses.

Essential Patent Claims may exist for which a Letter of Assurance has not been received. The IEEE is not responsible for identifying Essential Patent Claims for which a license may be required, for conducting inquiries into the legal validity or scope of Patents Claims, or determining whether any licensing terms or conditions provided in connection with submission of a Letter of Assurance, if any, or in any licensing agreements are reasonable or non-discriminatory. Users of this standard are expressly advised that determination of the validity of any patent rights, and the risk of infringement of such rights, is entirely their own responsibility. Further information may be obtained from the IEEE Standards Association.

IMPORTANT NOTICE

IEEE Standards do not guarantee or ensure safety, security, health, or environmental protection, or ensure against interference with or from other devices or networks. IEEE Standards development activities consider research and information presented to the standards development group in developing any safety recommendations. Other information about safety practices, changes in technology or technology implementation, or impact by peripheral systems also may be pertinent to safety considerations during implementation of the standard. Implementers and users of IEEE Standards documents are responsible for determining and complying with all appropriate safety, security, environmental, health, and interference protection practices and all applicable laws and regulations.

³ Available at: <https://ieeexplore.ieee.org/browse/standards/collection/ieee>.

⁴ Available at: <https://standards.ieee.org/standard/index.html>.

⁵ Available at: <https://standards.ieee.org/about/sasb/patcom/materials.html>.

Participants

At the time this standard was completed, the Personal Health Devices Working Group had the following membership:

Daidi Zhong, *Chair*
Malcolm Clarke, *Vice Chair*
Raymond Krasinski, *Secretary*

Karsten Aalders
 Charles R. Abbruscato
 Nabil Abujbara
 Maher Abuzaid
 James Agnew
 Manfred Aigner
 Jorge Alberola
 David Aparisi
 Lawrence Arne
 Diego B. Arquillo
 Serafín Arroyo
 Muhammad Asim
 Kit August
 Doug Baird
 David Baker
 Anindya Bakshi
 Ananth Balasubramanian
 Sunlee Bang
 M. Jonathan Barkley
 Gilberto Barrón
 David Bean
 John Bell
 Olivia Bellamou-Huet
 Rudy Belliardi
 Daniel Bernstein
 George A. Bertos
 Chris Biernacki
 Ola Björnsne
 Thomas Blackadar
 Thomas Bluethner
 Douglas P. Bogia
 Xavier Boniface
 Shannon Boucousis
 Lyle G. Bullock, Jr.
 Bernard Burg
 Chris Burns
 Jeremy Byford-Rew
 Satya Calloji
 Carole C. Carey
 Craig Carlson
 Santiago Carot-Nemesio
 Seungchul Chae
 Yao Chen
 Jing Cheng
 Peggy Chien
 David Chiu
 Jinyong Choi
 Chia-Chin Chong
 Jinhan Chung
 John A. Cogan
 John T. Collins
 Cory Condek
 Todd H. Cooper
 Sandra Costanzo
 Douglas Coup

Nigel Cox
 Hans Crommenacker
 Tomio Crosley
 Allen Curtis
 Jesús Daniel Trigo
 David Davenport
 Russell Davis
 Sushil K. Deka
 Ciro de la Vega
 Jim Dello Stritto
 Kent Dicks
 Hyoungdo Do
 Fangjie Dong
 Jonathan Dougherty
 Xiaolian Duan
 Sourav Dutta
 Jakob Ehrensverd
 Fredrik Einberg
 Javier Escayola Calvo
 Mark Estes
 Leonardo Estevez
 Bosco T. Fernandes
 Christoph Fischer
 Morten Flintrup
 Russell Foster
 Eric Freudenthal
 Matthias Frohner
 Ken Fuchs
 Jing Gao
 Marcus Garbe
 John Garguilo
 Liang Ge
 Rick Geimer
 Igor Gejdos
 Ferenc Gerbovics
 Alan Godfrey
 Nicolae Goga
 Julian Goldman
 Raul Gonzalez Gomez
 Chris Gough
 Channa Gowda
 Charles M. Gropper
 Amit Gupta
 Jeff Guttmacher
 Rasmus Haahr
 Christian Habermann
 Michael Hagerty
 Jerry Hahn
 Robert Hall
 Shu Han
 Nathaniel Hamming
 Rickey L. Hampton
 Sten Hanke
 Aki Harma
 Jordan Hartmann

Kai Hassing
 Avi Hauser
 Nathaniel Heintzman
 Charles Henderson
 Jun-Ho Her
 Timothy L. Hirou
 Allen Hobbs
 Alex Holland
 Arto Holopainen
 Kris Holtzclaw
 Robert Hoy
 Anne Huang
 Guiling Huang
 Haofei Huang
 Zhiyong Huang
 David Hughes
 Robert D. Hughes
 Jiyoung Huh
 Hugh Hunter
 Philip O. Isaacson
 Atsushi Ito
 Michael Jaffe
 Praduman Jain
 Hu Jin
 Danny Jochelson
 Akiyoshi Kabe
 Steve Kahle
 Tomio Kamioka
 James J. Kang
 Kei Kariya
 Andy Kaschl
 Junzo Kashiara
 Ralph Kent
 Laurie M. Kermes
 Sanjay R. Kharche
 Ahmad Kheirandish
 Junhyung Kim
 Minh Kim
 Min-Joon Kim
 Taekon Kim
 Tetsuya Kimura
 Michael J. Kirwan
 Alfred Kloos
 Edward Koch
 Jeongmee Koh
 Jean-Marc Koller
 John Koon
 Patty Krantz
 Alexander Kraus
 Ramesh Krishna
 Geoffrey Kruse
 Falko Kuester
 Rafael Lajara
 Pierre Landau
 Jaechul Lee

JongMuk Lee
 Kyong Ho Lee
 Rami Lee
 Sungkee Lee
 Woojae Lee
 Jing Li
 Qiong Li
 Xiangchen Li
 Xiaoyu Li
 Patrick Lichter
 Jisoon Lim
 Wei-Jung Lo
 Charles Lowe
 Ling Luo
 Don Ludolph
 Christian Lusztick
 Bob MacWilliams
 Srikanth Madhurboothswaran
 Miriam L. Makhoul
 M. Sabarimalai Manikandan
 Romain Marmot
 Sandra Martinez
 Miguel Martínez de EsproncedaCámara
 Peter Mayhew
 Jim McCain
 LászlóMeleg
 Alexander Mense
 Behnaz Minaei
 Jinsei Miyazaki
 Madhu Mohan
 Erik Moll
 Darr Moore
 Chris Morel
 Carsten Mueglitz
 Soundharya Nagasubramanian
 Alex Neefus
 Trong-Nghia Nguyen-Dobinsky
 Michael E. Nidd
 Jim Niswander
 Hiroaki Niwamoto
 Thomas Norgall
 Yoshiteru Nozoe
 Abraham Ofek

 Brett Olive
 BegonyaOtal
 Marco Paleari
 Bud Panjwani
 Carl Pantiskas
 Harry P. Pappas
 Hanna Park
 Jong-Tae Park
 Myungeun Park

Phillip E. Pash
 Tong Bi Pei
 Soren Petersen
 James Petisce
 Peter Piction
 Michael Pliskin
 Varshney Prabodh
 Jeff Price
 Harald Prinzhorn
 Lifei Qian
 Harry Qiu
 Tanzilur Rahman
 Lin Ran
 Phillip Raymond
 Terrie Reed
 Barry Reinhold
 Brian Reinhold
 John G. Rhoads
 Jeffrey S. Robbins
 Chris Roberts
 Moskowitz Robert
 Stefan Robert
 Scott M. Robertson
 Timothy Robertson
 Sean Roche
 David Rosales
 Bill Saltzstein
 Giovanna Sannino
 Jose A. Santos-Cadenas
 Stefan Sauermann
 John Sawyer
 Alois Schloegl
 Paul S. Schluter
 Mark G. Schnell
 Richard A. Schrenker
 Antonio Scorpiniti
 KwangSeok Seo
 Riccardo Serafin
 Sid Shaw
 Frank Shen
 Min Shih
 Mazen Shihabi
 Redmond Shouldice
 Sternly K. Simon
 Marjorie Skubic
 Robert Smith
 Ivan Soh
 Motoki Sone
 Emily Sopensky
 Rajagopalan Srinivasan
 Nicholas Steblay
 Lars Steubesand
 John (Ivo) Stivoric

Hermann Suominen
 Lee Surprenant
 Ravi Swami
 Ray Sweidan
 Na Tang
 Haruyuyki Tatsumi
 Isabel Tejero
 Chn Jonas Tirén
 Janet Traub
 Gary Tschautscher
 Masato Tsuchid
 Ken Tubman
 Akib Uddin
 Sunil Unadkat
 Fabio Urbani
 Philipp Urbauer
 Laura Vanzago
 Alpo Värri
 Andrei Vasilateanu
 Dalimar Velez
 Martha Velez
 Rudi Voon
 Isobel Walker
 David Wang
 Jerry P. Wang
 Shiwei Wang
 Yao Wang
 Yi Wang
 Steve Warren
 Fujio Watanabe
 Toru Watsuji
 Kathleen Wible
 Paul Williamson
 Jia-Rong Wu
 Will Wykeham
 Arton Xhafa
 Ricky Yang
 Melanie S. Yeung
 Qiang Yin
 Done-Sik Yoo
 Zhi Yu
 Jianchao Zeng
 Jason Zhang
 Zerui Zhang
 Shibai Zhao
 Yu Zhao
 Liang Zheng
 Yuanhong Zhong
 Qing Zhou
 Miha Zoubek
 Szymon Zyskotter

The following members of the individual Standards Association balloting group voted on this standard. Balloters may have voted for approval, disapproval, or abstention.

Bjoern Andersen
 William Byrd
 Randy Carroll
 Keith Chow
 Malcolm Clarke

Immanuel Freedman
 David Fuschi
 Randall Groves
 Werner Hoelzl
 Noriyuki Ikeuchi

Atsushi Ito
 Raj Jain
 Ghalib Muhammad Waqas Janjua
 Piotr Karocki

Raymond Krasinski
H. Moll
Nick S.A. Nikjoo

Beth Pumo
Stefan Schlichting
Walter Struppler

Oren Yuen
Jia Zheng

When the IEEE SA Standards Board approved this standard on 30 March 2023, it had the following membership:

David J. Law, *Chair*
Ted Burse, *Vice Chair*
Gary Hoffman, *Past Chair*
Konstantinos Karachalios, *Secretary*

Sara R. Biyabani
Doug Edwards
Ramy Ahmed Fathy
Guido R. Hiertz
Yousef Kimiagar
Joseph L. Koepfinger*
Thomas Koshy
John D. Kulick

Joseph S. Levy
Howard Li
Gui Lin
Johnny Daozhuang Lin
Xiaohui Liu
Kevin W. Lu
Daleep C. Mohla
Andrew Myles

Paul Nikolich
Annette D. Reilly
Robby Robson
Lei Wang
F. Keith Waters
Karl Weber
Philip B. Winston
Don Wright

*Member Emeritus

Introduction

This introduction is not part of IEEE Std 11073-10471-2023, IEEE Standard for Health informatics—Device Interoperability—Part 10471: Personal health device communication—Device specialization—Independent living activity hub.

ISO/IEEE 11073 standards enable communication between medical devices and external computer systems. This document uses the optimized framework created in ISO/IEEE 11073:20601 and describes a specific, interoperable communication approach for the Independent Living Activity Hub of personal health devices. These standards align with and draw on the existing clinically focused standards to provide support for communication of data from clinical or personal health devices (PHDs).

Contents

1. Overview	1
1.1 Scope	1
1.2 Purpose	1
1.3 Word usage	1
1.4 Context	1
2. Normative references	2
3. Definitions, acronyms, and abbreviations	2
3.1 Definitions	2
3.2 Acronyms and abbreviations	3
4. Introduction to ISO/IEEE 11073 personal health devices	3
4.1 General	3
4.2 Introduction to IEEE Std 11073-20601 modeling constructs	4
4.3 Compliance with other standards	4
5. Independent living activity hub device concepts and modalities	5
5.1 General	5
5.2 Concepts	5
5.3 Collected data	6
5.4 Reporting data from multiple persons	11
6. Independent living activity hub domain information model	12
6.1 Overview	12
6.2 Class extensions	12
6.3 Object instance diagram	12
6.4 Types of configuration	15
6.5 Medical device system object	16
6.6 Numeric objects	19
6.7 Real-time sample array objects	33
6.8 Enumeration objects	33
6.9 PM-store objects	50
6.10 Scanner objects	50
6.11 Class extension objects	50
6.12 Independent living activity hub information model extensibility rules	50
7. Independent living activity hub service model	51
7.1 General	51
7.2 Object access services	51
7.3 Object access event report services	52
8. Independent living activity hub communication model	52
8.1 Overview	52
8.2 Communication characteristics	53
8.3 Association procedure	53
8.4 Configuring procedure	55
8.5 Operating procedure	55
8.6 Time synchronization	56

9. Test associations 56

 9.1 Behavior with standard configuration 56

 9.2 Behavior with extended configurations 56

10. Conformance 57

 10.1 Applicability 57

 10.2 Conformance specification 57

 10.3 Levels of conformance 57

 10.4 Implementation conformance statements (ICSs) 58

Annex A (informative) Bibliography 62

Annex B (normative) Any additional ASN.1 definitions 63

Annex C (normative) Allocation of identifiers 66

 C.1 Systematic derivations of terms and codes for independent living monitoring measurements 73

Annex D 100

(informative) Message sequence examples 100

Annex E (informative) Protocol data unit examples 102

 E.1 General 102

 E.2 Association information exchange 102

 E.3 Configuration information exchange—extended configuration 104

 E.4 GET MDS attributes service 106

 E.5 Data reporting 107

 E.6 Disassociation 108

Annex F (informative) Revision History 109

Health Informatics—Device Interoperability— Part 10471: Personal Health Device Communication—Device Specialization—Independent Living Activity Hub

1. Overview

1.1 Scope

Within the context of the ISO/IEEE 11073 family of standards for device communication, this standard establishes a normative definition of the communication between independent living activity hubs and managers (e.g., cell phones, personal computers, personal health appliances, and set top boxes) in a manner that enables plug-and-play interoperability. It leverages appropriate portions of existing standards, including ISO/IEEE 11073 terminology and information models. It specifies the use of specific term codes, formats, and behaviors in telehealth environments restricting ambiguity in base frameworks in favor of interoperability. This standard defines a common core of communication functionality for independent living activity hubs. In this context, independent living activity hubs are defined as devices that communicate with simple situation monitors (binary sensors), normalize information received from the simple environmental monitors, and provide this normalized information to one or more managers. This information can be examined, for example, to determine when a person's activities/behaviors have deviated significantly from what is normal for them such that relevant parties can be notified. Independent living activity hubs will normalize information from the following simple situation monitors (binary sensors) for the initial release of the proposed standard: fall sensor, motion sensor, door sensor, bed/chair occupancy sensor, light switch sensor, smoke sensor, (ambient) temperature threshold sensor, personal emergency response system (PERS), and enuresis sensor (bed-wetting).

1.2 Purpose

This standard addresses a need for an openly defined, independent standard for controlling information exchange to and from personal health devices and managers (e.g., cell phones, personal computers, personal health appliances, and set top boxes). Interoperability is the key to growing the potential market for these devices and to enabling people to be better informed participants in the management of their health.

1.3 Word usage

The word *shall* indicates mandatory requirements strictly to be followed in order to conform to the standard and from which no deviation is permitted (*shall* equals *is required to*).^{6,7}

The word *should* indicates that among several possibilities one is recommended as particularly suitable, without mentioning or excluding others; or that a certain course of action is preferred but not necessarily required (*should* equals *is recommended that*).

The word *may* is used to indicate a course of action permissible within the limits of the standard (*may* equals *is permitted to*).

The word *can* is used for statements of possibility and capability, whether material, physical, or causal (*can* equals *is able to*).

1.4 Context

See IEEE Std 11073-20601™ for an overview of the environment within which this standard is written.⁸

⁶ The use of the word *must* is deprecated and cannot be used when stating mandatory requirements; *must* is used only to describe unavoidable situations.

⁷ The use of *will* is deprecated and cannot be used when stating mandatory requirements; *will* is only used in statements of fact.

⁸ Information on references can be found in Clause 2.

This document, IEEE Std 11073-10471™, defines the device specialization for the independent living activity hub, being a specific agent type, and provides a description of the device concepts, its capabilities, and its implementation according to this standard.

This standard is based on IEEE Std 11073-20601, which in turn draws information from both IEEE Std 11073-10201™ [B5] and ISO/IEEE 11073:20101 [B9].⁹ The medical device encoding rules (MDERs) used within this standard are fully described in IEEE Std 11073-20601.

This standard reproduces relevant portions of the nomenclature found in ISO/IEEE 11073:10101 [B8] and adds new nomenclature codes for the purposes of this standard. Between this standard and IEEE Std 11073-20601, all required nomenclature codes for implementation are documented.

The object classes and attributes in this standard are identified by nomenclature codes. Each code consists of a reference identifier (RefID) string and an integer code value. By using a consistent nomenclature, interoperability is enhanced as all implementations maintain the same semantic meaning for the numeric codes. This standard leverages the existing nomenclature codes in ISO/IEEE 11073:10101.

This standard defines specialized nomenclature codes that will be collected in IEEE Std 11073-10101™. Between this standard, IEEE Std 11073-10101, IEEE Std 11073-20601, and other IEEE Std 11073-104xx, all required nomenclature codes for implementation are documented. New codes may be defined in newer versions/revisions of each of these documents. In the case of a conflict, where one term code has been assigned to two separate semantic concepts with different RefIDs, in general the oldest definition that is in actual use should take precedence. The same policy applies when one RefID has two different code values assigned in different specifications. The resolution of such conflicts will be determined through joint action by the responsible workgroups and other stakeholders and any corrective action published as corrigenda.

NOTE—In this standard, IEEE Std 11073-104zz is used to refer to the collection of device specialization standards that utilize IEEE Std 11073-20601, where zz can be any number from 01 to 99, inclusive.¹⁰

2. Normative references

The following referenced documents are indispensable for the application of this document (i.e., they must be understood and used, so that each referenced document is cited in text and its relationship to this document is explained). For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments or corrigenda) applies.

IEEE Std 11073-20601™-2019, Health informatics—Personal health device communication—Part 20601: Application profile—Optimized Exchange Protocol.^{11,12}

ISO/IEEE 11073:10101-2020, Health informatics—Point-of-care medical device communication—Part 10101: Nomenclature.¹³

See Annex A for all informative material referenced by this standard.

3. Definitions, acronyms, and abbreviations

For the purposes of this standard, the following terms and definitions apply. *The IEEE Standard Dictionary Online* **Error! Reference source not found.** should be referenced for terms not defined in this clause.

3.1 Definitions

agent: A node that collects and transmits personal health data to an associated manager.

⁹ The numbers in brackets correspond to those of the bibliography in Annex A.

¹⁰ Notes in text, tables, and figures are given for information only and do not contain requirements needed to implement the standard.

¹¹ IEEE publications are available from The Institute of Electrical and Electronics Engineers, Inc. (<http://standards.ieee.org/>).

¹² The IEEE standards or products referred to in this clause are trademarks of The Institute of Electrical and Electronics Engineers, Inc.

¹³ ISO publications are available from the International Organization for Standardization (<https://www.iso.org/>) and the American National Standards Institute (<https://www.ansi.org/>).

class: In object-oriented modeling, it describes the attributes, methods, and events that objects instantiated from the class utilize.

compute engine: *See:* **manager**.

device: A term used to refer to a physical apparatus implementing either an agent or a manager role.

handle: An unsigned 16-bit number that is locally unique and identifies one of the object instances within an agent.

manager: A node receiving data from one or more agent systems. Some examples of managers include a cellular phone, health appliance, set top box, or a computer system.

obj-handle: *See:* **handle**.

object: In object-oriented modeling, a particular instantiation of a class. The instantiation realizes attributes, methods, and events from the class.

personal health device: A device used in personal health applications.

personal telehealth device: *See:* **personal health device**.

sensor: An apparatus that measures physical properties. These comprise the primary inputs to an independent living activity hub agent.

3.2 Acronyms and abbreviations

APDU	application protocol data unit
ASN.1	Abstract Syntax Notation One
DIM	domain information model
EUI-64	extended unique identifier (64 bits)
ICS	implementation conformance statement
MDC	medical device communication
MDER	medical device encoding rules
MDS	medical device system
MOC	managed object class
OID	object identifier
PDU	protocol data unit
PERS	personal emergency response system
PHD	personal health device
PIR	Passive Infrared
RT-SA	real-time sample array
VMO	virtual medical object
VMS	virtual medical system

4. Introduction to ISO/IEEE 11073 personal health devices

4.1 General

This standard and the remainder of the series of ISO/IEEE 11073 personal health device (PHD) standards fit in the larger context of the ISO/IEEE 11073 series of standards. The full suite of standards enables agents to interconnect and interoperate with managers and with computerized health-care information systems. See IEEE Std 11073-20601 for a description of the guiding principles for this series of ISO/IEEE 11073 personal health device standards.

IEEE Std 11073-20601 supports the modeling and implementation of an extensive set of personal health devices. This standard defines aspects of the independent living activity hub device. It describes all aspects necessary to implement the application layer services and data exchange protocol between an ISO/IEEE 11073 PHD independent living activity hub agent and a manager. This standard defines a subset of the objects and functionality contained in IEEE Std 11073-20601 and extends and adds definitions where appropriate. All new definitions are given in Annex B in abstract syntax notation one (ASN.1) **Error! Reference source not found.** All nomenclature codes referenced in this standard are collected in Annex C.

4.2 Introduction to IEEE Std 11073-20601 modeling constructs

4.2.1 General

The ISO/IEEE 11073 series of standards, and in particular IEEE Std 11073-20601, is based on an object-oriented systems management paradigm. The overall system model is divided into three principal components: the domain information model (DIM), the service model, and the communication model. See IEEE Std 11073-20601 for a detailed description of the modeling constructs.

4.2.2 Domain information model (DIM)

The DIM is a hierarchical model that describes an agent as a set of objects. These objects and their attributes represent the elements that control behavior and report on the status of the agent and data that an agent can communicate to a manager. Communication between the agent and the manager is defined by the application protocol in IEEE Std 11073-20601.

4.2.3 Service model

The service model defines the conceptual mechanisms for the data exchange services. Such services are mapped to messages that are exchanged between the agent and the manager. Protocol messages within the ISO/IEEE 11073 series of standards are defined in ASN.1. The messages defined in IEEE Std 11073-20601 can coexist with messages defined in other standard application profiles defined in the ISO/IEEE 11073 series of standards.

4.2.4 Communication model

In general, the communication model supports the topology of one or more agents communicating over logical point-to-point connections to a single manager. For each logical point-to-point connection, the dynamic system behavior is defined by a connection state machine, as specified in IEEE Std 11073-20601.

4.2.5 Implementing the models

An agent implementing this standard shall implement all mandatory elements of the information, service, and communication models as well as all conditional elements where the condition is met. The agent should implement the recommended elements, and it may implement any combination of the optional elements. A manager implementing this standard shall utilize at least one of the mandatory, conditional, recommended, or optional elements. In this context, “utilize” means to use the element as part of the primary function of the manager device. For example, a manager whose primary function is to display data would need to display a piece of data in the element in order to utilize it.

4.3 Compliance with other standards

Devices that comply with this standard may also be required to comply with other domain-and device specific standards that supersede the requirements of this standard with respect to issues including safety, reliability, and risk management. A user of this standard is expected to be familiar with all other such standards that apply and to comply with any higher specifications thus imposed.

Typically, medical devices should comply with the IEC 60601-1:2005 [B1] base standards with respect to electrical and mechanical safety as well as any device-specific standard as might be defined in the IEC 60601-2 [B2] series of standards. Software aspects may apply through standards such as IEC 62304:2006/EN 62304:2006 [B3]. Devices that comply with this standard implement higher layers of network software and utilize lower layers as appropriate to the application. The requirements on performance of such applications and conformance are defined elsewhere and are outside the scope of this standard. Moreover, the use of any medical equipment is subject to risk assessment and risk management appropriate to the

application. Some relevant examples are ISO 14971:2007 [B7] and IEC 80001-1:2010 [B4]. The requirements of such risk assessment and risk management and conformance are outside the scope of this standard. The applicable versions of the referenced safety related standards may differ per country.

5. Independent living activity hub device concepts and modalities

5.1 General

This clause presents the general concepts of independent living activity hub devices. In the context of personal health devices in this family of standards, an independent living activity hub is a device that aggregates activity data sensor events from multiple sensor data sources, all of which are used in the support of the independent living of one or more occupants. The occupants' environment may vary greatly and encompass a varying mixture of sensors; therefore, the activity data sensor events reported by any particular agent have a corresponding variance.

5.2 Concepts

While many data generating sensors exist, they have a number of properties in common that influence the design of this standard. Note that these are only generalities employed in the design and there may be instances where an activity data generating sensor exceeds the following properties:

- Price: usually, they are inexpensive sensors.
- Power: typically, they are low power sensors.
- Communication: typically, they use an inexpensive communication technique and are quite often wireless.
- Frequency: typically, they communicate infrequently or only when an event occurs.
- Quantity: there may be a wide range of sensors and many instances of any one sensor type.

It is the responsibility of the independent living activity hub agent to manage all these sensors. Management of the sensors is likely to be within the proprietary purview of the agent both because there is no acceptable existing industry standard, and the desire to integrate existing legacy solutions with the IEEE 11073-20601 framework. Therefore, this responsibility is outside the scope of this standard. Only the communication between the independent living activity hub agent and the manager is covered by this standard.

Additionally, due to the many different types of sensors that may be employed in any particular installation, there is a range of functionality that the corresponding independent living activity hub agent presents to the manager. On the one hand, a fully functional independent living activity hub agent could represent a significant number of sensors and have a complex conversation with the associated manager. On the other hand, a subset of this protocol could be employed by just one sensor such that it would appear to a manager as an independent living activity hub agent with a single sensor.

Figure 1 shows a typical scenario with the type of sensors that might be fitted within a domicile of a person being monitored for their activities of daily living (ADL) and to ensure their safety. Sensors to determine activity and location within the room include passive infrared (PIR) motion, chair and sofa occupancy sensors, and a position sensor worn on the person to provide accurate tracking within the room. Sensors detect when appliances such as the table lamp and television are used. Sensors for safety include a smoke detector and temperature sensor, and a window sensor detects any unexpected opening that might indicate unauthorized entry. Sensors on the door can be used to determine exit and entry and warn of wander. The person also wears a sensor that can detect if they have fallen, and includes a button they can press (PERS) to summon assistance.




Figure 1—Independent living activity hub—typical scenario

In this scenario, sensors, such as PIR motion, temperature, and smoke are wired back to a common independent living activity hub (or use proprietary legacy wireless connection) that monitors each of the sensors and acts as a proxy to send their data and report their status to the manager. Other sensors, such as occupancy and usage, act as independent living activity hub agents that send their data and report their status direct to the manager.

Position is based on a radio beacon system. The corner of the room is used as the origin for the coordinates, and the beacon at this location is designated as the datum base. The location of the base beacon may be reported as the exact geographic location, as a meaningful descriptive location, or no location, as appropriate for an application. The orientation of the room (base angle) may be reported as the angle between geographic north and the Y axis of the coordinate system.

5.3 Collected data

5.3.1 General

This clause provides an overview of the kinds of sensors and activity data that could be collected. This is not to imply that all independent living activity hub agents would necessarily report values for all of these sensors. Furthermore, this standard is not concerned with the form of the data nor the communication between any actual sensor and the independent living activity hub; rather, only the activity data sensor events derived as a result of that data are considered part of this standard. See Clause 6 for the normative definition of this derived data.

5.3.2 Fall sensor

This sensor is used to notify the monitoring system that a fall sensor event has taken place. This could take the form of a sensor of the type that detects a person's fall and automatically generates the event. A pre-fall event can be sent if the sensor detects a stumble.

5.3.3 PERS sensor

The personal emergency response system (PERS) sensor is used to notify the monitoring system that a personal emergency sensor event has taken place. This would typically take the form of a button that the person presses to indicate some sort of perceived emergency (“panic button”).

5.3.4 Environmental alert sensors

These sensors generate a sensor event whenever they sense an environmental aspect that is beyond a preset threshold. Examples include smoke sensors, carbon monoxide sensors, water sensors, and natural gas/liquid propane (LP) gas sensors.

5.3.5 Motion sensor

This sensor generates a sensor event whenever it has sensed movement within its range above a preset level. This type of sensor is typically employed in two manners: general motion and intruder detection.

In the case of general motion, the detection of the motion causes an immediate generation of a sensor event and subsequent action. This may be used for tracking the activity level of the occupant to discern whether behavior patterns have altered.

In the case of intruder detection, the motion sensor event could be used to trigger intruder detection actions.

Optionally, in the case of the primary entrances to the building; for example, the subsequent action to the sensor event may be delayed for some period before taken. This is to allow for an authorized person to enter the premises and to disable the sensor event before the action is taken (e.g., in the case where a triggered sensor event will generate an intruder alert). Should the proper disabling not take place within the expected delay time period, the normal subsequent action would take place. There may also be a sensor event for the case when the motion sensor detects someone attempting to tamper with its function.

5.3.6 Property exit sensor

This sensor generates a sensor event whenever it has sensed the exit of an occupant from the premises. This is commonly employed when there is an occupant with some cognitive issues who would encounter difficulties in an unfamiliar environment. There may also be a sensor event for the case where the premises’ exit is left open.

5.3.7 Enuresis sensor

This sensor generates a sensor event when it detects occurrences of involuntary urination or bed-wetting. This sensor could be utilized in a range of settings such as a bed, chair, or any similar structure.

5.3.8 Contact closure sensor

This sensor issues a sensor event whenever a contact is opened or closed. This sensor reports the state of the contact after a transition, either from closed to open or from open to closed. Only a single sensor event is sent for each transition. Examples of where this sensor would be deployed are passageway doors, cupboard doors, drawers, windows, and pressure mats.

5.3.9 Usage sensor

This sensor issues a sensor event to denote the start of use (into a bed/chair, for example) or the end of use (out of a bed/chair, for example). It also issues a sensor event for an anticipated usage not occurring by an expected preset time (expected use start violation) as well as a sensor event for the usage continuing beyond an expected preset time (expected use stop violation). Additionally, there would be a sensor event generated if during an expected usage time, the usage is discontinued for longer than a preset period of time (intermittent absence violation). An example would be that during a normal sleep period, a person got out of the bed and did not return in an expected period of time.

5.3.10 Switch use sensor

This sensor issues a sensor event for a switch changing states either to the used state (ON) or to the unused state (OFF). Examples of this are light switches, fan switches, and other similar switches that control electrical apparatus.

5.3.11 Simple medication dispenser

The dispenser is a container that contains doses of one or more medications. The medications are to be taken in predetermined doses at predetermined times. A sensor event is generated for a presented dosage being taken from the dispenser (dosage taken) and/or for a dose not being taken after a predetermined amount of time (dosage missed). An event is generated when the dispenser becomes empty of medication and requires to be refilled.

NOTE—The sensor description presented here is of a conceptual model to aid in understanding. Be aware that the same derived sensor events could be generated through many other means (such as a user interaction with some screen interface), and devices may not be able to generate all of these events. This standard is only concerned with the generated sensor events.

5.3.12 Temperature alert sensor

This sensor monitors the temperature in an environment. It issues sensor events based on the sensor value being outside of a preset temperature limit. The sensor events could be that the ambient temperature has risen above a certain level (high threshold) or dropped below a certain level (low threshold), or that the rate-of-change is faster than a predetermined expected rate (rate-of-change). This can be used for detecting conditions such as the temperature of a dwelling being dangerously high/low, or that stove elements have been left on after cooking has been completed.

5.3.13 Relative humidity alert sensor

This sensor monitors the relative humidity in an environment. It issues sensor events based on the sensor value being outside of a preset humidity limit. The sensor events could be that the ambient relative humidity has risen above a certain level (high threshold). This can be used for detecting conditions such as the dangerously high humidity of a dwelling.

5.3.14 Ambient temperature sensor

This sensor monitors the temperature in an environment and reports the current temperature if capable, or the most recent temperature reading when the manager initiates a data request.

5.3.15 Ambient relative humidity sensor

This sensor monitors the humidity in an environment and reports the current relative humidity if capable, or the most recent relative humidity reading when the manager initiates a data request.

5.3.16 Ambient pressure sensor

This sensor monitors the pressure in an environment and reports the current pressure if capable, or the most recent pressure reading when the manager initiates a data request.

5.3.17 Pollen sensor

This sensor monitors the level of pollen in an environment and reports the level if capable, or the most recent pollen level reading when the manager initiates a data request. The pollen sensor may send readings when the level is elevated.

5.3.18 Dust particle sensor

This sensor monitors the level of dust particles in an environment and reports the current dust count if capable or the most recent dust particle count when the manager initiates a data request. The dust particle sensor may send readings when the level is elevated.

5.3.19 Utility accumulated usage sensor

This sensor reports on the total accumulated usage of a utility such electricity, gas, or water.

5.3.20 Wind speed

Wind speed is reported separately as velocity and direction.

5.3.20.1 Wind velocity

This sensor reports on velocity of the wind.

5.3.20.2 Wind direction

This sensor reports on the direction from which the wind is blowing. Expressed in degrees, 0 degrees indicates the wind is from due north.

5.3.21 Utility accumulated usage

These sensors report on the accumulated usage of a utility, including electricity, gas, and water for the household or individual appliance.

5.3.21.1 Electricity accumulated usage sensor

This sensor reports the total accumulated usage of electricity and reports the current accumulated reading if capable or the most recent accumulated reading when the manager initiates a data request. Usage over a period of time is determined by taking successive readings. This sensor could also report readings at predefined intervals.

5.3.21.2 Gas accumulated usage sensor

This sensor reports the total accumulated usage of gas and reports the most reports the current accumulated reading if capable or the most recent accumulated reading when the manager initiates a data request. Usage over a period of time is determined by taking successive readings. This sensor could also report readings at predefined intervals.

5.3.21.3 Water accumulated usage sensor

This sensor reports the total accumulated usage of water and reports the current accumulated reading if capable or the most recent accumulated reading when the manager initiates a data request. Usage over a period of time is determined by taking successive readings. This sensor could also report readings at predefined intervals.

5.3.22 Utility instantaneous usage sensor

These sensors report on the instantaneous usage of a utility, including electricity, gas, and water for the household or individual appliance.

5.3.22.1 Electricity instantaneous usage sensor

This sensor reports the instantaneous usage of electricity and reports the current reading if capable, or the most recent reading when the manager initiates a data request. This sensor could also report readings at predefined intervals or based on predefined trigger events such as usage.

5.3.22.2 Gas instantaneous usage sensor

This sensor reports the instantaneous usage of gas and reports the current reading if capable, or the most recent reading when the manager initiates a data request. This sensor could also report readings at predefined intervals or based on predefined trigger events such as usage.

5.3.22.3 Water instantaneous usage sensor

This sensor reports the instantaneous usage of water and reports the current reading if capable, or the most recent reading when the manager initiates a data request. This sensor could also report readings at predefined intervals or based on predefined trigger events such as usage.

5.3.23 Solar accumulated generation sensor

This sensor reports on the total accumulated generation of energy by photovoltaic cells installed in a location and reports the current accumulated generation reading if capable, or the most recent accumulated reading when the manager initiates a data request. Usage over a period of time is determined by taking successive readings. This sensor could also report readings at predefined intervals.

5.3.24 Solar instantaneous generation sensor

This sensor reports on the instantaneous generation of energy by photovoltaic cells installed in a location and reports the current reading if capable, or the most recent reading when the manager initiates a data request. This sensor could also report readings at predefined intervals or based on predefined trigger events such as usage.

5.3.25 Location

This sensor reports its current location as an enumeration of one of the location codes and can indicate if the sensor is at or not at the location. The sensor may also emit a string that provides a descriptive annotation of the location such as a human readable string or might report the identity of a beacon in a location system.

5.3.26 Relative location coordinates

Location within a defined space, such as a room, building, or site are reported as the Cartesian coordinates $\{X, Y, Z\}$ relative to a datum base coordinate (origin) or specified identified datum base location. If the datum base coordinate is specified, then geographic coordinates are used as the datum base X (latitude), Y (longitude), and Z (altitude) coordinates, together with the angle of the Y axis of the local coordinate space relative to geographic coordinates ($\pm 180^\circ$). Alternatively, the datum base coordinate may be specified as a well-defined location within the defined space together with the angle of the Y axis of the local coordinate space relative to geographic coordinates. The location may be specified as one of the location codes and/or a string to describe the location.

If a datum base coordinate or the angle is not specified, then the relative coordinates are interpreted as appropriate for the location system within the defined space, such as a space defined by a wireless location system.

5.3.26.1 Relative Location coordinates

This sensor reports the $\{X, Y, Z\}$ coordinates of the relative location of a device within a defined space.

5.3.26.2 Base coordinate

This metric object reports the geographic coordinates (latitude, longitude, and altitude) for the datum base coordinate (origin) of the defined space. The location is reported by two metric objects: longitude and latitude, and altitude.

5.3.26.3 Base angle

This metric object reports the angle of the Y axis of the local coordinate space relative to geographic coordinates.

5.3.26.4 Base coordinate location

This metric object reports the datum base coordinate (origin) of the defined space as a well-defined physical location. The location may be specified as one of the location codes and/or a string to describe the location.

5.3.27 GPS location coordinates

This sensor reports the location of the device as the GPS (geographic) coordinates of longitude, latitude, and altitude. In many cases only longitude and latitude are reported, and for this reason GPS location is reported by two metric objects: longitude and latitude, and altitude.

5.3.28 Speed and heading

These sensors report separately the speed and heading of the device.

5.3.28.1 Speed

This sensor reports the absolute speed of the device.

5.3.28.2 Heading

This sensor reports the heading relative to geographic coordinates

5.3.29 Identity

This object reports an opaque string that is used as a locally unique identifier of the person or object associated with other sensors in an independent living activity hub agent. This could include such information as identification from an RFID tag, or patient identity number. If the identity object is present, then those sensors associated with a person shall have the Source-Handle-Reference set to the value of the handle of the identity object.

5.3.30 Encounter

This object reports an opaque string that is used to provide a locally unique identifier of the person or object that has been encountered by this agent. This could include such information as identification from RFID tag, Bluetooth MAC address, or patient identity number. The form of encounter is not specified and could take the form of detection of being in close proximity.

5.4 Reporting data from multiple persons

5.4.1 Identification of a single person

IEEE 11073 contains several mechanisms that permit a device or observation to be associated with a specific person or object. The medical device system (MDS) of every IEEE 11073 device includes a unique extended unique identifier (64 bits) (EUI-64) that may be used to associate a specific device with a single user, and thereby associate all observations from that device with that patient or object.

5.4.2 Identification of multiple persons using multiple person dynamic data update

Where a single device is used to report observations from multiple persons, the multiple person form of the dynamic data update (6.5.3) may be used to report the observations of each individual. A person number is allocated to each individual and observations with that person number as header in the multiple persons dynamic data update are associated with that person. This approach has the disadvantage that the association of a sensor with a person is unknown until an observation is reported and the association of a sensor may be to multiple persons. There is a further disadvantage that the person number must be mapped a-priori to each specific person for whom observations will be reported.

5.4.3 Identification of multiple persons using the identity object

The independent living activity hub may act as proxy for a sensor that is used by people who have previously not been configured in the sensor. This may be the case for entry systems where a card reader is used to identify the individual when they present a card to gain entry to a building. Information regarding which observation is associated with which person is provided by using the identity object. An identity object is instantiated and used to report the locally unique identification as an observation. If the Source-Handle-Reference attribute of a sensor object references the identity object, then the locally unique identity of a person can be associated to an observation from the sensor.

Figure 2 shows how an enumeration object references an associated identity object.




Figure 2—Identification of multiple persons using the identity object

Figure 3 provides an example of the use of the identity object to provide the identify of a person entering though the front door of a property and later leaving by the back door. An enumeration object references the associated identify object by using the value of handle of the identity object in the source reference handle field. The time stamp fields shall have the same value.




Figure 3—Example use of the identity object

6. Independent living activity hub domain information model

6.1 Overview

This clause describes the domain information model of the independent living activity hub.

6.2 Class extensions

In this standard, no class extensions are defined with respect to IEEE Std 11073-20601.

6.3 Object instance diagram

The object instance diagram of the independent living activity hub domain information model, defined for the purposes of this standard, is shown as Figure 4 and Figure 5.

The generic DIM of the independent living activity hub that is presented in Figure 4 and Figure 5 defines all possible data objects. However, it would be expected that most independent living activity hubs would implement only a restricted subset of the data objects. An independent living activity hub shall implement at least one sensor instance.

The objects of the DIM, as shown in Figure 4 and Figure 5, are described in 6.4 to 6.12. This includes the medical device system (MDS) object (see 6.5), the numeric objects (see 6.6), the real-time sample array (RT-SA) objects (see 6.7), the enumeration objects (see 6.8), the PM-store objects (see 6.9), and the scanner objects (see 6.10). See 6.12 for rules for extending the independent living activity hub information model beyond elements as described in this standard. Each clause that describes an object of the independent living activity hub contains the following information:

- The nomenclature code used to identify the class of the object. One example of where this code is used is the configuration event, where the object class is reported for each object. This allows the manager to determine whether the class of the object being specified is a numeric, real-time sample array, enumeration, scanner, or PM-store class.
- The attributes of the object. Each object has attributes that represent and convey information on the activity data generating sensor and its data sources. Each object has a Handle attribute that identifies the object instance within an agent. Attribute values are accessed and modified using communication services such as GET and SET. Attribute types are defined using ASN.1. The ASN.1 definitions for new attribute types specific to this standard are in Annex B, and the ASN.1 definitions for existing attribute types referenced in this standard are in IEEE Std 11073-20601.
- The methods available on the object.
- The potential events generated by the object. Data are sent to the manager using events.
- The available services such as getting or setting attributes.

The attributes for each class are defined in tables that specify the name of the attribute, its value, and its qualifier. The qualifiers mean M—Attribute is Mandatory, C—Attribute is Conditional and depends on the condition stated in the Remark or Value column (if IEEE Std 11073-20601 is referenced, then it contains the conditions), R—Attribute is Recommended, NR—Attribute is Not Recommended, and O—Attribute is Optional. Mandatory attributes shall be implemented by the agent. Conditional attributes shall be implemented if the condition applies and may be implemented otherwise. Recommended attributes should be implemented by the agent. Not recommended attributes should not be implemented by the agent. Optional attributes may be implemented by the agent. For attributes with qualifiers set to R or NR, underlying requirements stated in the Remark and Value column in IEEE Std 11073-20601 shall be followed. If any attribute (from the DIM of IEEE Std 11073-20601) is not included in the definition of that object in this standard, it shall not be included in that object by an implementation, unless it is a vendor-specific attribute extended according to 6.12.

An attribute is further qualified as static, dynamic, or observational. Static attributes shall not change value during the life of an association. Dynamic attributes have a value that may change during the life of an association. The dynamic attribute value should be sent at configuration time and shall be sent at or before the time when the value would be needed for interpreting a reported observation. Observational attributes have a value that may change during the life of an association. When a set of observational attribute values are received, these values are combined with the available context information (i.e., all related dynamic and static attribute values) to represent the observation at the observation time.




Figure 4—Independent living activity hub—object instances




Figure 5—Independent living activity hub—object instances

6.4 Types of configuration

6.4.1 General

As specified in IEEE Std 11073-20601, there are two styles of configuration available. Subclauses 6.4.2 and 6.4.3 briefly introduce standard and extended configurations.

6.4.2 Standard configuration

This standard does not define any standard configurations since the set of sensors for each agent configuration is likely to vary significantly for each deployment scenario. Therefore, all configurations shall be specified as extended configurations.

6.4.3 Extended configuration

In extended configurations, the agent's configuration is not predefined in a standard. The agent determines which objects, attributes, and values that it wants to use in a configuration and assigns a configuration identifier. When the agent associates with a manager, it negotiates an acceptable configuration. Typically, the manager does not recognize the agent's configuration on the first connection, so the manager responds that the agent must send its configuration information as a configuration event report. However, if the manager already understands the configuration, either because it was preloaded in some way or the agent had previously associated with the manager, then the manager responds that the configuration is known and no further configuration information needs to be sent.

6.5 Medical device system object

6.5.1 MDS object attributes

Table 1 summarizes the attributes of the independent living activity hub MDS object. The nomenclature code to identify the MDS class is MDC_MOC_VMS_MDS_SIMP.

Table 1—MDS object attributes

Attribute name	Value	Qual.
Handle	0	M
System-Type-Spec-List	{MDC_DEV_SPEC_PROFILE_AI_ACTIVITY_HUB,2}. If an agent follows a profile within this specialization this profile shall be specified in the System-Type-Spec-List.	M
System-Type	Attribute not present. See IEEE Std 11073-20601.	NR
System-Model	{“Manufacturer”, “Model”}.	M
System-Id	EUI-64.	M
Dev-Configuration-Id	Extended configs: 0x4000–0x7FFF.	M
Attribute-Value-Map	See IEEE Std 11073-20601.	C
Production-Specification	See IEEE Std 11073-20601.	O
Mds-Time-Info	See IEEE Std 11073-20601.	C
Date-and-Time	See IEEE Std 11073-20601.	C
Base-Offset-Time	See IEEE Std 11073-20601.	R
Relative-Time	See IEEE Std 11073-20601.	NR
HiRes-Relative-Time	See IEEE Std 11073-20601.	O
Date-and-Time-Adjustment	See IEEE Std 11073-20601.	C
Power-Status	<i>onBattery</i> or <i>onMains</i> .	R
Battery-Level	See IEEE Std 11073-20601.	R
Remaining-Battery-Time	See IEEE Std 11073-20601.	R
Reg-Cert-Data-List	See IEEE Std 11073-20601.	O
Confirm-Timeout	See IEEE Std 11073-20601.	O
Transport-Timeout	See IEEE Std 11073-20601.	O
NOTE—See IEEE Std 11073-20601 for information on whether an attribute is static, or dynamic.		

NOTE—See 6.3 for a description of the qualifiers.

In the response to a Get MDS object command, only implemented attributes and their corresponding values are returned.

See IEEE Std 11073-20601 for descriptive explanations of the individual attributes as well as for information on attribute ID and attribute type.

The Dev-Configuration-Id attribute holds a locally unique 16-bit identifier that identifies the device configuration. For an independent living activity hub agent with extended configuration, this identifier is chosen in the range of extended-config-start to extended-config-end (see IEEE Std 11073-20601) as shown in Table 1.

The agent sends the Dev-Configuration-Id during the Associating state (see 8.3) to identify its configuration for the duration of the association. If the manager already holds the configuration information relating to the Dev-Configuration-Id, it recognizes the Dev-Configuration-Id. Then the Configuring state (see 8.4) is skipped; the agent and manager then enter the Operating state. If the manager does not recognize the Dev-Configuration-Id, the agent and manager enter the Configuring state.

If an agent implements multiple IEEE 11073-104zz specializations, System-Type-Spec-List is a list of type/version pairs, each referencing the respective device standard and version of that standard.

6.5.2 MDS object methods (actions)

Table 2 defines the methods (actions) of the MDS object. These methods are invoked using the Action service. In Table 2, the Subservice type name column defines the name of the method. The Mode column defines whether the method is invoked as an unconfirmed action (i.e., roiv-cmip-action from IEEE Std 11073-20601) or a confirmed action (i.e., roiv-cmip-confirmed-action). The Subservice type (action-type) column defines the nomenclature code used in the action-type field of an action request and response (see IEEE Std 11073-20601). The Parameters (action-info-args) column defines the associated ASN.1 data structure (see IEEE Std 11073-20601 for ASN.1 definitions) to use in the action message for the action-info-args field of the request. The Results (action-info-args) column defines the structure to use in the action-info-args of the response.

Table 2—MDS object methods (actions)

Service	Subservice type name	Mode	Subservice type (action-type)	Parameters (action-info-args)	Results (action-info-args)
ACTION	Set-Time	Confirmed	MDC_ACT_SET_TIME	SetTimeInvoke	—
	Set-Base-Offset-Time	Confirmed	MDC_ACT_SET_BO_TIME	SetBOTimeInvoke	—
	MDS-Data-Request	Confirmed	MDC_ACT_DATA_REQUEST	DataRequest	DataResponse

— **Set-Time:**

This method allows the manager to set a real-time clock in the agent with the absolute time. The agent indicates whether the Set-Time command is valid using the mds-time-capab-set-clock bit in the Mds-Time-Info attribute (see IEEE Std 11073-20601). Agents with an internal real-time clock shall indicate this capability by also setting the mds-time-capab-real-time-clock bit in the Mds-Time-Info attribute. The Set-Time method can only be supported if the Absolute-Time-Stamp attribute is supported.

— **Set-Base-Offset-Time:**

This method allows the manager to set a real-time clock in the agent with the base time and offset. The agent indicates whether the Set-Base-Offset-Time command is valid using the mds-time-capab-set-clock bit in the Mds-Time-Info attribute (see IEEE Std 11073-20601). The Set-Base-Offset-Time method can only be supported if the Base-Offset-Time-Stamp attribute is supported.

— **MDS-Data-Request:**

This method allows the manager system to request measurement data from an agent (see IEEE Std 11073-20601 for a description).

Agents following only this device specialization and no others shall send event reports (see 6.5.3) using agent-initiated measurement and/or manager-initiated measurement data transmission. Agents following this device specialization shall send event reports or request data in the appropriate fashion. During the association procedure (see 8.3), DataReqModeCapab shall be set to the appropriate value for the event report or data request style. As a result, the manager shall assume that if the

independent living activity hub agent supports MDS-Data-Request, then it may access the object value only if the object's Metric-Spec-Small attribute has its mss-acc-manager-initiated bit set (see IEEE Std 11073-20601 for additional information).

6.5.3 MDS object events

Table 3 defines the events sent by the independent living activity hub MDS object:

Table 3—Independent living activity hub MDS object events

Service	Subservice type name	Mode	Subservice type (event-type)	Parameters (event-info)	Results (event-reply-info)
EVENT REPORT	MDS-Configuration-Event	Confirmed	MDC_NOTI_CONFIG	ConfigReport	ConfigReportRsp
	MDS-Dynamic-Data-Update-Var	Confirmed	MDC_NOTI_SCAN_REP ORT_VAR	ScanReportInfoVar	—
	MDS-Dynamic-Data-Update-Fixed	Confirmed	MDC_NOTI_SCAN_REP ORT_FIXED	ScanReportInfoFixed	—
	MDS-Dynamic-Data-Update-MP-Var	Confirmed	MDC_NOTI_SCAN_REP ORT_MP_VAR	ScanReportInfoMPVar	—
	MDS-Dynamic-Data-Update-MP-Fixed	Confirmed	MDC_NOTI_SCAN_REP ORT_MP_FIXED	ScanReportInfoMPFixed	—

— **MDS-Configuration-Event:**

This event is sent by the independent living activity hub agent during the configuring procedure if the manager does not already know the independent living activity hub agent's configuration from past associations or because the manager has not been implemented to recognize the configuration according to the independent living activity hub device standard. The event provides static information about the supported measurement capabilities of the independent living activity hub agent.

— **MDS-Dynamic-Data-Update-Var:**

This event provides dynamic measurement data from the independent living activity hub agent for the independent living activity data objects. These data are reported using a generic attribute list variable format. The event is sent as an unsolicited message by the agent (i.e., an agent-initiated measurement data transmission). See 8.5.3 for more information on unsolicited event reporting.

— **MDS-Dynamic-Data-Update-Fixed:**

This event provides dynamic measurement data from the independent living activity hub agent for the independent living activity data objects. These data are reported in the fixed format defined by the Attribute-Value-Map attribute of the object(s). The event is sent as an unsolicited message by the agent (i.e., an agent-initiated measurement data transmission). See 8.5.3 for more information on unsolicited event reporting.

— **MDS-Dynamic-Data-Update-MP-Var:**

This is the same as MDS-Dynamic-Data-Update-Var, but allows inclusion of data from multiple people.

— **MDS-Dynamic-Data-Update-MP-Fixed:**

This is the same as MDS-Dynamic-Data-Update-Fixed, but allows inclusion of data from multiple people.

NOTE—IEEE Std 11073-20601 requires that managers support all of the MDS object events listed above.

6.5.4 Other MDS services

6.5.4.1 GET service

An independent living activity hub agent shall support the GET service to retrieve the values of all implemented attributes of the MDS object. The GET service can only be invoked after the manager has confirmed selection of the configuration of the Agent.

The GET request for all attributes shall be supported. An attribute-id-list parameter may be supported.

The manager may request the MDS object attributes of the independent living activity hub agent; in which case, the manager shall send the “Remote Operation Invoke | Get” message (see roiv-cmip-get in IEEE Std 11073-20601) with the reserved MDS handle value of 0. The independent living activity hub agent shall report its MDS object attributes to the manager using the “Remote Operation Response | Get” message (see rors-cmip-get in IEEE Std 11073-20601). See Table 4 for a summary of the GET service including some message fields.

Table 4—Independent living activity hub MDS object GET service

Service	Subservice type name	Mode	Subservice type	Parameters	Results
GET	<na>	<implied confirmed>	<na>	GetArgumentSimple=(obj-handle=0), attribute-id-list <optional>	GetResultSimple=(obj-handle=0), attribute-list

See 8.5.2 for details on the procedure for getting the MDS object attributes.

6.5.4.2 SET service

The independent living activity hub standard does not require an implementation to support the MDS object SET service.

6.6 Numeric objects

6.6.1 General

The independent living activity hub requires one numeric data object for each supported sensor instance. Objects instantiated from the numeric class are used to report the environmental data. The nomenclature code to identify the enumeration class is MDC_MOC_VMO_METRIC_NU. The attributes of a numeric data object are shown in Table 5. The attribute structure shown in Table 5 is common to all sensor types. Later clauses define the precise definitions for each sensor type.

Sometimes, the interpretation of one attribute value in an object depends on other attribute values in the same object. For example, Unit-Code and Unit-LabelString provide context for the observed values. Whenever a contextual attribute changes, the agent shall report these changes to the manager using an MDS object event (see 6.5.3) prior to reporting any of the dependent values.

Table 5—Activity data numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601.	M
Type	Any of the MDC_PART_PHD_AI, MDC_AI_TYPE_* ID's. See Annex C.	M

Attribute name	Extended configuration	
Supplemental-Types	The value that denotes the sensor location. This location is a 16-bit value constructed where the high 11 bits are one of the MDC_AI_LOCATION family of constants as denoted in Annex C, and the lower 5 bits are the unique instance of that location. For example, the high 11 bits could be the value for bedroom, and the lower 5 bits could be the value for bedroom 3 (1 being the first bedroom). As a list, this allows for a hierarchical location, such as {building_3, bedroom_5, chair_2}. All of these values are from the Partition ID: MDC_PART_PHD_AI. If the sensor reports general household values or is not in a static location (e.g., GPS location of a person), then its use is optional. If reporting values for a specific location or appliance, then its use is mandatory.	C
Metric-Spec-Small	mss-avail-intermittent mss-avail-stored-data mss-upd-aperiodic mss-msmt-aperiodic mss-acc-agent-initiated mss-acc-manager-initiated mss-acc-manager-initiated-immediate (See note)	M
Metric-Structure-Small	See IEEE Std 11073-20601.	NR
Measurement-Status	See IEEE Std 11073-20601.	C
Metric-Id	See IEEE Std 11073-20601.	NR
Metric-Id-List	See IEEE Std 11073-20601.	C
Metric-Id-Partition	See IEEE Std 11073-20601.	C
Unit-Code	See IEEE Std 11073-20601.	C
Attribute-Value-Map	See IEEE Std 11073-20601.	C
Source-Handle-Reference	The value references the handle of the identity object to which the sensor is associated if applicable.	C
Label-String	See IEEE Std 11073-20601.	O
Unit-LabelString	See IEEE Std 11073-20601.	O
Absolute-Time-Stamp	See IEEE Std 11073-20601.	C
Base-Offset-Time-Stamp	See IEEE Std 11073-20601.	C
Relative-Time-Stamp	See IEEE Std 11073-20601.	NR
HiRes-Time-Stamp	See IEEE Std 11073-20601.	C
Measure-Active-Period	See IEEE Std 11073-20601.	NR
Simple-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Compound-Simple-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Compound-Basic-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Nu-Observed-Value	See IEEE Std 11073-20601.	C
Compound-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Accuracy	See IEEE Std 11073-20601.	O
Context-Id	See IEEE Std 11073-20601.	NR
NOTE 1—See IEEE Std 11073-20601 for information on whether an attribute is static, dynamic, or observational. NOTE 2—The settings for Metric-Spec-Small with respect to agent-initiated data transfer and manager initiated request for mss-acc-agent-initiated mss-acc-manager-initiated mss-acc-manager-initiated-immediate and associated settings will depend on implementation and explicit use of the sensor.		

NOTE—See 6.3 for a description of the qualifiers.

The numeric object does not support any methods, events, or other services.

See IEEE Std 11073-20601 for descriptive explanations on the individual attributes as well as for information on attribute ID and attribute type.

6.6.2 Ambient temperature

Table 6 summarizes the attributes of the ambient temperature numeric object.

Table 6—Ambient temperature numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601.	M
Type	{MDC_PART_SCADA, MDC_TEMP_ROOM }	M
Supplemental-Types	The value that denotes the sensor location. This location is a 16-bit value constructed where the high 11 bits are one of the MDC_AI_LOCATION family of constants as denoted in Annex C, and the lower 5 bits are the unique instance of that location. For example, the high 11 bits could be the value for bedroom, and the lower 5 bits could be the value for bedroom 3 (1 being the first bedroom). All of these values are from the Partition ID: MDC_PART_PHD_AI.	C
Metric-Spec-Small	mss-avail-intermittent mss-avail-stored-data mss-upd-aperiodic mss-msmt-aperiodic mss-acc-agent-initiated mss-acc-manager-initiated mss-acc-manager-initiated-immediate	M
Unit-Code	MDC_DIM_DEGC or MDC_DIM_FAHR	M
Absolute-Time-Stamp	See IEEE Std 11073-20601.	C
Base-Offset-Time-Stamp	See IEEE Std 11073-20601.	C
HiRes-Time-Stamp	See IEEE Std 11073-20601.	C
Simple-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Nu-Observed-Value	See IEEE Std 11073-20601.	C

6.6.3 Ambient relative humidity

Table 7 summarizes the attributes of the current humidity numeric object.

Table 7—Ambient relative humidity numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601.	M
Type	{MDC_PART_SCADA, MDC_REL_HUMIDITY_AMBIENT }	M

Attribute name	Extended configuration	
Supplemental-Types	The value that denotes the sensor location. This location is a 16-bit value constructed where the high 11 bits are one of the MDC_AI_LOCATION family of constants as denoted in Annex C, and the lower 5 bits are the unique instance of that location. For example, the high 11 bits could be the value for bedroom, and the lower 5 bits could be the value for bedroom 3 (1 being the first bedroom). All of these values are from the Partition ID: MDC_PART_PHD_AI.	C
Metric-Spec-Small	mss-avail-intermittent mss-avail-stored-data mss-upd-aperiodic mss-msmt-aperiodic mss-acc-agent-initiated mss-acc-manager-initiated mss-acc-manager-initiated-immediate	M
Unit-Code	MDC_DIM_PERCENT	M
Absolute-Time-Stamp	See IEEE Std 11073-20601.	C
Base-Offset-Time-Stamp	See IEEE Std 11073-20601.	C
HiRes-Time-Stamp	See IEEE Std 11073-20601.	C
Simple-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Nu-Observed-Value	See IEEE Std 11073-20601.	C

6.6.4 Ambient pressure

Table 8 summarizes the attributes of the current pressure numeric object.

Table 8—Ambient pressure numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601.	M
Type	{MDC_PART_SCADA, MDC_PRESS_AIR_AMBIENT}	M
Supplemental-Types	The value that denotes the sensor location. This location is a 16-bit value constructed where the high 11 bits are one of the MDC_AI_LOCATION family of constants as denoted in Annex C, and the lower 5 bits are the unique instance of that location. For example, the high 11 bits could be the value for bedroom, and the lower 5 bits could be the value for bedroom 3 (1 being the first bedroom). All of these values are from the Partition ID: MDC_PART_PHD_AI.	C
Metric-Spec-Small	mss-avail-intermittent mss-avail-stored-data mss-upd-aperiodic mss-msmt-aperiodic mss-acc-agent-initiated mss-acc-manager-initiated mss-acc-manager-initiated-immediate	M
Unit-Code	MDC_DIM_KILO_PASCAL, MDC_DIM_BAR, MDC_DIM_MILLI_BAR, MDC_DIM_MMHG, MDC_DIM_INHG, MDC_DIM_PSI.	M
Absolute-Time-Stamp	See IEEE Std 11073-20601.	C
Base-Offset-Time-Stamp	See IEEE Std 11073-20601.	C
HiRes-Time-Stamp	See IEEE Std 11073-20601.	C
Simple-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Nu-Observed-Value	See IEEE Std 11073-20601.	C

6.6.5 Pollen

Table 9 summarizes the attributes of the current pollen numeric object.

Table 9—Pollen numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.

Attribute name	Extended configuration	
Handle	See IEEE Std 11073-20601.	M
Type	{MDC_PART_SCADA, MDC_LEVEL_POLLEN}	M
Supplemental-Types	The value that denotes the sensor location. This location is a 16-bit value constructed where the high 11 bits are one of the MDC_AI_LOCATION family of constants as denoted in Annex C, and the lower 5 bits are the unique instance of that location. For example, the high 11 bits could be the value for bedroom, and the lower 5 bits could be the value for bedroom 3 (1 being the first bedroom). All of these values are from the Partition ID: MDC_PART_PHD_AI.	C
Metric-Spec-Small	mss-avail-intermittent mss-avail-stored-data mss-upd-aperiodic mss-msmt-aperiodic mss-acc-agent-initiated mss-acc-manager-initiated mss-acc-manager-initiated-immediate	M
Unit-Code	MDC_DIM_PER_DECI_L, MDC_DIM_MILLI_G_PER_M_CUBE, MDC_DIM_MICRO_G_PER_M_CUBE	M
Absolute-Time-Stamp	See IEEE Std 11073-20601.	C
Base-Offset-Time-Stamp	See IEEE Std 11073-20601.	C
HiRes-Time-Stamp	See IEEE Std 11073-20601.	C
Simple-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Nu-Observed-Value	See IEEE Std 11073-20601.	C

6.6.6 Dust Count

Table 10 summarizes the attributes of the current dust count numeric object.

Table 10—Dust count numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601.	M
Type	{MDC_PART_SCADA, MDC_LEVEL_DUST}	M
Supplemental-Types	The value that denotes the sensor location. This location is a 16-bit value constructed where the high 11 bits are one of the MDC_AI_LOCATION family of constants as denoted in Annex C, and the lower 5 bits are the unique instance of that location. For example, the high 11 bits could be the value for bedroom, and the lower 5 bits could be the value for bedroom 3 (1 being the first bedroom). All of these values are from the Partition ID: MDC_PART_PHD_AI.	C
Metric-Spec-Small	mss-avail-intermittent mss-avail-stored-data mss-upd-aperiodic mss-msmt-aperiodic mss-acc-agent-initiated mss-acc-manager-initiated mss-acc-manager-initiated-immediate	M
Unit-Code	MDC_DIM_PER_DECI_L, MDC_DIM_MILLI_G_PER_M_CUBE, MDC_DIM_MICRO_G_PER_M_CUBE	M
Absolute-Time-Stamp	See IEEE Std 11073-20601.	C
Base-Offset-Time-Stamp	See IEEE Std 11073-20601.	C
HiRes-Time-Stamp	See IEEE Std 11073-20601.	C
Simple-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Nu-Observed-Value	See IEEE Std 11073-20601.	C

6.6.7 Wind velocity

Table 11 summarizes the attributes of the current wind velocity numeric object.

Table 11—Wind velocity numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601.	M
Type	{ MDC_PART_SCADA, MDC_WIND_VELOCITY }	M
Supplemental-Types	The value that denotes the sensor location. This location is a 16-bit value constructed where the high 11 bits are one of the MDC_AI_LOCATION family of constants as denoted in Annex C, and the lower 5 bits are the unique instance of that location. For example, the high 11 bits could be the value for bedroom, and the lower 5 bits could be the value for bedroom 3 (1 being the first bedroom). All of these values are from the Partition ID: MDC_PART_PHD_AI.	C
Metric-Spec-Small	mss-avail-intermittent mss-avail-stored-data mss-upd-aperiodic mss-msmt-aperiodic mss-acc-agent-initiated mss-acc-manager-initiated mss-acc-manager-initiated-immediate	M
Unit-Code	MDC_DIM_M_PER_SEC, MDC_DIM_KILO_M_PER_HR MDC_DIM_MPH	M
Absolute-Time-Stamp	See IEEE Std 11073-20601.	C
Base-Offset-Time-Stamp	See IEEE Std 11073-20601.	C
HiRes-Time-Stamp	See IEEE Std 11073-20601.	C
Simple-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Nu-Observed-Value	See IEEE Std 11073-20601.	C

6.6.8 Wind direction

Table 12 summarizes the attributes of the current wind direction numeric object.

Table 12—Wind direction numeric object attributes

Attribute name	Extended configuration	
	Value	Qual.
Handle	See IEEE Std 11073-20601.	M
Type	{ MDC_PART_SCADA, MDC_WIND_DIRECTION }	M
Supplemental-Types	The value that denotes the sensor location. This location is a 16-bit value constructed where the high 11 bits are one of the MDC_AI_LOCATION family of constants as denoted in Annex C, and the lower 5 bits are the unique instance of that location. For example, the high 11 bits could be the value for bedroom, and the lower 5 bits could be the value for bedroom 3 (1 being the first bedroom). All of these values are from the Partition ID: MDC_PART_PHD_AI.	C
Metric-Spec-Small	mss-avail-intermittent mss-avail-stored-data mss-upd-aperiodic mss-msmt-aperiodic mss-acc-agent-initiated mss-acc-manager-initiated mss-acc-manager-initiated-immediate	M
Unit-Code	MDC_DIM_ANG_DEG	M
Absolute-Time-Stamp	See IEEE Std 11073-20601.	C
Base-Offset-Time-Stamp	See IEEE Std 11073-20601.	C
HiRes-Time-Stamp	See IEEE Std 11073-20601.	C
Simple-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Basic-Nu-Observed-Value	See IEEE Std 11073-20601.	C
Nu-Observed-Value	See IEEE Std 11073-20601.	C