


IEEE Recommended Practice for Measurement of Field Intensity Above 300 MC from Radio-Frequency Industrial, Scientific, and Medical Equipments



IEEE Std 139-1952

Documents

Document Sharing Hub



Published by The Institute of Electrical and Electronics Engineers, Inc., 345 East 47th Street, New York, NY 10017, USA

Authorized licensed use limited to: Tsinghua University. Downloaded on February 24, 2026 at 15:06:06 UTC from IEEE Xplore. Restrictions apply.

Recommended Practice
for
MEASUREMENT *of* FIELD INTENSITY ABOVE 300 MC
***from* RADIO-FREQUENCY INDUSTRIAL, SCIENTIFIC,**
***and* MEDICAL EQUIPMENTS**

**THIS IS NOT AN AIEE STANDARD AND THE RECOMMENDATIONS
ARE NOT INTENDED TO BE MANDATORY OR RESTRICTIVE**



**Report of the Subcommittee on
Radiation Measurements Above 300 MC
of the AIEE Electric Heating Committee**

Published by
AMERICAN INSTITUTE OF ELECTRICAL ENGINEERS
33 West Thirty-ninth Street, New York 18, N. Y.

IEEE Standards documents are developed within the Technical Committees of the IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Board. Members of the committees serve voluntarily and without compensation. They are not necessarily members of the Institute. The standards developed within IEEE represent a consensus of the broad expertise on the subject within the Institute as well as those activities outside of IEEE which have expressed an interest in participating in the development of the standard.

Use of an IEEE Standard is wholly voluntary. The existence of an IEEE Standard does not imply that there are no other ways to produce, test, measure, purchase, market, or provide other goods and services related to the scope of the IEEE Standard. Furthermore, the viewpoint expressed at the time a standard is approved and issued is subject to change brought about through developments in the state of the art and comments received from users of the standard. Every IEEE Standard is subjected to review at least once every five years for revision or reaffirmation. When a document is more than five years old, and has not been reaffirmed, it is reasonable to conclude that its contents, although still of some value, do not wholly reflect the present state of the art. Users are cautioned to check to determine that they have the latest edition of any IEEE Standard.

Comments for revision of IEEE Standards are welcome from any interested party, regardless of membership affiliation with IEEE. Suggestions for changes in documents should be in the form of a proposed change of text, together with appropriate supporting comments.

Interpretations: Occasionally questions may arise regarding the meaning of portions of standards as they relate to specific applications. When the need for interpretations is brought to the attention of IEEE, the Institute will initiate action to prepare appropriate responses. Since IEEE Standards represent a consensus of all concerned interests, it is important to ensure that any interpretation has also received the concurrence of a balance of interests. For this reason IEEE and the members of its technical committees are not able to provide an instant response to interpretation requests except in those cases where the matter has previously received formal consideration.

Comments on standards and requests for interpretations should be addressed to:

Secretary, IEEE Standards Board
345 East 47th Street
New York, NY 10017
USA

CONTENTS

I	Introduction	5
A.	Need for Standard Measuring Method	5
B.	FCC Requirements	5
C.	Committee Action	5
II	Symbols, Units and Abbreviations	5
III	Methods of Measurement	
A.	General	6
B.	Standard-Antenna Method	6
C.	Standard-Field Method	6
IV	Antenna Design	
A.	General	7
B.	Half-Wave-Dipole Design	7
C.	Corner-Reflector Design	8
E.	Impedance Matching and Bandwidth	9
D.	Horn Antenna Design	9
V	Antenna Gain Measurements	
A.	General	10
B.	Direct Measurement Method	10
C.	Reflection Method	11
D.	Comparison Method	11
VI	Associated Measuring Equipment	
A.	General	11
B.	Conventional Power Meters	11
C.	Receivers	12
D.	Waveguide, Cables, and Connectors	12
E.	Tuners and Standing-Wave Indicators	12
VII	Measurements on Production Equipment	
A.	General	12
B.	Preliminary Measurement Procedure	12
C.	Final Measurement Procedure	13
VIII	Measurement Precautions and Factors Affecting Accuracy of Measurement	
A.	Precautions	13
B.	Factors Affecting the Accuracy of Measurement	13
Appendix	References	14

ACKNOWLEDGMENT

The Institute wishes to acknowledge its indebtedness to those who have so freely given their time and knowledge and have conducted experimental work on which many of the AIEE publications are based.

The personnel of the Subcommittee on Radiation Measurements Above 300 Mc, a Subcommittee of the AIEE Committee on Electric Heating, which prepared this publication is as follows:

R. M. Baker, *Chairman*

H. H. Beizer

Paul Hines

Marvin Bock

G. E. Leavitt (alternate)

E. W. Chapin

W. L. Roberts

H. E. Dinger

R. M. Showers

J. E. Eiselein

H. E. Sorrows

C. W. Frick

A. J. Vadasz