

INTERNATIONAL STANDARD

**ISO
10647**

First edition
1996-12-15

Procedures for calibrating and determining the response of neutron-measuring devices used for radiation protection purposes

*Méthodes d'étalonnage et de détermination de la réponse des
instruments de mesure des neutrons utilisés en radioprotection*

This material is reproduced from ISO documents under International Organization for Standardization (ISO) Copyright License number IHS/ICC/1996. Not for resale. No part of these ISO documents may be reproduced in any form, electronic retrieval system or otherwise, except as allowed in the copyright law of the country of use, or with the prior written consent of ISO (Case postale 56, 1211 Geneva 20, Switzerland, Fax +41 22 734 10 79), IHS or the ISO Licensor's members.



Reference number
ISO 10647:1996(E)

AVA Documents
Document Sharing Hub

Contents

Page

1	Scope.....	1
2	Normative references.....	1
3	Devices covered in this International Standard	2
3.1	Individual dosimeters	2
3.2	Devices for surveying and area monitoring.....	2
4	Definitions	2
4.1	Metrological terms	3
4.1.1	Reading	3
4.1.2	Response	3
4.1.3	Calibration factor	3
4.1.4	Energy dependence of response.....	3
4.1.5	Angular dependence of response	3
4.1.6	Gamma-ray sensitivity	3
4.1.7	Free-field quantities	3
4.2	Reference and monitoring instruments	3
4.2.1	Primary standard instrument	3
4.2.2	Secondary standard instrument.....	3
4.2.3	Transfer instrument	3
4.2.4	Monitoring instrument	3
4.3	Devices to be tested	4
4.3.1	Neutron dosimeter.....	4
4.3.2	Neutron dose-ratemeter	4
4.3.3	Neutron detector.....	4
5	Symbols (see annex G).....	4
6	Traceability of the calibration of the reference radiation field.....	4
6.1	Traceability for radionuclide neutron sources.....	4

© ISO 1996

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

6.2	Traceability for neutron produced by an accelerator	5
6.3	Traceability for reactor neutron beams.....	5
7	Calibration principles for radionuclide neutron sources	5
7.1	General principles	5
7.2	Important features of a neutron calibration facility.....	6
7.2.1	Source	6
7.2.2	Irradiation set-up	6
7.2.3	Irradiation room	6
7.3	Sources of scattered neutrons	6
7.3.1	Room-scatter.....	7
7.3.2	Air-attenuation (air-outscatter).....	7
7.3.3	Room-inscatter.....	7
7.3.4	Scattering from support structures.....	7
7.3.5	Spectral effects	7
7.4	Effective point of measurement and effective calibration distance	7
7.4.1	Spherical devices	7
7.4.2	Cylindrical devices.....	7
7.4.3	Irradiation of individual dosimeters upon a phantom.....	7
7.5	Effect of photon radiation.....	8
8	Correcting for scattering effects for radionuclide neutron sources	8
8.1	Initial measurements.....	8
8.2	Geometry correction	8
8.3	Analysis of data	9
8.3.1	Semi-empirical method	9
8.3.2	Shadow-cone technique.....	9
8.3.3	Polynomial fit method	10
8.4	Choice of methods	10
8.4.1	Semi-empirical method	10
8.4.2	Shadow-cone method	10
8.4.3	Polynomial fit method	11
9	Routine calibrations with radionuclide neutron sources	11
9.1	Linearity determination	11
9.2	Determination of the angular dependence of response.....	12
10	Estimated uncertainties	12
10.1	Components of the uncertainty applicable to radionuclide source calibrations.....	12
10.1.1	Uncertainty in neutron source strength, B	12

10.1.2	Uncertainty in anisotropy function, $F_1(\theta)$	12
10.1.3	Uncertainty in calibration distance, l	12
10.1.4	Uncertainty in geometry-correction factor $F_1(l)$	12
10.1.5	Scattering correction uncertainty	12
10.1.6	Statistical uncertainty in readings, M	13
10.1.7	Timing uncertainty	13
10.1.8	Uncertainty in spectrum-averaged values of fluence to-dose equivalent conversion	13
11	Calibrations using accelerators and reactors	13
11.1	Introduction	13
11.2	Accelerator-produced neutrons	14
11.2.1	Neutron fluence rate	14
11.2.2	Monitoring	14
11.2.3	Energy dependence of response	14
11.2.4	Contaminant sources of neutrons	14
11.2.5	Neutron energy and energy spread	14
11.2.6	Variation of neutron spectrum with emission angle	14
11.2.7	Target-scatter corrections	14
11.2.8	Effects of neutron scattering	15
11.3	Reactor neutron beams	15
11.3.1	Filtered neutron beams	15
11.3.2	Thermal neutrons	15
11.4	Estimated uncertainties for accelerator and reactors	15
11.4.1	Neutron energy and energy spread	16
11.4.2	Variation of neutron spectrum with emission angle	16
11.4.3	Target-scatter corrections	16
11.4.4	Effects of neutron scattering	16
11.4.5	Fluence determination	16
11.5	Combining uncertainties	16
Annexes		
A	Scheme of physical characteristics to be tested	17
B	Recommendations on phantoms	18
C	Angular source strength characteristics of two radionuclide neutron sources	19
D	Minimum room length for 40% room return	20
E	Air-attenuation correction factors and total air-scatter correction	21
F	Criteria for construction and use of shadow cones	23

G	List of symbols used in ISO 8529.....	24
H	Bibliography	25

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 10647 was prepared by Technical Committee ISO/TC 85, *Nuclear energy*, Subcommittee SC 2, *Radiation protection*.

Annexes A to F form an integral part of this International Standard. Annexes G and H are for information only.

Procedures for calibrating and determining the response of neutron-measuring devices used for radiation protection purposes

1 Scope

This International Standard specifies procedures for the calibration of neutron-measuring devices used for radiation protection purposes, and for determining their response as a function of energy, angle of incidence and dose equivalent rate, using the neutron reference radiations specified in ISO 8529.

Since, according to ICRU Report 39 and ICRU Report 43, neutrons of all energies are strongly penetrating, area monitors should be calibrated in terms of ambient dose equivalent [$H^*(10)$], and individual dosimeters should be calibrated to measure individual dose equivalent, penetrating [$H_p(10)$]. The procedures given in this International Standard are, however, quite general, and may be used with any system of dose equivalent quantities. A diagrammatic scheme of the physical characteristics to be tested is given as annex A.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 8529:1989, *Neutron reference radiations for calibrating neutron-measuring devices used for radiation protection purposes and for determining their response as a function of neutron energy*.

IEC 1005, 1990: *Portable neutron ambient dose equivalent rate-meters for use in radiation protection*.

ICRP Publication 21 (1973): *Data for Protection Against Ionizing Radiation from External Sources*. (Supplement to ICRP Publication 15.)

ICRU Report 13 (1969): *Neutron Fluence, Neutron Spectra and Kerma*. International Commission on Radiation Units and Measurements, Washington, D.C., USA.

ICRU Report 20 (1971): *Radiation Protection Instrumentation and its Application*. International Commission on Radiation Units and Measurements, Washington, D.C., USA.

ICRU Report 26 (1977): *Neutron Dosimetry for Biology and Medicine*. International Commission on Radiation Units and Measurements, Washington, D.C., USA.

ICRU Report 33 (1980): *Radiation Quantities and Units*. International Commission on Radiation Units and Measurements, Washington, D.C., USA.

1) ICRP: International Commission on Radiological Protection.

ICRU Report 39 (1985): *Determination of Dose Equivalents Resulting from External Radiation Sources*. International Commission on Radiation Units and Measurements, Bethesda, MD, USA.

ICRU Report 43 (1988): *Determination of Dose Equivalents Resulting from External Radiation Sources — Part 2*. International Commission on Radiation Units and Measurements, Bethesda, MD, USA.

3 Devices covered in this International Standard

This International Standard applies to individual dosimeters and to portable devices for surveying and area monitoring, some examples of which are given in 3.1 and 3.2. Some of these devices may also be used as reference instruments.

3.1 Individual dosimeters

Individual dosimeters include devices such as

- nuclear emulsions;
- solid-state nuclear track detectors;
- albedo dosimeters;
- pocket ionization chambers;
- bubble, or superheated drop, detectors;
- semiconductor detectors.

These shall be calibrated on a suitable phantom (see annex B). Reviews of the physical characteristics of individual dosimeters are given by Griffith et al. [1]. Reviews of calibration procedures are given by Eisenhauer et al. [2] and by Burger and Schwartz [3].

3.2 Devices for surveying and area monitoring

Dose equivalent meters, dose equivalent ratemeters or monitors for surveying and area monitoring of neutrons are generally portable devices, calibrated free-in-air, rather than on a phantom. These devices include

- moderating devices with thermal-neutron detectors for the measurement of neutrons over a wide range of energies;
- low-pressure tissue-equivalent proportional counters for the measurement of absorbed dose and dose equivalent;
- large ionization chambers for the measurement of tissue kerma;
- BF₃ and ³He proportional counters for the measurement of thermal neutrons;
- bubble, or superheated drop, detectors;
- recombination dose equivalent ratemeters.

More details concerning the characteristics of portable neutron dose-equivalent ratemeters, and of their calibration requirements and procedures are given in references [3] and [4] in annex H, and in IEC 1005. Complete definitions of radiation quantities and units can be found in ICRP Publication 21, ICRU Report 13, ICRU Report 33, ICRU Report 39, ICRU Report 43, and ISO 8529.

4 Definitions

For the purposes of this International Standard, the following definitions apply.

4.1 Metrological terms

4.1.1 reading, M : Value of the quantity indicated by an instrument.

4.1.2 response, R : Reading divided by the magnitude of the quantity causing it.

The type of response shall be specified, for example:

"fluence response",

$$R_{\Phi} = \frac{M}{\Phi} \quad \dots (1)$$

"dose equivalent response",

$$R_H = \frac{M}{H} \quad \dots (2)$$

or "photon dose equivalent response",

$$R_{H\gamma} = \frac{M}{H_{\gamma}} \quad \dots (3)$$

If M is a measurement of a rate, the quantities fluence, Φ , and dose equivalent, H , are replaced by fluence rate, ϕ , and dose equivalent rate, H , respectively.

The concept of dose equivalent, H , in this International Standard is intended to be general. For example, it may apply to the dose equivalent calculated at the maximum of the depth-dose equivalent curve, i.e. MADE (ICRP Publication 21) or the ambient dose equivalent (ICRU Report 39).

4.1.3 calibration factor, C : Reciprocal of the response; factor by which the reading, M , is multiplied to obtain the value of the quantity to be measured.

4.1.4 energy dependence of response, $R_{\Phi}(E)$ or $R_H(E)$: Response, R , with respect to fluence, Φ , or dose equivalent, H , to monoenergetic neutrons as a function of neutron energy, E .

4.1.5 angular dependence of response: Response as a function of the direction of incidence of neutrons on the device.

4.1.6 gamma-ray sensitivity: Change in the neutron reading of a device when gamma-rays are added to a neutron field.

(Compare with "photon dose equivalent response" in 4.1.2).

4.1.7 free-field quantities: Quantity which would exist if irradiations were performed in free space with no scatter or background effects.

4.2 Reference and monitoring instruments

4.2.1 primary standard instrument: Instrument that is capable of determining the required radiation quantity from known physical data only.

4.2.2 secondary standard instrument: Instrument whose radiation response characteristics have been determined by comparison with a primary standard instrument.

4.2.3 transfer instrument: Instrument whose radiation response characteristics, response precision and long-term stability make it suitable for comparing the measurement of a radiation quantity in one laboratory with that in another.

4.2.4 monitoring instrument: Instrument with suitable physical properties to monitor the radiation field characteristics in a given laboratory over a long term.

More details concerning primary standards, secondary standards, and transfer instruments can be found in ICRU Report 20 and ICRU Report 26.

4.3 Devices to be tested

4.3.1 neutron dosimeter: Device for the determination of neutron dose equivalent, kerma, or absorbed dose.

4.3.2 neutron dose-ratemeter: Device for the determination of neutron dose equivalent rate, kerma rate, or absorbed dose rate.

4.3.3 neutron detector: Device sensitive to neutrons.

5 Symbols

For the purposes of this International Standard, the following symbols, and certain symbols listed in annex G, apply.

Symbol	Meaning
k	Characteristic constant for source-detector combination
r	Detector radius
A	Total air-scatter component
C	Calibration factor
F_1	Geometry-correction factor
F_A	Air outscatter factor
F_I	Anisotropy correction factor
M_C	Instrument reading under free-field conditions
M_S	Instrument reading due to inscattered neutrons alone, during a shadow-cone calibration procedure
M_T	Total instrument reading during a calibration procedure
S	Room back-scatter component
δ	Effectiveness factor
Δ	Lateral dosimeter distance on phantom surface
Σ	Linear attenuation coefficient (energy averaged)

6 Traceability of the calibration of the reference radiation field

The neutron fluence rate of a radiation field established in a calibration laboratory in accordance with this International Standard shall be traceable to a recognized national or international standard. The method used to provide this calibration link is dependent upon the type of reference radiation field, but measurement traceability is usually achieved through the utilization of a transfer standard. This may be, for example, a radionuclide source (6.1) or an approved transfer instrument (see 6.2). The calibration of the field is valid in exact terms only at the time of the calibration, and thereafter shall be inferred, for example, from a knowledge of the half-life and isotopic composition of the radionuclide source, or knowledge of the properties of the transfer instrument.

The measurement technique used by the laboratory in the calibration of a neutron-measuring device shall also be approved by a body or institution as required by national regulations. An instrument of the same, or similar, type as that routinely calibrated by the laboratory shall be calibrated by both the reference laboratory and the calibration laboratory. These measurements shall be performed within each laboratory using its own approved calibration methods. In order to demonstrate that adequate traceability has been achieved, the calibration laboratory should obtain the same calibration factor, within agreed-upon limits, as that obtained by the reference laboratory.

The frequency of field calibrations shall be such that there is reasonable confidence that its value will not move outside the limits of its specification between successive calibrations. The frequency of calibration of the radionuclide neutron sources is given in ISO 8529. The calibration of the laboratory-approved transfer instrument, and the check on the measurement techniques used by the calibration laboratory, shall be carried out at least every five years, or whenever there are significant changes in the laboratory environment.

6.1 Traceability for radionuclide neutron sources

For calibrations using neutron fields produced by radionuclide neutron sources, traceability shall be provided either by using a radionuclide source whose angular source strength has been determined by a reference laboratory (see 7.2.1 for angular source strength), or by determining the fluence rate at the instrument test position using an