



International
Standard

ISO 11979-2

Ophthalmic implants — Intraocular lenses —

Part 2: Optical properties and test methods

*Implants ophtalmiques — Lentilles intraoculaires —
Partie 2: Propriétés optiques et méthodes d'essai*

Third edition
2024-10

Reference number
ISO 11979-2:2024(en)

Full document details: [ISO 11979-2-2024](#)

© ISO 2024  **Documents**
Document Sharing Hub



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	1
4.1 General	1
4.2 Dioptric power	2
4.2.1 General	2
4.2.2 Dioptric power for toric IOL (TIOL)	2
4.2.3 Dioptric power for simultaneous vision IOL (SVIOL)	2
4.2.4 Dioptric power for accommodating IOL (AIOL)	3
4.3 Imaging quality	3
4.3.1 General	3
4.3.2 Monofocal IOL	4
4.3.3 Toric IOL (TIOL)	4
4.3.4 Simultaneous vision IOL (SVIOL)	4
4.3.5 Accommodating IOL (AIOL)	4
4.3.6 Combination of optical principles	4
4.3.7 Exceptions	4
4.4 Optical characterization	5
4.5 Spectral transmittance	5
4.5.1 Measurement of spectral transmittance	5
4.5.2 Cut-off wavelength	5
Annex A (normative) Measurement of dioptric power	6
Annex B (normative) Measurement of MTF	14
Annex C (normative) Optical characterization	18
Bibliography	21

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 170, *Ophthalmic optics*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 11979-2:2014), which has been technically revised.

The main changes are as follows:

- A new category of simultaneous vision IOLs (SVIOL) is introduced for non-accommodating lenses that provide simultaneous vision at multiple distances. It includes multifocal IOLs (MIOL), extended depth of focus IOLs (EDF), and full visual range IOLs (FVR).
- Dioptric power, imaging quality, and characterization clauses and annexes were modified to include requirements for SVIOLs.
- Respective units of mm^{-1} and degree^{-1} were adopted for linear and angular spatial frequencies per ISO 9334.
- The resolution efficiency and associated annex have been removed from this document due to advancements in optical designs and the availability of modulation transfer function (MTF) imaging quality measurement methods.
- A new [Annex C](#) with associated requirements for all IOL categories has been added.
- Clarified description of UV cut-off wavelength.
- New references were added to the Bibliography.

A list of all parts in the ISO 11979 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document initially addressed monofocal IOLs and now includes the optical requirements and test methods for monofocal, toric, simultaneous vision, and accommodating IOLs. This document generally provides specific test methods and requirements connected to the optical function of intraocular lenses. In some cases, test methods do not have specified requirements, including:

- the spectral transmittance test that provides information related to UV transmission and potential exposure situations, e.g. when using laser light sources for diagnosis and treatment;
- optical characterization testing that informs potential optical design risks and guides potential clinical investigation design.

The specified dioptric power and imaging quality limits result from the analysis of extensive interlaboratory testing of the original spherical monofocal IOLs. Based on these studies, the respective dioptric power repeatability and reproducibility were about 0,5 % and 1 %, respectively, of the dioptric power as described in Reference [1]. Additionally, for IOLs in the 10 D to 30 D range, the respective expected imaging quality repeatability and reproducibility were 0,09 and 0,16 modulation transfer function values as described in Reference [2]. For other non-monofocal IOL designs, manufacturers should utilize model-specific repeatability and reproducibility precision limits to establish reliable final release criteria.

During the interlaboratory testing, some problems were encountered with measuring dioptric power, as described in Reference [1]. Specifically, the accuracy in determining dioptric power has an error that is not negligible in relation to the half dioptre steps in which intraocular lenses are commonly labelled. The dioptric power tolerances take this fact into account. Hence the limits set may lead to some overlap into the next labelled power, especially for high dioptre lenses. Reference [1] further discusses this subject.

Historically, imaging quality was tested using either

- a) Air Force target-based resolution efficiency, or
- b) MTF using a minimal spherical aberration model eye, or
- c) a manufacturer-defined spherical aberration model eye using modulation transfer function (MTF) testing.

Since the test method with Air Force target-based resolution efficiency is not optimal for quantifying image contrast, and better methods using MTF measurements have become mainstream in the industry, Air Force target-based resolution efficiency is not included in this revision as a reference method. The model eye with manufacturer-defined spherical aberration includes the option of having a model eye with minimal spherical aberration. Therefore, the original model eye with minimal spherical aberration is removed from this document. For lenses that have already been approved using the measurements in the previous edition, it is not necessary to retest these lens models with the method in this document.

[Annex B](#) describes a test method used to establish quality criteria for IOLs. The quality criteria assure consistent IOL optical quality. This document also includes a new normative optical characterization text (see [Annex C](#)), that is meant to provide preclinical assessments to inform of risks and benefits associated with the optical design and guide the design of the potential clinical investigation. The additional optical characterization is required only for lens models to be approved after publication of this document.

Ophthalmic implants — Intraocular lenses —

Part 2: Optical properties and test methods

1 Scope

This document specifies requirements and test methods for certain optical properties of intraocular lenses (IOLs) with monofocal, toric, simultaneous vision, and/or accommodative optics. The generic descriptor 'IOL' used throughout this document also includes phakic intraocular lenses (PIOL).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes the requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 9334, *Optics and photonics — Optical transfer function — Definitions and mathematical relationships*

ISO 9335, *Optics and photonics — Optical transfer function — Principles and procedures of measurement*

ISO 11979-1, *Ophthalmic implants — Intraocular lenses — Part 1: Vocabulary*

ISO 11979-4, *Ophthalmic implants — Intraocular lenses — Part 4: Labelling and information*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11979-1 and ISO 9334 apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

4 Requirements

4.1 General

The manufacturer shall assure that the entire range of available powers meets the specifications herein. All optical properties apply at in situ conditions, either by being measured at simulated in situ conditions, or being measured at other conditions and then corrected to in situ conditions.

For IOLs where the optic is intended to be deformed during implantation, it shall be demonstrated that dioptric power and imaging quality are retained at in situ or equivalent conditions following surgical manipulation and recovery. See ISO 11979-3^[3] for more details.

The test methods described in this document are reference methods. Alternative methods that produce equivalent results to those obtained with the reference methods may be used if the manufacturer can demonstrate that the IOLs meet the minimum dioptric power and imaging quality requirements.

For rotationally symmetric IOLs the manufacturer shall assure that lenses meet the requirements in all meridians, for example by selecting an arbitrary meridian for measurement.

4.2 Dioptric power

4.2.1 General

The base power of lenses as stated by the manufacturer in the IOL labelling per ISO 11979-4 shall be within the tolerance limits specified in [Table 1](#). Manufacturers shall consider measurement precision when establishing IOL release specifications.

Table 1 — Tolerance limits on spherical dioptric power, S

Nominal base power ^a D	Tolerance limits on spherical dioptric power D
$0 \leq S \leq 15$	$\pm 0,3$
$15 < S \leq 25$	$\pm 0,4$
$25 < S \leq 30$	$\pm 0,5$
$30 < S $	$\pm 1,0$
^a The dioptric power ranges apply to positive and negative dioptric powers.	

4.2.2 Dioptric power for toric IOL (TIOL)

When determined by any of the methods in [Annex A](#), the spherical equivalent (SE) power shall be within the tolerance limits for dioptric power specified in [Table 1](#). Additionally, the cylindrical power calculated as the absolute difference between the powers of the meridian of highest dioptric power and the meridian of lowest dioptric power shall be within the cylindrical power tolerance limits specified in [Table 2](#).

Table 2 — Tolerance limits on cylindrical dioptric power, C

Nominal cylindrical dioptric power D	Tolerance limits on cylindri- cal dioptric power D $SE < 25 D$	Tolerance limits on cylindri- cal dioptric power D $SE \geq 25 D$
$0 < C \leq 2,5$	$\pm 0,3$	$\pm 0,4$
$2,5 < C \leq 4,5$	$\pm 0,4$	$\pm 0,4$
$4,5 < C$	$\pm 0,5$	$\pm 0,5$

The TIOL shall have a physical axis indicator such as a mark, engraving, or label that aligns with the meridian of lowest dioptric power and is visible to the surgeon during implantation. The angle difference between the physical axis indicator and the meridian with the lowest dioptric power shall be less than or equal to $5,0^\circ$.

4.2.3 Dioptric power for simultaneous vision IOL (SVIOL)

Methods [A.3](#) to [A.4](#) can be applied to SVIOLs for determining the far power and any designed distinct addition power(s). The dioptric power of the far power shall be within the tolerance limits specified in [Table 1](#), and the dioptric power of designed distinct addition power(s) shall be within the tolerances in [Table 3](#). For SVIOLs that do not have designed distinct addition powers, the manufacturer shall develop MTF through-focus response specifications per [4.3.4](#).

Table 3 — Tolerance limits on addition dioptric power, A

Nominal addition dioptric power D	Tolerance limits on addition dioptric power D far power < 25 D	Tolerance limits on addition dioptric power D far power $\geq$ 25 D
$0 < A \leq 2,5$	$\pm 0,3$	$\pm 0,4$
$2,5 < A \leq 4,5$	$\pm 0,4$	$\pm 0,4$
$4,5 < A$	$\pm 0,5$	$\pm 0,5$

4.2.4 Dioptric power for accommodating IOL (AIOL)

The power associated with the far power configuration of an AIOL shall be determined by one of the methods in [Annex A](#). When determined by one of these methods, the dioptric power tolerances specified in [Table 1](#) shall apply to the power associated with the far power configuration of the AIOL. The dioptric power response of the lens or system in the eye shall be determined in a theoretical or laboratory eye model that simulates the intended accommodating mechanism of action.

4.3 Imaging quality

4.3.1 General

Reported imaging quality is dependent upon compatibility between the optical design, manufactured lens quality, and conditions that are used to evaluate optical performance. Imaging quality shall be specified in relation to theoretical lens performance in terms of a modulation transfer function (MTF) value at one or more specified spatial frequencies or the area under the MTF curve between two spatial frequencies for a given aperture. Manufacturers shall consider measurement precision when establishing IOL release specifications.

A method for measuring MTF and example model eye specifications are given in [Annex B](#). Alternatively, the manufacturer can specify an equivalent method or model eye with optical properties for the intended use and design. In this case, the model eye and the method shall be fully described, and a justification for the use thereof be provided. The imaging quality specifications apply to all available powers unless stated otherwise.

NOTE 1 The test apertures given in [4.3](#) and in [Annexes A, B, and C](#) represent the exposed central area of the IOL under test.

NOTE 2 Throughout this document, optical resolution is specified using spatial frequencies that are presented in cycles per millimetre (mm^{-1}). Alternatively, equivalent values for the generally accepted vision science convention of cycles per degree (degree^{-1}) can be used:

- where the document specifies 100 mm^{-1} , alternatively 30 degree^{-1} can be used;
- where the document specifies 50 mm^{-1} , alternatively 15 degree^{-1} can be used;
- where the document specifies 25 mm^{-1} , alternatively $7,5 \text{ degree}^{-1}$ can be used.

If conversion back from these values in degree^{-1} to mm^{-1} is needed for different lens powers, the following approximative conversion can be used, as shown in [Formula \(1\)](#)

with:

- SF = spatial frequency, expressed in degree^{-1} ;
- sf = spatial frequency, expressed in mm^{-1} ;
- EFL(P) = effective focal length of the model eye, with an IOL with power P (in D) in place;
- so that EFL(20) = EFL of the model eye with an IOL of 20 D.

Then: