



**Publicly
Available
Specification**

ISO/PAS 7020

Sizing parameters of surgical valve prostheses: Requirements regarding the application of ISO 5840-2

*Désignation des dimensions des prothèses valvulaires
chirurgicales: Exigences relatives à l'application de l'ISO 5840-2*

**First edition
2023-05**

**Corrected version
2024-03**

Reference number
ISO/PAS 7020:2023(en)

© ISO 2023  **Documents**
Document Sharing Hub



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

© ISO 2023 – All rights reserved

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Abbreviated terms	4
5 Information required for the outer container labelling	4
5.1 General	4
5.2 Intended valve to be replaced	5
5.3 Inflow orifice diameter	5
5.3.1 General	5
5.3.2 Mechanical valves	5
5.3.3 Stented bioprosthetic valves	5
5.4 Effective orifice diameter	7
5.4.1 General	7
5.4.2 Steady flow method	7
5.5 Valve housing external diameter	8
5.6 External sewing ring diameter	9
6 Labelling format	9
Bibliography	10

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This corrected version of ISO/PAS 7020:2023 incorporates the following corrections:

- in 5.5, second sentence of third paragraph, "If the measured diameter is based on fitting a circle constructed from selected points around the inflow inner surface of the valve, ..." has been replaced by: "If the measured diameter is based on fitting a circle constructed from selected points around the outer circumference of the valve housing, ...";
- in 5.6, second sentence of second paragraph, "If the measured diameter is based on fitting a circle constructed from selected points around the inflow inner surface of the valve, ..." has been replaced by: "If the measured diameter is based on fitting a circle constructed from selected points around the outer circumference of the sewing ring, ...".

Introduction

0.1 General

In the past, inconsistencies have been reported with the labelling and instructions for use associated with sizing parameters and sizing procedures for surgical replacement heart valves, specifically, mechanical and stented bioprosthetic valves. These inconsistencies have led to confusion among some users about which size valve to implant in a particular patient and have also led to challenges in comparing results (published or otherwise) from one valve model to another. A solution to the problem can be achieved by providing more complete and accurate sizing information to the clinicians, which will ultimately benefit the patients.

ISO 5840-2 identifies a number of sizing parameters that are required in the labelling (including on the unit box, see ISO 5840-2:2021, 6.3.3, and instructions for use) to inform the selection of a surgical heart valve prosthesis to be implanted in a specific patient. However, no guidance is offered in ISO 5840-2 on how these parameters should be obtained.

0.2 Clinical rationale for additional sizing information

Successful valve replacement therapy requires that an adequate size surgical heart valve substitute is used, based on patient body size and the native valve annulus size. An understanding of valve sizing parameters and appropriate choice of size is critical to post-procedure success since a valve substitute that is too small for the patient can result in prosthesis-patient mismatch. For aortic valve replacements, severe mismatch has been reported in 5 % to 15 % of patients.^[1] Severe prosthesis-patient mismatch leads to increased early, mid-term and late mortality, especially if the left ventricular ejection fraction (LVEF) is reduced.^{[2]-[5]} In the mid-term, it causes a higher incidence of heart failure^[5] and limits left ventricular mass regression.^[5] In the long term, it can also contribute to accelerated structural valve degeneration (SVD).^[6] Patients with severe prosthesis-patient mismatch can require replacement of the valve substitute with another having a larger effective orifice area (EOA). However, re-intervention has significant risk of mortality and morbidity.

The best approach to prosthesis-patient mismatch is prevention. This requires the surgeon to have clear and accurate information about the sizing parameters and EOA of each valve substitute.

A surgical heart valve substitute is described by a labelled size given by the manufacturer, which is assumed to be broadly consistent with the size of the patient native valve annulus for which the valve is intended. Literature reviews^[1] and studies of haemodynamic function commonly compare valve substitutes by labelled size, but there can be major differences between the patient native valve annulus diameter and the labelled size of the valve substitute.^{[8]-[11]} Intraoperative sizing is further complicated by the need for aortic supra-annular valves to fit within the aortic sinus. The disparity between labelled size and actual size means that echocardiographic or clinical comparisons based on labelled size can be misleading.

The issue of valve sizing is a complex problem and is being addressed in a stepwise fashion. The working group revising ISO 5840-2 proposed a first step toward greater transparency by requiring additional information be added to the unit box, namely, internal orifice diameter and effective orifice diameter. Although this information does not necessarily inform the surgeon on whether the valve would fit in the patient's annulus, it helps to estimate the internal orifice available for blood flow and thus indirectly the EOA. It is not feasible to use clinically measured EOA's since sizing information must be available before a surgical heart valve substitute is released for use in patients. Indeed, it can take a number of years to gather sufficient echocardiographic data to confirm the clinical EOAs. Furthermore, the use of echocardiographic data to help avoid prosthesis-patient mismatch has been criticized because of variability in the measurements obtained in vivo.^[12] In vitro steady flow data have less variability and allow meaningful comparison of every design and size of surgical heart valve substitute under the same flow conditions. This information can be used by the surgeon to choose a specific valve substitute type and size based on more standardized parameters than labelled valve size. It is anticipated that further steps toward a standardised approach to sizing will be addressed in subsequent editions of ISO 5840-2.

This document provides further specifications to explain these two parameters (i.e. internal orifice diameter and effective orifice diameter) and other sizing parameters. This document also guides the manufacturer in selecting reproducible methods to obtain these parameters and the degree of accuracy required.

