
**Measurement of fluid flow by means of
pressure-differential devices — Guidelines
for specification of nozzles and orifice
plates beyond the scope of ISO 5167-1**

*Mesurage du débit des fluides au moyen d'appareils déprimogènes —
Lignes directrices pour les spécifications des tuyères et diaphragmes non
couverts par l'ISO 5167-1*



Contents

1 Scope 1

2 Reference..... 1

3 Symbols 1

4 Principle of the method of measurement and computation..... 1

5 Square-edged orifice plates and nozzles with drain holes, in pipes below 50 mm diameter and as inlet and outlet devices 1

6 Orifice plates (except square-edged)..... 7

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The main task of technical committees is to prepare International Standards, but in exceptional circumstances a technical committee may propose the publication of a Technical Report of one of the following types:

- type 1, when the required support cannot be obtained for the publication of an International Standard, despite repeated efforts;
- type 2, when the subject is still under technical development or where for any other reason there is the future but not immediate possibility of an agreement on an International Standard;
- type 3, when a technical committee has collected data of a different kind from that which is normally published as an International Standard (“state of the art”, for example).

Technical Reports of types 1 and 2 are subject to review within three years of publication, to decide whether they can be transformed into International Standards. Technical Reports of type 3 do not necessarily have to be reviewed until data they provide are considered to be no longer valid or useful.

ISO/TR 15377, which is a Technical Report of type 3, was prepared by Technical Committee ISO/TC 30, *Measurement of fluid flow in closed conduits*, Subcommittee SC 2, *Differential pressure methods*.

Measurement of fluid flow by means of pressure-differential devices — Guidelines for specification of nozzles and orifice plates beyond the scope of ISO 5167-1

1 Scope

This Technical Report describes the geometry and method of use for conical-entrance orifice plates, quarter-circle orifice plates and eccentric orifice plates. Recommendations are also given for square-edged orifice plates and nozzles under conditions outside the scope of ISO 5167-1.

2 Reference

ISO 5167-1:1991, *Measurement of fluid flow by means of pressure differential devices — Part 1 : Orifice plates, nozzles and Venturi tubes inserted in circular cross-section conduits running full.*

3 Symbols

The symbols used in this Technical Report are given in table 1.

4 Principle of the method of measurement and computation

The principle of the method of measurement and computation is as specified in clause 5 of ISO 5167-1:1991.

5 Square-edged orifice plates and nozzles with drain holes, in pipes below 50 mm diameter and as inlet and outlet devices

5.1 Drain holes through the upstream face of the square-edged orifice plate or nozzle

5.1.1 General

Square-edged orifice plates and nozzles with drain holes may be used, installed, and manufactured in accordance with the following guidelines.

Table 1 — Symbols

Symbols	Represented quantity	Dimensions M: mass L: length T: time	SI unit
a	Pressure tapping hole diameter	L	m
C	Discharge coefficient	dimensionless	
d	Diameter of orifice or throat of primary device at operating conditions	L	m
D	Upstream internal pipe diameter at operating conditions	L	m
e	Thickness of bore	L	m
E, E_1	Thickness of orifice plate	L	m
F_E	Correction factor	dimensionless	
k	Uniform equivalent roughness	L	m
p	Static pressure of the fluid	$ML^{-1}T^{-2}$	Pa
q_m	Mass rate of flow	MT^{-1}	kg/s
r	Radius of profile	L	m
R_a	Roughness criterion	L	m
Re	Reynolds number	dimensionless	
Re_D, Re_d	Reynolds number referred to D or d	dimensionless	
β	Diameter ratio, $\beta = \frac{d}{D}$	dimensionless	
Δp	Differential pressure	$ML^{-1}T^{-2}$	Pa
ε	Expansibility (expansion) factor	dimensionless	
κ	Isentropic exponent	dimensionless	
ρ	Mass density of the fluid	ML^{-3}	kg/m ³
τ	Pressure ratio, $\tau = \frac{p_2}{p_1}$	dimensionless	

NOTE 1 Other symbols used in this Technical Report are defined at their place of use.

NOTE 2 Some of the symbols used in this Technical Report are different from those used in ISO 5167-1.

NOTE 3 Subscript 1 refers to the cross-section at the plane of the upstream pressure tapping. Subscript 2 refers to the cross-section at the plane of the downstream pressure tapping.

5.1.2 Square-edged orifice plates

If a drain hole is drilled through the orifice plate, the coefficient values specified in ISO 5167-1: 1991 should not be used unless the following conditions are observed.

- The pipe diameter should be larger than 100 mm.
- The diameter of the drain hole should not exceed $0,1d$ and no part of the hole should lie within a circle, concentric with the orifice, of diameter $(D - 0,2d)$. The outer edge of the drain hole should be as close to the pipe wall as practicable.

- c) The drain hole should be deburred and the upstream edge should be sharp.
- d) Single pressure tapplings should be orientated so that they are between 90° and 180° to the position of the drain hole.
- e) The measured orifice diameter, d_m , should be corrected to allow for the additional orifice area represented by the drain hole of diameter d_k as shown in the following equations.

$$d \approx d_m \left\{ 1 + 0,55 \left(\frac{d_k}{d_m} \right)^2 \right\}$$

$$d_m \approx d \left\{ 1 - 0,55 \left(\frac{d_k}{d} \right)^2 \right\}$$

NOTE These equations are based on the assumption that the value for $C\epsilon(1 - \beta^4)^{-0,5}$ for flow through the drain hole is 10 % greater than the value for flow through the orifice.

An additional uncertainty equivalent to 100 % of the drain hole correction should be added arithmetically to the discharge coefficient uncertainty when estimating the overall uncertainty of the flow measurement.

5.1.3 ISA 1932 nozzles

If a drain hole is drilled through the nozzle upstream face, the coefficient values specified in ISO 5167-1:1991 should not be used unless the following conditions are observed.

- a) The value of β should be less than 0,625.
- b) The diameter of the drain hole should not exceed 0,1 d and no part of the hole should lie within a circle, concentric with the throat, of diameter $(D - 0,2d)$.
- c) The length of the drain hole should not exceed 0,1 D .
- d) The drain hole should be deburred and the upstream edge should be sharp.
- e) Single pressure tapplings should be orientated so that they are between 90° and 180° to the position of the drain hole.
- f) The measured diameter, d_m , should be corrected to allow for the additional throat area represented by the drain hole of diameter d_k as shown in the following equations.

$$d \approx d_m \left\{ 1 + 0,40 \left(\frac{d_k}{d_m} \right)^2 \right\}$$

$$d_m \approx d \left\{ 1 - 0,40 \left(\frac{d_k}{d} \right)^2 \right\}$$

NOTE These equations are based on the assumption that the value for $C\epsilon(1 - \beta^4)^{-0,5}$ for flow through the drain hole is 10 % greater than the value for flow through the throat of the nozzle.

An additional uncertainty equivalent to 100 % of the drain hole correction should be added arithmetically to the discharge coefficient uncertainty when estimating the overall uncertainty of the flow measurement.

5.1.4 Long radius nozzles

Drain holes through these primary elements should not be used.

5.2 Square-edged orifice plates installed in pipes of diameter $25 \text{ mm} \leq D < 50 \text{ mm}$

5.2.1 General

Orifice plates should be installed and manufactured in accordance with ISO 5167-1.

5.2.2 Limits of use

When square-edged orifice plates are installed in pipes of bore 25 mm up to 50 mm the following conditions should be strictly observed.

- a) The pipes should have high quality internal surfaces such as drawn copper or brass tubes, glass or plastics pipes or drawn or fine-machined steel tubes. The steel tubes should be of stainless steel for use with corrosive fluids such as water. The uniform equivalent roughness, k , should be $< 0,03 \text{ mm}$ for all diameter ratios. If the pipe is machined, the surface finish should be better than $0,3 \text{ }\mu\text{m}$.
- b) Corner taps should be used, preferably of the carrier ring type detailed in figure 6 a) of ISO 5167-1: 1991.
- c) The diameter ratio, β , should be within the range $0,23 \leq \beta \leq 0,7$ where $0,032 \leq C\beta^2(1 - \beta^4)^{-0,5} \leq 0,350$.

5.2.3 Discharge coefficients and corresponding uncertainties

The Stolz equation for corner tapplings given in 8.3.2.1 of ISO 5167-1: 1991 should be used for deriving the discharge coefficients provided the minimum pipe Reynolds numbers are above the following values.

$$Re_d \geq 40\,000 \beta^2 \text{ for } 0,23 \leq \beta \leq 0,5$$

$$Re_d \geq 10\,000 \text{ for } 0,5 \leq \beta \leq 0,7$$

An additional uncertainty of 1,0 % should be added arithmetically to the uncertainty derived from 8.3.3 of ISO 5167-1:1991.

5.3 No upstream or downstream pipeline

5.3.1 General

This clause should apply where there is no pipeline on either or both the upstream or downstream sides of the device, that is for flow from a large space into a pipe or vice versa, or flow through a device installed in the partition wall between two large spaces.

5.3.2 Flow from a large space (no upstream pipeline) into a pipeline or another large space

5.3.2.1 Upstream and downstream tapplings

The space on the upstream side of the device should be considered large if:

- a) there is no wall closer than $4d$ to the axis of the device or to the plane of the upstream face of the orifice or nozzle,
- b) the velocity of the fluid at any point more than $4d$ from the device is less than 3 % of the velocity in the orifice or throat,
- c) the diameter of the downstream pipeline is not less than $2d$.

NOTE 1 The first condition implies, for example, that an upstream pipeline of diameter greater than $8d$ (that is where $\beta < 0,125$) may be regarded as a large space. The second condition, which excludes upstream disturbances due to draughts, swirl and jet effects, implies that the fluid is to enter the space uniformly over an area of not less than 33 times the area of the orifice or throat. For example, if the flow is provided by a fall in level of a liquid in a tank, the area of the liquid surface is not to be less than 33 times the area of the orifice or throat through which the tank is discharged.