



JAPANESE
INDUSTRIAL
STANDARD

Translated and Published by
Japanese Standards Association

Ⓔ JIS C 1102-5 : 1997
(IEC 51-5 : 1985)

**Direct acting indicating analogue
electrical measuring instruments
and their accessories
Part 5 : Special requirements
for phase meters, power factor
meters and synchrosopes**

ICS 17.220.20

Descriptors : electrical measuring instruments, phase measurement (electric), power factor, phase (electric)

Reference number : JIS C 1102-5 : 1997 (E)



Documents

Document Sharing Hub

85

C 1102-5 : 1997 (IEC 51-5 : 1985)

Foreword

This translation has been made based on the original Japanese Industrial Standard established by the Minister of International Trade and Industry through deliberations at Japanese Industrial Standards Committee in accordance with the Industrial Standardization Law. By this establishment JIS C 1102 : 1981 was withdrawn and replaced with this Standard.

Constitution of Standard

JIS C 1102 consists of the following 9 parts under the common title *Direct acting indicating analogue electrical measuring instruments and their accessories*.

- Part 1 : *Definitions and general requirements common to all parts*
- Part 2 : *Special requirements for ammeters and voltmeters*
- Part 3 : *Special requirements for wattmeters and varmeters*
- Part 4 : *Special requirements for frequency meters*
- Part 5 : *Special requirements for phase meters, power factor meters and synchroscopes*
- Part 6 : *Special requirements for ohmmeters (impedance meters) and conductance meters*
- Part 7 : *Special requirements for multi-function instruments*
- Part 8 : *Special requirements for accessories*
- Part 9 : *Recommended test methods*

Parts 2 to 9 are not complete in themselves and shall be read in conjunction with Part 1.

Parts 1 to 8 are arranged in the same format and a standard relationship between subject and clause number is maintained throughout. In addition, tables, figures and annexes add a suffix to the part number in order to differentiate the parts. This re-arrangement will assist the reader of this Standard to distinguish information relating to the different types of instruments.

Date of Establishment: 1997-10-20

Date of Public Notice in Official Gazette: 1997-10-20

Investigated by: Japanese Industrial Standards Committee
Divisional Council on Electricity

JIS C 1102-5:1997, First English edition published in 2000-07

Translated and published by: Japanese Standards Association
4-1-24, Akasaka, Minato-ku, Tokyo, 107-8440 JAPAN

In the event of any doubts arising as to the contents,
the original JIS is to be the final authority.

© JSA 2000

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Printed in Japan

Contents

	Page
Introduction	1
1 Scope	1
2 Definitions	1
3 Description, classification and compliance	1
4 Reference conditions and intrinsic errors	2
5 Nominal range of use and variations	3
6 Further electrical and mechanical requirements	6
7 Constructional requirements	8
8 Information, general markings and symbols	9
9 Markings and symbols for terminals	9
10 Tests to prove compliance with this standard	9

**Direct acting indicating analogue
electrical measuring instruments
and their accessories
Part 5 : Special requirements for
phase meters, power factor
meters and synchrosopes**

Introduction This Japanese Industrial Standard has been prepared based on IEC 51-5 *Direct acting indicating analogue electrical measuring instruments and their accessories Part 5 : Special requirements for phase meters, power factor meters and synchrosopes* issued in 1985 as the fourth edition without changing the technical contents.

1 Scope

1.1 Part 5 applies to direct acting indicating analogue phase meters, power factor meters and synchrosopes.

1.2 This part also applies to non-interchangeable accessories (as defined in Sub-clause 2.1.15.3 of JIS C 1102-1) used with phase meters, power factor meters and synchrosopes.

1.3 This part also applies to a phase meter or power factor meter whose scale marks do not correspond directly to its electrical input quantity, provided that the relationship between them is known.

1.4 to 1.8 See JIS C 1102-1.

2 Definitions

See JIS C 1102-1.

3 Description, classification and compliance

3.1 Description

Phase meters, power factor meters and synchrosopes shall be described according to their method of operation as given in Sub-clause 2.2 of JIS C 1102-1.

3.2 Classification

Phase meters, power factor meters and synchrosopes shall be classified in one of the accuracy classes denoted by the following class indices:

0.1, 0.2, 0.3, 0.5, 1, 1.5, 2, 2.5, 3, 5

3.3 Compliance with the requirements of this standard

See JIS C 1102-1.