



JAPANESE  
INDUSTRIAL  
STANDARD

Translated and Published by  
Japanese Standards Association

---

---

JIS T 3222 : 2023

(MTJAPAN/JSA)

Sterile winged intravenous devices

---

ICS 11.040.25

Reference number : JIS T 3222 : 2023 (E)

PROTECTED BY COPYRIGHT

T 3222 : 2023

Date of Establishment: 2005-03-25

Date of Revision: 2023-06-01

Date of Public Notice in Official Gazette: 2023-06-01

Investigated by: Japanese Industrial Standards Committee  
Standards Board for ISO area  
Technical Committee on Medical Equipment

---

JIS T 3222 : 2023, First English edition published in 2024-05

Translated and published by: Japanese Standards Association  
Mita Avanti, 3-11-28, Mita, Minato-ku, Tokyo, 108-0073 JAPAN

---

In the event of any doubts arising as to the contents,  
the original JIS is to be the final authority.

© JSA 2024

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Printed in Japan

PROTECTED BY COPYRIGHT

HN  
 **Documents**  
Document Sharing Hub

## Contents

|                                                              | Page |
|--------------------------------------------------------------|------|
| 1 Scope .....                                                | 1    |
| 2 Normative references .....                                 | 1    |
| 3 Terms and definitions .....                                | 1    |
| 4 Configuration and name of each part .....                  | 2    |
| 5 Performance .....                                          | 3    |
| 5.1 Leakage .....                                            | 3    |
| 5.2 Injection site .....                                     | 4    |
| 6 Physical requirements .....                                | 4    |
| 6.1 Needle tube .....                                        | 4    |
| 6.2 Winged intravenous device .....                          | 4    |
| 6.3 Tubing .....                                             | 4    |
| 6.4 Needle point .....                                       | 5    |
| 6.5 Lubricant .....                                          | 5    |
| 7 Chemical requirements .....                                | 5    |
| 8 Biological safety .....                                    | 6    |
| 9 Sterility assurance .....                                  | 6    |
| 10 Dimensions and structures .....                           | 6    |
| 10.1 Tolerance on dimensions .....                           | 6    |
| 10.2 Colour code .....                                       | 7    |
| 10.3 Taper fittings of female conical fitting .....          | 7    |
| 11 Packaging .....                                           | 7    |
| 11.1 Primary container .....                                 | 7    |
| 11.2 Secondary container .....                               | 7    |
| 12 Labelling .....                                           | 7    |
| 12.1 Primary container .....                                 | 7    |
| 12.2 Secondary container .....                               | 7    |
| 12.3 Use of symbols .....                                    | 8    |
| Annex A (normative) Test method 2 for dissolved matter ..... | 9    |

## Foreword

This Japanese Industrial Standard has been revised by the Minister of Health, Labour and Welfare through deliberations at the Japanese Industrial Standards Committee as the result of proposal for revision of Japanese Industrial Standard submitted by Medical Technology Association of Japan (MTJAPAN)/Japanese Standards Association (JSA) with a draft being attached, based on the provision of Article 12, paragraph (1) of the Industrial Standardization Act applied mutatis mutandis pursuant to the provision of Article 16 of the said Act. This edition replaces the previous edition (**JIS T 3222** : 2011), which has been technically revised.

This **JIS** document is protected by the Copyright Act.

Attention is drawn to the possibility that some parts of this Standard may conflict with patent rights, published patent application or utility model rights. The relevant Minister and the Japanese Industrial Standards Committee are not responsible for identifying any of such patent rights, published patent application or utility model rights.

# Sterile winged intravenous devices

## 1 Scope

This Japanese Industrial Standard specifies requirements for sterile single-use winged intravenous devices (hereafter referred to as winged intravenous devices).

However, **JIS T 3222 : 2011** remains valid for three years from the date of public notice of the revision of this Standard.

## 2 Normative references

Part or all of the provisions of the following standards, through reference in this text, constitute provisions of this Standard. For standards with the year indication, only the editions of the indicated year shall be applied and the revisions (including amendments) made thereafter shall not be applied. For those without the indication of the year, the most recent editions (including amendments) shall be applied.

JIS G 4305 : 2021 *Cold-rolled stainless steel plate, sheet and strip*

JIS T 0993-1 *Biological evaluation of medical devices — Part 1 : Evaluation and testing within a risk management process*

JIS T 3210 : 2022 *Sterile injection syringes*

ISO 3696 *Water for analytical laboratory use — Specification and test methods*

ISO 6009 *Hypodermic needles for single use — Colour coding for identification*

ISO 9626 : 2016 *Stainless steel needle tubing for the manufacture of medical devices — Requirements and test methods*

ISO 80369-7 *Small-bore connectors for liquids and gases in healthcare applications — Part 7: Connectors for intravascular or hypodermic applications*

## 3 Terms and definitions

For the purpose of this Standard, the following terms and definitions apply.

### 3.1

#### **nominal outside diameter**

dimension of the outside diameter of needle tube, which is indicated on the encapsulation or container of the winged intravenous device

### 3.2

#### **nominal length**

dimension of the length of needle tube, which is indicated on the encapsulation or container of the winged intravenous device