



JAPANESE
INDUSTRIAL
STANDARD

Translated and Published by
Japanese Standards Association

JIS T 3226-1 : 2023

(MTJAPAN/JSA)

Needle-based injection systems for
medical use — Part 1: Needle-based
injection systems — Requirements and
test methods

ICS 11.040.20

Reference number: JIS T 3226-1 : 2023 (E)

PROTECTED BY COPYRIGHT

T 3226-1 : 2023

Date of Establishment: 2005-03-25

Date of Revision: 2023-07-01

Date of Public Notice in Official Gazette: 2023-07-03

Investigated by: Japanese Industrial Standards Committee
Standards Board for ISO area
Technical Committee on Medical Equipment

JIS T 3226-1 : 2023, First English edition published in 2024-09

Translated and published by: Japanese Standards Association
Mita Avanti, 3-11-28, Mita, Minato-ku, Tokyo, 108-0073 JAPAN

In the event of any doubts arising as to the contents,
the original JIS is to be the final authority.

© JSA 2024

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Printed in Japan

PROTECTED BY COPYRIGHT

HN
 **Documents**
Document Sharing Hub

Contents

	Page
Introduction	1
1 Scope	1
2 Normative references	1
3 Terms and definitions	3
4 Symbols and abbreviated terms	5
5 Requirements	7
5.1 General	7
5.2 System designations	7
5.3 Risk analysis requirements	8
5.4 Uncertainty of measurement and conformance with specifications	8
5.5 General design requirements	8
6 Reagent and apparatus	10
6.1 General	10
6.2 Test liquid	10
6.3 Balance	11
6.4 Test surface for free-fall testing	11
7 Determination of dose accuracy	11
7.1 General	11
7.2 Dosing regions	12
7.3 Dose settings	12
7.4 Assessment	13
8 Preparation and operation of NISs	17
9 Test matrix	17
10 Test descriptions	19
10.1 General	19
10.2 Cool, standard and warm atmosphere testing	19
10.3 Last-dose testing (system designations A and C only)	20
10.4 Life-cycle testing (systems designations A and B only) —Pre-conditioning	20
10.5 Free-fall testing	20
10.6 Dry-heat and cold-storage testing — Pre-conditioning	22
10.7 Damp-heat testing (system designations A and B only) — Pre-conditioning	22
10.8 Cyclical testing (system designations A and B only) — Pre-conditioning	23
10.9 Vibration testing — Pre-conditioning	23

10.10	Electromagnetic disturbance (EMD) (systems with electronics only)	23
11	Inspection	23
11.1	Visual inspection	23
11.2	Container inspection	24
11.3	Dose accuracy acceptance criteria	24
12	Test report	24
13	Information supplied by the manufacturer	25
13.1	General	25
13.2	Marking	25
13.3	Instructions for use	26
Annex A (informative)	Dose replicates, accuracy and testing rationale	28
Annex B (informative)	One- and two-sided tolerance limit factors, k	32
Annex JA (normative)	Requirements for containers	43
Bibliography		44
Annex JB (informative)	Comparison table between JIS and corresponding International Standard	45

Foreword

This Japanese Industrial Standard has been revised by the Minister of Health, Labour and Welfare through deliberations at the Japanese Industrial Standards Committee as the result of proposal for revision of Japanese Industrial Standard submitted by Medical Technology Association of Japan (MTJAPAN)/Japanese Standards Association (JSA) with a draft being attached, based on the provision of Article 12, paragraph (1) of the Industrial Standardization Act applied mutatis mutandis pursuant to the provision of Article 16 of the said Act. This edition replaces the previous edition (**JIS T 3226-1 : 2015**), which has been technically revised.

This **JIS** document is protected by the Copyright Act.

Attention is drawn to the possibility that some parts of this Standard may conflict with patent rights, published patent application or utility model rights. The relevant Minister and the Japanese Industrial Standards Committee are not responsible for identifying any of such patent rights, published patent application or utility model rights.

JIS T 3226 series consists of the following 2 parts under the general title *Needle-based injection systems for medical use* —:

Part 1: Needle-based injection systems — Requirements and test methods

Part 2: Needles — Requirements and test methods

Blank

Copyrighted material

Needle-based injection systems for medical use — Part 1: Needle-based injection systems — Requirements and test methods

Introduction

This Japanese Industrial Standard has been prepared based on **ISO 11608-1 : 2014**, Edition 3, with some modifications of the technical contents to reflect the local conditions in Japan.

Annex JA contains requirements for containers that are unique to **JIS** and not given in the corresponding International Standard.

The dotted underlines indicate changes from the corresponding International Standard. A list of modifications with the explanations is given in Annex JB.

1 Scope

This Standard specifies requirements and test methods for needle-based injection systems (NISs) intended to be used with needles and with replaceable or non-replaceable containers. Containers covered in this Standard include single- and multi-dose syringe-based and cartridge-based systems, filled either by the manufacturer or by the end-user.

Additional guidance for NISs equipped with electronic or electromechanical components and NISs equipped with automated functions is given in **ISO 11608-4** and **ISO 11608-5** respectively.

Needle-free injectors, and requirements relating to methods or equipment associated with end-user filling of containers, are outside the scope of this Standard.

NOTE The International Standard corresponding to this Standard and the symbol of degree of correspondence are as follows.

ISO 11608-1 : 2014 *Needle-based injection systems for medical use — Requirements and test methods — Part 1: Needle-based injection systems* (MOD)

In addition, symbols which denote the degree of correspondence in the contents between the relevant International Standard and **JIS** are IDT (identical), MOD (modified), and NEQ (not equivalent) according to **ISO/IEC Guide 21-1**.

The previous edition of this Standard, **JIS T 3226-1 : 2015**, remains valid for three years from the date of public notice of the revision of this Standard.

2 Normative references

Part or all of the provisions of the following standards, through reference in this text, constitute provisions of this Standard. For standards with the year indication, only the

editions of the indicated year shall be applied and the revisions (including amendments) made thereafter shall not be applied. For those without the indication of the year, the most recent editions (including amendments) shall be applied.

JIS B 0641-1 *Geometrical product specifications (GPS) — Inspection by measurement of workpieces and measuring equipment — Part 1: Decision rules for proving conformance or nonconformance with specifications*

NOTE Normative reference in the corresponding International Standard: ISO 14253-1 *Geometrical product specifications (GPS) — Inspection by measurement of workpieces and measuring equipment — Part 1: Decision rules for verifying conformity or nonconformity with specifications*

JIS C 60068-2-6 : 2010 *Environmental testing — Part 2-6: Tests — Test Fc: Vibration (sinusoidal)*

NOTE Normative reference in the corresponding International Standard: IEC 60068-2-6 : 2007 *Environmental testing — Part 2-6: Tests — Test Fc: Vibration (sinusoidal)*

JIS C 60068-2-30 : 2011 *Environmental testing — Part 2-30: Tests — Test Db: Damp heat, cyclic (12 h + 12 h cycle)*

NOTE Normative reference in the corresponding International Standard: IEC 60068-2-30 : 2005 *Environmental testing — Part 2-30: Tests — Test Db: Damp heat, cyclic (12 h + 12 h cycle)*

JIS Q 13485 : 2018 *Medical devices — Quality management systems — Requirements for regulatory purposes*

NOTE 1 Normative reference in the corresponding International Standard: ISO 13485 : 2003 *Medical devices — Quality management systems — Requirements for regulatory purposes*

NOTE 2 To update the contents, the most recent edition of **JIS Q 13485** that corresponds to **ISO 13485 : 2016** is listed above instead of the edition corresponding to **ISO 13485 : 2003**.

JIS T 0601-1-2 *Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral Standard: Electromagnetic disturbances — Requirements and tests*

NOTE Normative reference in the corresponding International Standard: IEC 60601-1-2 *Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral standard: Electromagnetic compatibility — Requirements and tests*

JIS T 0993-1 *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

NOTE Normative reference in the corresponding International Standard: ISO 10993-1 *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

JIS T 2304 *Medical device software — Software life cycle processes*

NOTE Normative reference in the corresponding International Standard: IEC

62304 *Medical device software — Software life cycle processes*

JIS T 3226-2 *Needle-based injection systems for medical use — Part 2: Needles — Requirements and test methods*

NOTE Normative reference in the corresponding International Standard: ISO 11608-2 *Needle-based injection systems for medical use — Requirements and test methods — Part 2: Needles*

JIS T 14971 *Medical devices — Application of risk management to medical devices*

NOTE Normative reference in the corresponding International Standard: ISO 14971 *Medical devices — Application of risk management to medical devices*

JIS T 62366-1 *Medical devices — Part 1: Application of usability engineering to medical devices*

NOTE Normative reference in the corresponding International Standard: IEC 62366 *Medical devices — Application of usability engineering to medical devices*

ISO 11608-4 : 2006 *Pen-injectors for medical use — Part 4: Requirements and test methods for electronic and electromechanical pen-injectors*

ISO 11608-5 *Needle-based injection systems for medical use — Requirements and test methods — Part 5: Automated functions*

ISO/IEC Guide 98-3 *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement* (GUM: 1995)

IEC 60601-1-11 *Medical electrical equipment — Part 1-11: General requirements for basic safety and essential performance — Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

3 Terms and definitions

For the purpose of this Standard, the following terms and definitions apply.

3.1

cap

part of the NIS intended to protect the injector and its contents

3.2

container

unit packaging that contains the medicinal product for injection (either single-compartment or multi-compartment)

3.3

dose delivery efficiency

ratio of expelled dose to fill volume

Note 1 to entry Dose delivery efficiency is expressed as a percentage.

Note 2 to entry Delivery efficiency can be used to evaluate dose accuracy for NISs designed to fully empty single-dose containers filled by the user.

3.4

dialling resolution

smallest possible increment to be selected between dose amounts

3.5

dose accuracy

accuracy with which the NIS delivers a pre-set dose of medicinal product

3.6

“dose delivered” indication

dose number shown in the dose window indicating the amount of medicinal product delivered

Note 1 to entry This applies to variable multi-dose NISs that allow the setting of a dose greater than the remaining volume.

Note 2 to entry If the dose window indicates the amount of medicinal product yet to be delivered, then the “dose delivered” indication can be determined as the intended dose minus the indication of medicinal product yet to be delivered.

3.7

manufacturer-filled

container supplied to the user pre-filled by the manufacturer of medicinal products

Note 1 to entry This medicinal product can be in liquid form or lyophilized with diluent in the same container.

3.8

minimum deliverable dose

minimum dose that is ensured by the manufacturer to be delivered in a single-dose manufacturer-filled NIS designed to fully empty the container

3.9

needle-based injection system, NIS

injection system intended for parenteral administration by injection of medicinal products using a needle and a multi-dose or single-dose container

3.10

pre-setting

procedure by which individual amounts of medicinal product can be selected for injection by the user

Note 1 to entry The doses may be pre-set by the manufacturer.

3.11

residual scale

graduated scale which indicates the remainder of medicinal product in the container

3.12

user packaging

what is provided to the user with one or a collection of devices of the same item and from the same manufacturing batch, including the directions for use

3.13

user-filled

container that is filled or reconstituted (if in lyophilized form) by the user from a separate medicinal product or diluent container

3.14

cylinder

body of the cartridge

3.15

plunger

part to seal hermetically one end of the cartridge and expel the medicinal products in conjunction with the injection mechanism

3.16

disk

component (plug) to seal hermetically that end of the cartridge which is opposite to the plunger

3.17

cap

part to attach the disk to the cartridge

4 Symbols and abbreviated terms

NIS Needle-based injection system.

V_{set} One of the three pre-set doses (expressed as a volume, in millilitres) used in determining the dose accuracy for a given NIS. V_{set} is defined as one of the following:

- minimum dose ($V_{\text{set}} = V_{\text{min}}$) [specified in the instructions for use (or in precautions and other information)];
- maximum dose ($V_{\text{set}} = V_{\text{max}}$) [specified in the instructions for use (or in precautions

and other information)];

- c) midpoint dose ($V_{\text{set}} = V_{\text{mid}}$), where V_{mid} is defined as the injector setting closest to $(V_{\text{min}} + V_{\text{max}})/2$.

NOTE 1 Recommended doses as specified in the instructions for use (or in precautions and other information) may differ from the pre-set doses used for determining the dose accuracy.

NOTE 2 System designations B1 and D1 define V_{set} to be equal to the manufacturer-filled or user-filled volumes. System designations B2 and D2 define V_{set} to be equal to a single pre-set dose representing a portion of the manufacturer-filled or user-filled volumes. In the case of last-dose accuracy assessments for system designations A and C, V_{set} is equal to V_{mid} , the TP, or dose error (evaluated over a range of doses within a specified percentage of the TP).

V_{meas} The volumetric measurement value for a given V_{set} , expressed in millimetres.

G_{meas} The gravimetric measurement value for a given V_{set} , expressed in grams.

ρ Density, expressed in grams per millilitre.

p Probability content.

Y Number of NISs required for a given test.

R Number of replicates required for a given test. A replicate is a random sequence of V_{min} , V_{mid} , and V_{max} . There are six possible replicates.

n Number of measurements, V_{meas} , to be made for each V_{set} .

$\bar{x}$ The sample mean; when based on a random sample, an estimate of the true mean:

$$\bar{x} = \sum \frac{V_{\text{meas}}}{n}$$

s The sample standard deviation; when based on a random sample, an estimate of the true standard deviation:

$$s = \left[\sum \frac{(V_{\text{meas}} - \bar{x})^2}{(n-1)} \right]^{1/2}$$

k k value, or tolerance limit factor, determined from the confidence level (95 %), probability content, p , and number of accuracy measurements, n , conducted at each dose setting.

k_{act} Actual k value, determined from the following equations:

Two-sided

$$\min \left[\frac{(U - \bar{x})}{s}, \frac{(\bar{x} - L)}{s} \right]$$

One-sided

$$\left[\frac{(\bar{x} - L)}{s} \right] \text{ or } \left[\frac{(U - \bar{x})}{s} \right]$$

- k_{tar} Target k value, found from the look-up table in Annexes D and E of **ISO 16269-6** : 2014 [3] or Annex B of this Standard.
- DR Dialling resolution, the minimum dialling increment of the NIS.
- α Absolute error, in millimetres, used to define the upper and lower specification limits for a pre-set dose in absolute terms.
- β Relative error, as a percentage, used to define the upper and lower specification limits for a pre-set dose in relative terms.
- TP The transition point volume, in millilitres, at which the upper and lower specification limits for V_{set} change from absolute terms to relative terms (i.e. V_{set} where α and β are equal):
- $$\text{TP} = (100 \times \alpha) / \beta$$
- U Upper specification limit for a given V_{set} .
- L Lower specification limit for a given V_{set} .

5 Requirements

5.1 General

Companies wishing to verify a NIS shall ensure that the system meets the requirements of this Standard.

In addition, companies shall ensure that the appropriate components (e.g. needles and containers) and features (e.g. electromechanical drive systems and automated functions) specified for use in the system satisfy the relevant parts of **ISO 11608**.

5.2 System designations

Given the differences in device designs and containers (e.g. multi-dose, single dose with partial evacuation, and single-dose with full evacuation), the following system designations are provided to clearly associate the appropriate test and dose accuracy method with the injection system under consideration.

Containers can be either manufacturer-filled or user-filled.

Table 1 shows the various needle-based injector system designations.

Table 1 System designations

Multi-dose container	Single-dose container
A Needle-based injection device with replaceable container. Each container holds multiple doses, the size of which may be fixed or variable (pre-set by the user).	B1 Needle-based injection device with replaceable container. Each container holds a single dose, whereby the entire deliverable volume is expelled.
—	B2 Needle-based injection device with replaceable container. Each container holds a single dose, whereby a portion of the deliverable volume is expelled.
C Needle-based injection device with integrated non-replaceable container. Each container holds multiple doses, the size of which may be fixed or variable (pre-set by the user).	D1 Needle-based injection device with integrated non-replaceable container. Each container holds a single dose, whereby the entire deliverable volume is expelled.
—	D2 Needle-based injection device with integrated non-replaceable container. Each container holds a single dose, whereby a portion of the deliverable volume is expelled.

5.3 Risk analysis requirements

The manufacturer shall conduct risk assessments in accordance with **JIS T 14971**. These risk assessments shall consider all aspects of the development, manufacture and intended use of the NIS for medical use. The NIS shall conform to the usability requirements specified in **JIS T 62366-1**.

5.4 Uncertainty of measurement and conformance with specifications

Measurement uncertainty shall be evaluated and expressed by the laboratory performing the test in accordance with **ISO/IEC Guide 98-3**.

Conformance with specifications is established in accordance with **JIS B 0641-1**.

5.5 General design requirements

- The container holder shall allow visibility of the deliverable volume. The manufacturer shall determine, by risk analysis, if a residual scale is required and how much of the deliverable volume shall be visible.
- With the exception of system designations B2 and D2, NISs shall be designed in such a way that they are able to accurately deliver the entire labelled volume from the container for which they are designed.
- NISs with system designation B1 where the container is user-filled shall be designed in such a way that they are capable of delivering the maximum volume needed to fill the container, as specified in the labelling.

- d) When the injection system requires the user to pre-set the dose, the injector shall provide an indication of the dose that has been set. This information can be displayed in drug-specific units (e.g. millilitres, milligrams, international units) or in a setting specified by the physician (e.g. number, letter, percentage) as appropriate for the drug to be delivered. When the dose has been pre-set by the manufacturer, the dose can be indicated by the device or the system labelling, as appropriate.
- e) There shall be an indication of the pre-setting by visual and either tactile and/or audible means.
- f) The NIS shall indicate, at least by visual means, that it is ready for injection.
- g) The state of the NIS, when ready to deliver a dose, shall be different from its state when the dose has been delivered. The difference shall be visible.
- h) The NIS shall indicate, by visual, audible or tactile means, or any combination of these, that the injection stroke has been completed.
- i) NISs with system designation D2 shall be designed in such a way that it is impossible to deliver the remaining volume following the actuation and that it is impossible to reactivate the device.
- j) Variable multi-dose NISs (system designations A and C) shall be designed so that they:
 - 1) do not allow a larger dose to be pre-set than is left in the container, or
 - 2) do not allow dose delivery if the pre-set amount exceeds the amount of medicinal product left in the container, or
 - 3) indicate the amount of medicinal product delivered, or
 - 4) indicate the amount of medicinal product not delivered (of the pre-set dose).
- k) Fixed multi-dose NISs shall not allow pre-setting of the dose if a volume that is insufficient for the full fixed dose remains.
- l) The NIS shall be designed to function with its specified needles. **JIS T 3226-2** provides guidance for needles.
- m) The NIS shall be designed to function with its specified containers. Annex JA provides requirements for containers.
- n) If the NIS has automated functions, the requirements of **ISO 11608-5** shall be fulfilled.
- o) If the NIS contains electronic or electromechanical components and/or software, the requirements of **ISO 11608-4** shall be fulfilled.
- p) To avoid inadvertent disabling of the NIS containing replaceable batteries, it shall not be possible to remove the batteries unless two independent movements are applied.
- q) If designed with small parts that might be swallowed, the NIS labelling shall include warnings preventing access by children under the age of 3 years.

- r) If the NIS contains batteries, it shall be designed to allow the user to determine the state of the power supply.
- s) If the NIS contains software, the software shall be designed based on a life-cycle model in accordance with **JIS T 2304**. The NIS shall fulfil the applicable requirements of **JIS T 2304** including connection to other equipment.
- t) The risk analysis shall take into consideration the use of alarms, as appropriate, as described in **IEC 60601-1-11**.
- u) Adverse effects of the medicinal product contact with the NIS shall be assessed and mitigated through risk assessment.
- v) Biological requirements of the NIS shall be established in accordance with **JIS T 0993-1**.

NOTE It is preferable that the design process incorporate environmentally conscious design (see **IEC 60601-1-9 [4]**).

- w) Where requirements in this Standard provide a test method without acceptance criteria, the manufacturer shall establish a specification and acceptance criteria appropriate for the intended use of the device, using a risk-based approach (consistent with **JIS T 14971** and **JIS T 62366-1**).

6 Reagent and apparatus

6.1 General

Any suitable test system can be used, when the required accuracy (calibration) and precision (Gauge R&R) can be obtained. The repeatability and reproducibility (Gauge R&R) of the test apparatus shall be no greater than 20 % of the allowed tolerance range for any given measurement. For destructive test measurements, the Gauge R&R shall be no greater than 30 % of the allowed tolerance range. At a minimum, the Gauge R&R should cover ± 2 standard deviations (thereby covering approximately 95 % or more of the variation).

Example A measurement system with a measurement specification limit of ± 0.01 ml (range of 0.02 ml) comes out of the Gauge R&R with a Gauge R&R/tolerance range ratio of 20 %, which means that the Gauge R&R (4 standard uncertainties) equals $0.02 \text{ ml}/5 = 0.004 \text{ ml}$. The uncertainty of the measurement is ± 2 standard deviations (see **ISO/IEC Guide 98-1**), which equals 0.002 ml.

All doses, V_{set} , delivered are recorded gravimetrically, G_{meas} , (expressed in grams). These recordings are converted to volumes, V_{meas} , by using the density, ρ , (expressed in grams per millilitre) for the test liquid at environmental conditions. The following equation can be used to convert gravimetric measurements to volumetric:

$$V_{\text{meas}} = \frac{G_{\text{meas}}}{\rho} \dots\dots\dots (1)$$

6.2 Test liquid The test liquid is the original medicinal product intended to be injected by the NIS, or a liquid with similar physical properties.