



JAPANESE
INDUSTRIAL
STANDARD

Translated and Published by
Japanese Standards Association

JIS T 3239 : 2011
(JMED/JSA)

**Catheters for drainage use in the
stomach and esophagus**

ICS 11.040.25

Reference number : JIS T 3239 : 2011 (E)

PROTECTED BY COPYRIGHT

Date of Establishment: 2005-03-25

Date of Revision: 2011-07-29

Date of Public Notice in Official Gazette: 2011-07-29

Investigated by: Japanese Industrial Standards Committee
Standards Board
Technical Committee on Medical Equipment

JIS T 3239 : 2011, First English edition published in 2012-03

Translated and published by: Japanese Standards Association
4-1-24, Akasaka, Minato-ku, Tokyo, 107-8440 JAPAN

In the event of any doubts arising as to the contents,
the original JIS is to be the final authority.

© JSA 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Printed in Japan

PROTECTED BY COPYRIGHT

Contents

	Page
Introduction	1
1 Scope	1
2 Normative references.....	1
3 Terms and definitions	1
4 Configuration and name of each part	2
4.1 General	2
4.2 Tip part	2
4.3 Part at hand.....	2
5 Requirements.....	4
5.1 Tensile strength	4
5.2 Airtightness	4
5.3 Pressure resistance.....	5
5.4 Appearance and cleanliness.....	5
5.5 Biological safety	5
5.6 Assurance of sterility	5
6 Packaging	5
6.1 Primary container.....	5
6.2 Secondary container	5
7 Labelling	5
7.1 Shaft	5
7.2 Primary container.....	5
7.3 Secondary container	6
7.4 Use of symbols	7
Annex A (normative) Test for tensile strength	8
Annex B (normative) Test for leakage by pressurization	10

Foreword

This translation has been made based on the original Japanese Industrial Standard revised by the Minister of Health, Labour and Welfare through deliberations at the Japanese Industrial Standards Committee as the result of proposal for revision of Japanese Industrial Standard submitted by Japan Medical Devices Manufacturers Association (JMED)/Japanese Standards Association (JSA) with the draft being attached, based on the provision of Article 12 Clause 1 of the Industrial Standardization Law applicable to the case of revision by the provision of Article 14.

Consequently **JIS T 3239 : 2005** is replaced with this Standard.

This **JIS** document is protected by the Copyright Law.

Attention is drawn to the possibility that some parts of this Standard may conflict with a patent right, application for a patent after opening to the public, or utility model right which have technical properties. The relevant Minister and the Japanese Industrial Standards Committee are not responsible for identifying the patent right, application for a patent after opening to the public or utility model right which has the said technical properties.

Catheters for drainage use in the stomach and esophagus

Introduction

This Japanese Industrial Standard is the standard which is revised so that the contents such as terms and constitution of documents are altered for the convenience of users in the review of **JIS T 3239** established in 2005.

1 Scope

This Standard specifies the catheters for drainage use in the stomach and esophagus (hereafter referred to as “catheters”), which are designed to be nasogastrically or orally inserted into and indwelled in the stomach and esophagus for the purpose of aspiration, drainage, evacuation, cleaning and removal of foreign objects, etc.

This Standard does not include requirements for X-ray permeability of catheters.

Moreover, **JIS T 3239 : 2005** may be applied until July 28th, 2014.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this Standard. For standards with the year indication, only the edition of the indicated year shall apply and the revisions (including amendments) made thereafter shall not apply.

JIS T 0307 : 2004 *Medical devices—Symbols to be used with medical device labels, labelling and information to be supplied*

JIS T 0993-1 : 2005 *Biological evaluation of medical devices—Part 1: Evaluation and testing*

ISO 594-1 : 1986 *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment—Part 1: General requirements*

3 Terms and definitions

For the purposes of this Standard, the following terms and definitions apply:

3.1 catheters for drainage use in the stomach and esophagus

medical device which is composed of flexible tube and designed so that it can be nasogastrically or orally inserted into and indwelled in the stomach and esophagus for the purpose of aspiration, drainage, evacuation, cleaning and removal of foreign objects, etc.

3.2 tip part

tip of catheters to be inserted into the body