

JIS

JAPANESE INDUSTRIAL STANDARD

Standard Atmospheric Conditions for Testing

JIS Z 8703—1983

Translated and Published

by

Japanese Standards Association

Printed in Japan

AVA Documents
Document Sharing Hub

Translation without guarantee
In the event of any doubt arising, the original
standard in Japanese is to be evidence

JAPANESE INDUSTRIAL STANDARD

J I S

Standard Atmospheric Conditions for Testing

Z 8703-1983

1. Scope

This Japanese Industrial Standard specifies the standard atmospheric conditions concerning the temperature, humidity and atmospheric pressure in the place where the tests (including the conditioning for the tests) in the mining and manufacturing industries are carried out, provided that the atmospheric conditions for the testing of things that are required to be tested in special atmospheric conditions are not included.

2. Standard Atmospheric Conditions

2.1 Temperature of Standard Atmospheric Conditions The temperature of the standard atmospheric conditions shall be 20°C, 23°C or 25°C, according to the purpose of testing.

Corresponding International Standards:

ISO 554-1976 Standard atmospheres for conditioning and/or testing-Specifications

IEC Publication 160-1963 Standard atmospheric conditions for test purposes