

KES

Checking standard for Strength Tester machine

KES Q – B205A

(EST) 1986-08-08

(REV) 1987-04-21

Contents

- 1.Application Range**
- 2.Checking Standard**
- 3.Record and Control**

Additional Rules



1. Application

This standard is described the checking of accuracy for Dynamic Toe-in tester.

2. Checking standard

It will be followed sheet 1.

3. Record and control

It will be followed KMS33900. (prescription of accuracy control for measurement equipment)

Additional rules

This manual will be carried out on established date and revised date.

Sheet 1 (Checking method)

No.	Checking item	Method	Figure	Judgement
1	Appearance	<ol style="list-style-type: none"> 1. To be cleared mark and marling and check defect of scale by eyes. 2. Check painting and plating condition at each point by eyes. 3. Check shape and grinding condition at each point by eyes. 		No defect
2	Operation	<ol style="list-style-type: none"> 1. To be equipped pulling equipment for test piece with loading, engage equipment for test piece and indicator for given load or recording equipment. And, To be boar maximum capacity. 2. For the applying equipment for load, it has to smoothly give load up to maximum weight on test piece. 		No defect