



Correlation Test Method by Sham Gas

KES T – D606A

(EST) 1986-10-30

(REV) 1988-08-26

Contents

1. Purpose
2. Scope of Application
3. Order of Correlation Test Method by Sham Gas

Supplementary Provision



1. Purpose

This shall intend to explain the Correlation Test Method by Sham Gas.

2. Scope of Application

This shall regulate the Correlation Test Method by Sham Gas.

3. Order of Correlation Test Method by Sham Gas

Just analyzer shall be suitable for leak standard of sample line.

Order of test	Notes
(1) Proceed the leak test of the bag to be used. (2) Attach Pressue adjusting gauge for check to Sham gas bombe. (3) Set CPU. (4) Proceed Zero Span Adjustment of T.HC 25 PPMC, CO 100PPM, CO ₂ 1% of analyzer with CPU. (5) Connect bag to bombe of content (2), inject gas for 5 seconds by opening valve of bombe. (6) Connect bag of content (5) to bag selector and exhaust gas completely with exhaust mode (7) Connect bag to bombe of content (2). (8) Open valve of bombe and inject gas into bag for 30 seconds. (9) Connect bag to analyzer and proceed Meas with CPU. (10) when the pointer becomes stable record by reading pointed value of every component. (11) Print out value of CPU.	Be suitable leak standard. Use Pressue adjusting gauge for leak test of sample line. NOx 50 PPM Distinguish bombe Do not make selection of bag