



National Accreditation Board for Testing and Calibration Laboratories (NABL)

Specific Criteria for Site Testing and Site Calibration Laboratories

ISSUE NO.: 06
ISSUE DATE: 06-Feb-2020

AMENDMENT NO.: --
AMENDMENT DATE: --

AMENDMENT SHEET

S. No.	Page No.	Clause No.	Date of Amendment	Amendment	Reasons	Signature QA Team	Signature CEO
1							
2							
3							
4							
5							
6							
7							
8							
9							
10							

CONTENTS

S. No.	Title	Page No.
	Amendment Sheet	1
	Contents	2
1	Introduction	3
2	Scope	3
3	Definitions	4
4	Management Requirements	5
5	Technical Requirements	6
5.1	Personnel	6
5.2	Accommodation and Environment	6
5.3	Equipment and Reference Material	7
5.4	Calibration and Test Methods	8
5.5	Certificates and Reports	8
	Appendix A	9

1. INTRODUCTION

NABL has developed “Specific Criteria” in the fields of testing and calibrations (on requirements basis), to provide information on the additional requirements for each field/ discipline of accreditation. Laboratories seeking accreditation for site testing/ calibration must also meet the requirements outlined in the Specific Criteria relevant to that field /discipline (wherever applicable).

This specific criteria document is intended for the laboratories/ entities who are engaged in site testing or site calibration.

2. SCOPE

This specific criteria document describes additional accreditation requirements specifically applicable to the laboratories performing tests or calibrations at customer's premises (site) or at a location temporarily set up for a defined duration away from the permanent laboratory or in a mobile facility. Laboratories/ entities may be accredited for site testing or calibration regardless of their affiliation with an accredited “permanent laboratory”, as defined in Section 4 of this document. However, if any laboratory is accredited only for site testing / site calibration; the accreditation status of the site testing/ site calibration laboratory shall not be used to claim accreditation at the permanent laboratory i.e. the permanent laboratory (if any) shall not be considered as one of the sites. Further in such cases the permanent laboratory/entity shall apply for accreditation of its permanent laboratory separately.

National Accreditation Board for Testing and Calibration Laboratories				
Doc. No.: NABL 130	Specific Criteria for Site Testing and Site Calibration Laboratories			
Issue No.: 06	Issue Date: 06-Feb-2020	Amend. No.: --	Amend. Date: --	Page No: 3 of 9