



BSI Standards Publication

Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for circulating tumour cells (CTCs) in venous whole blood

Part 1: Isolated RNA

National foreword

This Published Document is the UK implementation of CEN ISO/TS 7552-1:2024. It is identical to ISO/TS 7552-1:2024. It supersedes PD CEN/TS 17390-1:2020, which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/212, IVDs.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

This publication is not to be regarded as a British Standard.

© The British Standards Institution 2024
Published by BSI Standards Limited 2024

ISBN 978 0 539 24372 7

ICS 11.100.10

Compliance with a Published Document cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 November 2024.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

English Version

Molecular in vitro diagnostic examinations - Specifications
for pre-examination processes for circulating tumour cells
(CTCs) in venous whole blood - Part 1: Isolated RNA
(ISO/TS 7552-1:2024)

Analyses de diagnostic moléculaire in vitro -
Spécifications relatives aux processus préanalytiques
pour les cellules tumorales circulantes (CTC) dans le
sang total veineux - Partie 1: ARN isolé (ISO/TS 7552-
1:2024)

Spezifikationen für präanalytische Prozesse für
zirkulierende Tumorzellen (CTC) in venösen
Vollblutproben - Teil 1: Isolierte RNA (ISO/TS 7552-
1:2024)

This Technical Specification (CEN/TS) was approved by CEN on 8 November 2024 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (CEN ISO/TS 7552-1:2024) has been prepared by Technical Committee ISO/TC 212 "Medical laboratories and in vitro diagnostic systems" in collaboration with Technical Committee CEN/TC 140 "In vitro diagnostic medical devices" the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TS 17390-1:2020.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO/TS 7552-1:2024 has been approved by CEN as CEN ISO/TS 7552-1:2024 without any modification.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General considerations	5
5 Activities outside the laboratory	6
5.1 Specimen collection	6
5.1.1 General	6
5.1.2 Information about the specimen donor/patient	6
5.1.3 Selection of the venous whole blood collection tube by the laboratory	7
5.1.4 Venous whole blood specimen collection from the patient/donor	7
5.2 Specimen storage and transport	8
5.2.1 General	8
5.2.2 Storage and transport using blood collection tubes with stabilizers	8
5.2.3 Storage and transport using blood collection tubes without stabilizers	9
6 Activities inside the laboratory	9
6.1 Specimen reception	9
6.2 Specimen storage after transport and reception	9
6.3 Enrichment of CTCs	9
6.3.1 General	9
6.3.2 Using a commercial CTC enrichment system intended for diagnostic use	10
6.3.3 Using the laboratory-developed CTC enrichment procedure	10
6.4 Quality of enriched CTCs	11
6.5 Storage of enriched CTCs	11
6.6 Isolation of CTCs	11
6.6.1 General	11
6.6.2 Using a commercial CTC isolation system intended for diagnostic use	11
6.6.3 Using the laboratory-developed CTC isolation procedure	12
6.7 Isolation of RNA from an enriched CTC sample	12
6.7.1 General	12
6.7.2 Using a commercial RNA isolation kit intended for diagnostic use	12
6.7.3 Using a laboratory-developed CTC RNA isolation procedure	13
6.8 Quantity and quality assessment of isolated RNA from enriched or isolated CTCs	13
6.8.1 General	13
6.8.2 Quantity assessment of CTC RNA	13
6.8.3 Quality assessment CTC RNA	14
6.9 Storage of isolated RNA from enriched CTCs	14
6.9.1 General	14
6.9.2 Storage of RNA isolated with a commercially available kit intended for diagnostic use	15
6.9.3 Storage of RNA isolated with the laboratory-developed procedure	15
Annex A (informative) Decision guideline for critical steps of the CTC pre-analytical workflow	16
Bibliography	18

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 212, *Medical laboratories and in vitro diagnostic systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 140, *In vitro diagnostic medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 7552 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Solid tumours release cells and bioanalytes into blood and other body fluids. This has opened the option of utilizing such body fluids (liquid biopsies) for a minimally-invasive procedure for tumour detection, diagnosis and characterization. Liquid biopsies can enable earlier detection and diagnosis of cancers and advance personalized patient treatment.^[20,21]

These applications have become one of the fastest growing segments of the entire diagnostic market.

Circulating tumour cells (CTCs) in venous whole blood can reflect the disease complexity that evolves during tumour progression, with distinct genetic, epigenetic and gene expression biomarkers.^[41]

Besides the prognostic role of CTC identification and enumeration in cancer progression, CTC molecular characterization can improve disease outcome prediction, therapeutic guidance and post-treatment monitoring of the patient.^[39]

CTCs are now considered as a surrogate of tumour tissue in cancer early development, progression and metastatic phase^[23].

Molecular characterization of CTCs can provide a strategy for monitoring cancer genotypes during systemic therapies,^[24] identifying mechanisms of disease progression, identifying novel targets for biological treatment^[25] and selecting targeted therapies^[39].

Moreover, CTC single-cell sequencing is emerging as an important tool for tumour genomic heterogeneity analysis.^[26-28] CTCs are fragile and tend to degrade within a few hours when collected in conventional blood collection tubes, e.g. EDTA containing tubes, without dedicated CTC stabilizers. CTCs are extremely rare, especially in early disease, e.g. less than 10 cells per 10 ml of blood, representing a ratio of approximately 1:10⁷ CTCs to white blood cells (WBCs). This low ratio represents a significant challenge to CTC enrichment required for examination.

RNA profiles of CTCs resemble gene expression profiles of tumours. For RNA profile analysis, measures to remove the WBCs are necessary in order to obtain sufficiently enriched CTC-specific RNA.

RNA profiles can change significantly after blood collection, during CTC enrichment and isolation. Therefore, special measures are necessary to obtain CTC samples of adequate quality and isolated RNA of appropriate quality for ensuring the specified RNA examination performance.^[29]

Standardization includes all steps of the pre-examination process, including blood collection and stabilization, transport, storage, CTC enrichment, CTC isolation (if included), and RNA isolation. This pre-examination standardization is crucial to ensure reliable examination results in current clinical use and is also critical to develop new CTC based diagnostic examinations and to establish these in clinical healthcare.^[30]

An illustration of critical steps of the CTC pre-analytical workflow is provided in [Annex A](#).

This document describes special measures to obtain appropriate quality and quantity of RNA from CTC-containing blood specimens for subsequent examination.

Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for circulating tumour cells (CTCs) in venous whole blood —

Part 1: Isolated RNA

1 Scope

This document specifies requirements and gives recommendations on the handling, storage, CTC enrichment and isolation, RNA isolation and storage, and documentation of venous whole blood specimens intended for the examination of RNA isolated from circulating tumour cells (CTCs) during the pre-examination phase before a molecular examination is performed.

This document is applicable to molecular in vitro diagnostic examinations including laboratory-developed tests performed by medical laboratories. It is also intended to be used by laboratory customers, in vitro diagnostics developers and manufacturers, biobanks, institutions, and commercial organizations performing biomedical research, and regulatory authorities.

This document does not cover the isolation of cellular RNA directly from venous whole blood containing CTCs. This is covered in ISO 20186-1.

This document does not cover the isolation of specific white blood cells and subsequent isolation of cellular RNA therefrom. This document does not cover pre-analytical workflow requirements for viable CTC cryopreservation and culturing.

NOTE 1 The requirements given in this document can also be applied to other circulating rare cells (e.g. foetal cells).

NOTE 2 International, national or regional regulations or requirements can also apply to specific topics covered in this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15189, *Medical laboratories — Requirements for quality and competence*

ISO 15190, *Medical laboratories — Requirements for safety*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 15189 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>