PD ISO/TR 16840-9:2015

BSI Standards Publication

Wheelchair seating

Part 9: Clinical interface pressure mapping guidelines for seating

bsi.

...making excellence a habit.™

Document Sharing Hub

Full document details: PD ISO/TR 16840-9:2015

National foreword

This Published Document is the UK implementation of ISO/TR 16840-9:2015.

The UK participation in its preparation was entrusted to Technical Committee CH/173/1, Wheelchairs.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2015. Published by BSI Standards Limited 2015

ISBN 978 0 580 87406 2

ICS 11.180.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 July 2015.

Amendments issued since publication

Date	Text affected

Documents
Document Sharing Hub

Full document details: PD ISO/TR 16840-9:2015

TECHNICAL REPORT

ISO/TR 16840-9

First edition
2015-07-01

Wheelchair seating —

Part 9: Clinical interface pressure mapping guidelines for seating

Sieges de fauteuils roulants —

Partie 9: Lignes directrices pour l'utilisation d'un système de mappage de pression

Reference number
ISO/TR 16840-9:2015(E)

Documents
Document Sharing Hub

© ISO 2015

Full document details: PD ISO/TR 16840-9:2015

COPYRIGHT PROTECTED DOCUMENT

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

© ISO 2015 – All rights reserved

Documents
Document Sharing Hub

Full document details: PD ISO/TR 16840-9:2015

Contents

Page

Foreword ..... **v**

Introduction ..... **vi**

1 Scope **1**

2 Definitions and glossary **1**
2.1 Calibration 1
2.2 Coefficient of Variation (CoV) 1
2.3 Conformity 1
2.4 Contact area 2
2.5 Creep 2
2.6 Dispersion Index (DI) 2
2.7 Envelopment 2
2.8 Fomite 2
2.9 Hysteresis error 2
2.10 Immersion 3
2.11 Pressure, stresses, and forces 3
2.11.1 Axial strain (ε) 4
2.11.2 Perpendicular force (FP) 4
2.11.3 Pressure (p) 4
2.11.4 Shear force (FS) 5
2.11.5 Shear stress (τ) 5
2.11.6 Shear strain (γ) 5
2.11.7 Friction force 5
2.12 Sensor saturation 6

3 Applications **6**
3.1 Cushion comparison 6
3.2 Wheelchair set-up 6
3.3 Client/caregiver education 7
3.4 Dynamic activity 7
3.5 Relevance to activities of daily living 7

4 Test protocol in a clinic **7**
4.1 Infection control 8
4.1.1 Wash hands 8
4.1.2 Avoid cross-infection 8
4.1.3 Protect the mat 8
4.2 Setting up 8
4.2.1 Set up mat 8
4.2.2 Prepare test protocol 8
4.2.3 Prepare client file 8
4.2.4 Be consistent 9
4.2.5 Check mat orientation 9
4.2.6 Check mat position 9
4.2.7 Smooth mat into cushion 9
4.2.8 Acclimatize test cushions 9
4.3 Transferring the client and getting the client into position 9
4.3.1 Transfer boards 9
4.3.2 Remove unnecessary items 9
4.3.3 Post-transfer checks 10
4.3.4 Consistency between tests 10
4.3.5 Settling time 10
4.4 Introducing pressure mapping to the client 10
4.4.1 Position of the screen 10
4.5 Capturing data 10
4.5.1 Identify position of bony prominences 10

AVA Documents
Document Sharing Hub

© ISO 2015 – All rights reserved

Full document details: PD ISO/TR 16840-9:2015

4.5.2 Record one or more images ... 11
4.5.3 Repeat the process ... 11
4.5.4 Check the images ... 11
4.5.5 Validate the images ... 11

5 Documentation ... 11
5.1 File naming protocol ... 12
5.2 File location ... 12
5.3 Make notes immediately ... 12
5.4 Reporting ... 12
5.5 Photo documentation ... 12
5.6 Data back-up ... 12

6 Interpretation ... 12
6.1 An unexpected image ... 13
6.2 An image that is not ideal ... 13
6.3 The image shows nothing abnormal ... 13
6.4 IPM views and metrics ... 13
6.4.1 Isobar (2-D)/contour view ... 13
6.4.2 3-D/surface view ... 14
6.4.3 Centre of pressure (CoP) ... 15
6.4.4 Gradient ... 15
6.4.5 Asymmetry ... 15
6.4.6 Interpolation ... 16
6.4.7 Smoothing ... 16

7 Limitations ... 16
7.1 Prediction of pressure injuries ... 16
7.2 Microclimate ... 17
7.3 Clothing ... 17
7.4 Repeatability and relative values ... 17
7.5 Influence of posture ... 17
7.6 Sampling frequency ... 17

Bibliography ... 18

© ISO 2015 – All rights reserved
Documents
Document Sharing Hub
Full document details: PD ISO/TR 16840-9:2015

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

The committee responsible for this document is ISO/TC 173, *Assistive products for persons with disability*, Subcommittee SC 1, *Wheelchairs*.

ISO 16840 consists of the following parts, under the general title *Wheelchair seating*:

— *Part 1: Vocabulary, reference axis convention and measures for body segments, posture and postural support surfaces*

— *Part 2: Determination of physical and mechanical characteristics of devices intended to manage tissue integrity — Seat cushions*

— *Part 3: Determination of static, impact and repetitive load strengths for postural support devices*

— *Part 4: Seating systems for use in motor vehicles*

— *Part 6: Simulated use and determination of the changes in properties of seat cushions*

— *Part 9: Clinical interface pressure mapping guidelines for seating* [Technical Report]

— *Part 10: Resistance to ignition of non-integrated seat and back support cushions — Requirements and test methods*

— *Part 11: Determination of perspiration dissipation characteristics of seat cushions intended to manage tissue integrity* [Technical Specification]

— *Part 12: Apparatus and method for cushion envelopment testing* [Technical Specification]

© ISO 2015 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TR 16840-9:2015

Introduction

The purpose of this Technical Report is to provide information as to where interface pressure mapping (IPM) can fit into a clinical assessment of an individual and their seating system, to highlight what can be achieved, and bring awareness of the limitations.

The use of IPM in the clinic is increasingly used to support clinicians in the evaluation of individual seating systems for individual clients. To get the best value from IPM requires a knowledge of the basic concepts of the interface pressure distribution (see Clause 2), the ability to define and apply a correct protocol for the application of IPM (see Clause 4), the ability to prepare correct documentation of the IPM (see Clause 5), and the skill in interpreting the collected data (see Clause 6).

The aim of an IPM test session can be different in a clinical or rehabilitation environment. In some cases, it can be a tool to compare the behaviour of different pressure care or postural cushions. It can also be used to support the clinician in finding the best match between cushion and client. Various applications looking at the broader picture are covered in Clause 3. Whichever the application, the main objective behind using an IPM needs to be clear from the beginning.

© ISO 2015 – All rights reserved
Documents
Document Sharing Hub
Full document details: PD ISO/TR 16840-9:2015

Wheelchair seating —

Part 9: Clinical interface pressure mapping guidelines for seating

1 Scope

This Technical Report has been produced to guide users in the performance of the tasks that are directly involved in the clinical use of interface pressure mapping (IPM) or are synergistic with its use in a comprehensive wheelchair seating evaluation.

This Technical Report does not cover other aspects of the clinical assessment process (e.g. taking a medical history), nor the prescription or treatment process which might arise from an assessment. These guidelines are not meant to be a substitute for clinical reasoning and judgement within the context of a complete assessment.

This Technical Report refers to the state of the art of IPM experiences in a seating scenario. Most of the principles covered can be extrapolated to whole body (in bed) or to foot assessments, for example.

2 Definitions and glossary

2.1 Calibration

Calibration is a process wherein the sensing mat is subjected to known forces. The sensor responses are monitored and modelled in the software.

NOTE A record is kept of the responses (called a calibration file) and whenever the sensors output a similar response, the result is related to the previously known forces. In most cases, this is done by placing the mat in a purpose-built chamber with an air-filled bladder. The bladder is inflated and the pressure in the bladder is measured. It is assumed to be evenly pressurized over the mat.

Calibration allows for the software to accommodate changes with time (creep) or pressure (hysteresis) exhibited by the sensors.

Recalibrate the mat whenever the readings look unreliable, after excessive use, or at the manufacturer's recommended interval. Keep track of the uses of the mat and the date of the last calibration. Old calibration files should be retained (old calibration files can be loaded for comparison to determine change over a period).

2.2 Coefficient of Variation (CoV)

The CoV is expressed as a percentage:

$$Coefficient\ of\ Variation = \frac{Standard\ deviation}{Mean} \tag{1}$$

NOTE This is one of the statistical measures available to assess how evenly the pressure is distributed across a support surface. The lower the CoV, the lower the variability in the data set.

2.3 Conformity

The ability of the IPM mat to adapt to irregular shapes without creasing.

© ISO 2015 – All rights reserved

Documents
Document Sharing Hub

Full document details: PD ISO/TR 16840-9:2015