PD ISO/TS 9320:2024

BSI Standards Publication

Health informatics — Standardized data set for transfer of hemodialysis patients

Documents
Document Sharing Hub

Full document details: PD ISO/TS 9320:2024

National foreword

This Published Document is the UK implementation of ISO/TS 9320:2024.

The UK participation in its preparation was entrusted to Technical Committee IST/35, Health informatics.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

This publication is not to be regarded as a British Standard.

© The British Standards Institution 2024
Published by BSI Standards Limited 2024

ISBN 978 0 539 22476 4

ICS 35.240.80

Compliance with a Published Document cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 August 2024.

Amendments/corrigenda issued since publication

Date	Text affected

Documents
Document Sharing Hub

Full document details: PD ISO/TS 9320:2024

Technical Specification

ISO/TS 9320

First edition
2024-08

Health informatics — Standardized data set for transfer of hemodialysis patients

Informatique de santé — Ensemble de données normalisées pour le transfert des patients en hémodialyse

Reference number
ISO/TS 9320:2024(en)

© ISO 2024

Documents
Document Sharing Hub

Full document details: PD ISO/TS 9320:2024

COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

Contents

Page

Foreword ..... **v**
Introduction ..... **vi**
1 Scope **1**
2 Normative references **1**
3 Terms, definitions and abbreviations **1**
3.1 Terms and definitions 1
3.2 Abbreviated terms 5
4 Significance of haemodialysis data set **5**
4.1 General 5
4.2 Unique elements in the haemodialysis data set 6
4.3 Purpose of a standardized data set 6
5 Use case **7**
5.1 General 7
5.2 Haemodialytic patient with no special problems 7
5.3 HIV positive haemodialytic patient 7
5.4 Haemodialytic patient with refractory heart failure 7
6 Data sets for transfer of haemodialysis **8**
6.1 General 8
6.2 Fields of required data 10
6.2.1 General 10
6.2.2 Last dialysis date 10
6.2.3 Haemodialysis interval 11
6.2.4 Dry weight 11
6.2.5 Haemodialysis machine model 12
6.2.6 Haemodialysis blood flow rate 12
6.2.7 Haemodialysis access type 12
6.2.8 Haemodialysis access site 13
6.2.9 Haemodialysis access status 13
6.2.10 Hepatitis type B antigen / antibody 14
6.2.11 Hepatitis C virus infection 14
6.2.12 Human immunodeficiency virus infection 15
6.2.13 Dialysis duration 15
6.2.14 Dialysate 16
6.3 Fields of conditionally-required data 16
6.3.1 General 16
6.3.2 Initial dialysis date 16
6.3.3 Last laboratory test date 17
6.3.4 Blood type 17
6.3.5 Diagnosis 18
6.3.6 Blood haemoglobin concentration 18
6.3.7 Blood sodium concentration 19
6.3.8 Blood potassium concentration 19
6.3.9 Blood calcium concentration 20
6.3.10 Blood phosphate concentration 20
6.3.11 Venereal disease research laboratory 21
6.3.12 Pre-dialysis vital sign 22
6.3.13 Post-dialysis vital sign 22
6.3.14 Interdialytic weight gain 23
6.3.15 Prescription 23
6.3.16 Remark note 24
6.4 Fields of optional data 24
6.4.1 General 24
6.4.2 Heparin concentration 24

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

6.4.3 Remnant kidney function ... 25
6.4.4 Other laboratory results ... 25
6.4.5 Radiology report ... 26
6.4.6 Allergy ... 26
6.4.7 Presence of heart failure ... 27
6.4.8 Heart ejection fraction ... 27
6.4.9 Haemodialysis access construction date ... 28

Annex A (informative) **Example of a haemodialysis referral paper** ... **29**

Annex B (informative) **Example of an FHIR resource** ... **30**

Bibliography ... **53**

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 215, *Health informatics*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

Introduction

Globally, the population with end-stage renal disease (ESRD) has steadily increased. Haemodialysis, which provides rapid clearance of solutes, is the most popular renal replacement therapy for ESRD patients. However, haemodialysis has a high risk of serious complications for receivers of the treatment. Well-known complications of haemodialysis include hypotension, electrolyte imbalance, infection, fluid overload, and dialysis disequilibrium. Due to these side-effects, patients who receive haemodialysis have lower life expectancy than the general population.

Moreover, haemodialysis requires a patient to frequently visit a hospital and to go through complex and time-consuming dialysis procedures. This is because patients need to know how to care for the haemodialysis access or fistula, exercise, regulate diet and monitor their blood pressure and weight. In addition, haemodialysis patients are more likely to be depressed because they are limited in many areas of their daily activities.

Referral of haemodialysis cases occurs often due to complications of chronic kidney disease, such as myocardial infarction, heart failure, atrial fibrillation and cerebrovascular accidents. Recently, a notable change in these referral cases is the increase of referrals from abroad. One of the reasons behind this increase is that haemodialysis which is done outside of a person's residential area is covered by many payers in the United States. As a result, international travel for haemodialysis is growing fast in east Asian countries.

Since haemodialysis includes a complex procedure and requires attentive monitoring, it is imperative to share important and accurate information about a patient, dialysis setup and haemodialysis access to ensure safe and timely haemodialysis. This would also allow patients who are travelling abroad for a haemodialysis to experience smooth referral.

The purpose of this document is to define the data set for referral of haemodialysis patients. This data set will ensure continuity of haemodialysis-related care. The data set provides optimal dialysate parameters and individualized dialysis settings. Providing this information can reduce the occurrence of dialysis-related complications. The data set can be used for referring haemodialysis patients, for surgery or intensive care, and it can also support safe and timely haemodialysis procedures for ESRD patients from abroad.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

Health informatics — Standardized data set for transfer of hemodialysis patients

1 Scope

The document defines a data set for the safe and timely transfer of haemodialysis procedure for end-stage renal disease (ESRD) patients. The necessary information for dialysis is provided through required, conditionally-required and optional data fields of the data set. Complicated use cases are also described in this document.

This document does not cover general quality requirements or system requirements for haemodialysis.

2 Normative references

There are no normative references in this document.

3 Terms, definitions and abbreviations

3.1 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at https://www.iso.org/obp

— IEC Electropedia: available at https://www.electropedia.org/

3.1.1
haemodialysis
dialysis
therapy where waste solutes from the blood are removed by diffusion across an artificial extracorporeal semipermeable membrane

3.1.2
haemodialysis interval
currently prescribed interval between *haemodialysis* (3.1.1) sessions

Note 1 to entry: The interval determines how effectively waste products and fluids are cleared from the body.

Note 2 to entry: A longer interval can lead to accumulation of harmful body products and water, resulting in complications such as oedema, electrolyte imbalance and increased blood pressure.

3.1.3
dry weight
optimal or target weight that a patient should achieve during or after a treatment

Note 1 to entry: Dry weight is measured without the excess fluid that builds up between dialysis treatments.

Note 2 to entry: Dry weight is an important concept because it helps clinicians determine how much excess fluid needs to be removed during each *haemodialysis* (3.1.1) session.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

3.1.4
haemodialysis machine model
information about haemodialysis machine manufacturer and model used in recent *haemodialysis* (3.1.1) sessions

Note 1 to entry: The product name of *dialysate* (3.1.9) varies depending on the machine model (the manufacturer's product).

Note 2 to entry: The information helps clinicians prescribe dialysate in the receiving unit.

3.1.5
haemodialysis blood flow rate
rate at which a patient's blood is pumped through the dialysis machine during *haemodialysis* (3.1.1) treatment

Note 1 to entry: Haemodialysis blood flow rate is typically measured in millilitres per minute (ml/min) or litres per hour (l/h).

Note 2 to entry: Clinicians need to determine the appropriate blood flow rate for each individual undergoing haemodialysis to balance the benefits and potential risks.

3.1.6
haemodialysis access type
method or route by which blood is withdrawn from the patient's body, filtered through the dialysis machine, and then returned to the patient

Note 1 to entry: Common types of haemodialysis access are arteriovenous fistula, arteriovenous graft and central venous catheter.

3.1.7
haemodialysis access site
anatomical location of *haemodialysis* (3.1.1) access

Note 1 to entry: The preferred access site for an arteriovenous fistula is usually in the forearm.

3.1.8
haemodialysis access status
condition and functionality of the access site through which *haemodialysis* (3.1.1) treatments are performed

Note 1 to entry: Monitoring and assessing access status is necessary to ensure that the access site remains functional, safe and free from complications.

Note 2 to entry: Haemodialysis access status usually refers to patency, infection, maturation and functionality.

3.1.9
dialysate
fluid used in dialysis to exchange solutes with the blood

Note 1 to entry: Dialysate is used to draw fluids and toxins out of the bloodstream and supply electrolytes and other chemicals to the bloodstream

3.1.10
initial dialysis date
date on which a patient with end-stage renal disease (ESRD) begins *haemodialysis* (3.1.1) treatment for the first time.

Note 1 to entry: Initial dialysis date is a significant event for patients, as it marks the beginning of a lifelong or long-term commitment to regular dialysis treatments.

3.1.11
blood haemoglobin concentration
concentration of haemoglobin in whole blood

Note 1 to entry: Anaemia is a common complication in patients with chronic kidney disease.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

Note 2 to entry: Haemoglobin is an important determinant for iron supplementation, blood transfusion or nutritional assessment.

3.1.12
blood sodium concentration
concentration of sodium in serum or plasma

Note 1 to entry: Sodium is key for electrolytes in the body.

Note 2 to entry: *Haemodialysis* (3.1.1) patients are at risk of fluid and electrolyte imbalance due to impaired kidney function and haemodialysis process.

3.1.13
blood potassium concentration
concentration of potassium in serum or plasma

Note 1 to entry: Potassium is key for electrolytes in the body.

3.1.14
blood calcium concentration
concentration of calcium in serum or plasma

Note 1 to entry: Calcium is key for electrolytes in the body.

3.1.15
blood phosphate concentration
concentration of phosphate in serum or plasma

Note 1 to entry: Phosphate is key for electrolytes in the body.

3.1.16
venereal disease research laboratory
nontreponemal test result for screening of syphilis

3.1.17
pre-dialysis vital sign
vital sign measured before *haemodialysis* (3.1.1) session

Note 1 to entry: Vital signs can be subcategorized in “blood pressure”, “heart rate”, “body temperature” and “respiration rate.”

Note 2 to entry: Pre-dialysis vital signs help ensure patient’s safety and provide information about their readiness for haemodialysis.

3.1.18
post-dialysis vital sign
vital signs measured after *haemodialysis* (3.1.1) session

Note 1 to entry: Vital sign can be subcategorized in “blood pressure”, “heart rate”, “body temperature” and “respiration rate.”

Note 2 to entry: Post-dialysis vital signs are needed to assess haemodialysis response and to ensure patient’s safety.

3.1.19
interdialytic weight gain
amount of weight gain between two *haemodialysis* (3.1.1) sessions

Note 1 to entry: Interdialytic weight gain is the result of salt and water intake.

Note 2 to entry: Interdialytic weight gain helps clinicians assess how effectively the patient is managing fluid intake and output between dialysis sessions.

Note 3 to entry: Interdialytic weight gain is used to determine the target fluid removal for each dialysis session.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

3.1.20
remark note
note containing information or precautions regarding *haemodialysis* (3.1.1), disease management and general patient-care

Note 1 to entry: A remark note is usually provided in plain text.

3.1.21
heparin concentration
concentration of an anticoagulant in *haemodialysis* (3.1.1) fluid required to prevent thrombosis in catheter, filter and circuit

Note 1 to entry: Heparin concentration usually refers to heparin bolus dose as initial and maintenance dose.

3.1.22
remnant kidney function
renal working capacity remaining after *haemodialysis* (3.1.1)

Note 1 to entry: Remnant kidney function is usually measured as daily urine output or *glomerular filtration rate* (3.1.29).

3.1.23
other laboratory results
laboratory examination of a blood sample other than those mentioned among required and conditionally-required data set

3.1.24
radiology report
text that represents the interpretation of a radiological study

3.1.25
allergy
altered bodily reactivity (such as hypersensitivity) to an antigen in response to a first exposure

3.1.26
presence of heart failure
presence of a condition in which the heart cannot pump enough blood to meet the body's needs

Note 1 to entry: *Haemodialysis* (3.1.1) patients are at an elevated risk of cardiovascular complications.

Note 2 to entry: Managing heart failure is important for fluid management, blood pressure control and dialysis procedure.

3.1.27
heart ejection fraction
measure of cardiac function reflecting the average fraction of emptying of the left ventricle with each contraction

Note 1 to entry: Heart ejection fraction is a crucial indicator of heart function and is commonly assessed using echocardiography.

3.1.28
haemodialysis access construction date
procedure date of current *haemodialysis* (3.1.1) access

Note 1 to entry: Haemodialysis access construction date helps clinicians check the duration of access use and complications that can arise over time.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

3.1.29
glomerular filtration rate
calculated measure of renal function which is expressed by the total volume of fluid filtered through all renal glomeruli in a minute

Note 1 to entry: The glomerular filtration rate is usually corrected for estimated body surface area, and reported in ml/min/1,73 m^2.

3.2 Abbreviated terms

ESRD	end-stage renal disease
HD	haemodialysis
HBsAg	surface antigen of the hepatitis B virus
HBV	hepatitis B virus
HCV	hepatitis C virus
HIV	human immunodeficiency virus
LOINC	Logical Observation Identifiers Names and Codes
FHIR	Fast Healthcare Interoperability Resources
VDRL	venereal disease research laboratory
GFR	glomerular filtration rate
NYHA	New York Heart Association

4 Significance of haemodialysis data set

4.1 General

Haemodialysis is one method of renal replacement therapy for ESRD patients. Since ESRD patients do not have functioning kidneys, patients experience symptoms such as water retention and accumulation of waste metabolites. Typically, ESRD patients require extracorporeal blood circulation that connects to the patient's vascular structures (vascular access), which are artificially constructed in the body.

The basic principle of haemodialysis is to remove unnecessary body water and toxic wastes, such as urea and creatinine, by extracorporeal filtration of patients' blood. Haemodialysis is conventionally performed and managed by highly trained staff. Furthermore, many clinics have specialized haemodialysis facilities such as high-quality water purification system and electricity system. The facility should apply additional measures to prevent microbial contamination.

Although the number of haemodialysis medical institutions and haemodialysis machines is steadily increasing, it is difficult to manage the quality of haemodialysis treatment for various reasons. Patients' lack of awareness of their dialysis data and their non-compliance to medical staff's instructions are obstacles to the improvement of patient's life quality and the reduction of mortality in ESRD patients. Additionally, ESRD patients usually suffer from severe anaemia, malnutrition, hyperparathyroidism, hyperphosphatemia, hypocalcaemia, metabolic acidaemia, hypertension and congestive heart failure at the start of dialysis. Hypoalbuminemia, anaemia and left ventricular hypertrophy are common side effects of dialysis treatment. Sometimes dialysis results in premature death. Details of these cases should be communicated through standardized data set so that this ultimately leads to improvement in dialysis treatment and better prognosis for ESRD patients.[8],[9]

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

4.2 Unique elements in the haemodialysis data set

There are unique data elements that exist only for haemodialysis and these elements are sometimes misleading. Procedural data is often misunderstood as observational data. Caution is required as the meaning of a specific data element can be conveyed inaccurately. Some data sets do not use standard medical terminology. Table 1 provides unique data elements which exist only in the haemodialysis domain and often lack standard terminology.

Table 1 — Unique data elements in haemodialysis

Elements	Element description and its significance
Dry weight	— Target weight of haemodialysis without fluid overload or hypovolemia. — It should not be misinterpreted as common body weight.
Haemodialysis blood flow rate	— Blood flow in haemodialysis machine. — It should not be misinterpreted as the blood flow in the patient's body. — It represents setting value in the haemodialysis machine.
Haemodialysis access	— Body site to reach the blood for haemodialysis.
Dialysate	— Kind of chemical bath used in dialysis to draw fluids and toxins out of the blood stream. — It is not considered as medication data.
Interdialytic weight gain	— Amount of weight gain as a result of salt and water intake between two haemodialysis sessions. — It is used as value to calculate target weight during haemodialysis. — It should not be misinterpreted as common weight gain.
Haemodialysis machine model	— Information of haemodialysis machine manufacturer and model. — Importance of model is often neglected. — Standard terminology not given.

4.3 Purpose of a standardized data set

Safe and timely haemodialysis requires a complex set of data and it is practically unfeasible for a patient to remember each element. Therefore, it has been a common practice to carry paper-based data when they are transferred (see Annex A). A standard haemodialysis data set benefits both patients and medical staff by electronically transferring relevant and accurate dialysis data.

Moreover, a standard haemodialysis data set can provide safe dialysis to patients by reducing the possibility of haemodialysis-related complications. Haemodialysis is a care-intensive process that requires the utmost attention to prevent various complications. For example, if haemodialysis parameters are set inappropriately, too much water can be removed from patients which can lead to hypovolemia, hypotension and fatigue. Therefore, it is critical to have data from previous dialysis about end-dialysis weight and dialysis speed. Such information can be accurately and electronically transferred from one medical institution to another using the standardized data set proposed in this document.

Timely haemodialysis is important, because delays in haemodialysis schedule can lead to worsening uraemia, increased infection susceptibility and other preventable complications. Sometimes, medical staff need to additionally contact the medical institution that issued the referral because patients did not know all the necessary information for their haemodialysis. Since the standardized data set proposed in this document provides basic parameters to start and monitor haemodialysis procedure, it can reduce the time for medical staff to prepare a haemodialysis session.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

5 Use case

5.1 General

Haemodialysis transfer is a multidisciplinary process because it involves wide range of agenda, including educating a patient on diet and exercise and explaining inadvertent complications of haemodialysis to the patient's family members.

A transfer can be done for various reasons, for example if the patient needs a procedure unrelated to kidney issues or if they move to another state. It can also occur if the patient's family planned a family trip to a near-by country.

Reviewing the data elements in the proposed data set provides helpful information to medical staff, especially in the case of a haemodialysis patient who is planned to be transferred. For instance, the time between receiving the data and the patient's visit would help medical staff to evaluate risk factors for the upcoming transfer and to decide on appropriate transfer time. Sending the proposed data set to a recipient hospital is more than a simple data transfer because it helps a recipient hospital prepare for unexpected events that can occur at any step of the transfer process. Detailed information on the transfer process should be shared with the patient, and discussions can be necessary with patient's custodians.

Patients can have very different complication risks depending on their age, race and comorbidities. Some patients can develop repetitive complications by haemodialysis and can develop similar complications in the recipient hospital. Repetitive adverse events can be notified and can be prevented. Sometimes it is necessary to contact the recipient hospital to query whether all necessary medical equipment is ready for expected emergency.

5.2, 5.3 and 5.4 are intended to provide several use cases with different risks of complication.

5.2 Haemodialytic patient with no special problems

In this use case, the haemodialysis patient has had stable haemodialysis repeatedly. The patient is young. Vital signs are stable. No adverse event has been notable in repeated haemodialysis session. The patient can be transferred with only the required data set. Other fields are either conditionally-required or optional. The required, conditionally-required and optional field are further explained in Clause 6.

5.3 HIV positive haemodialytic patient

When a haemodialytic patient is HIV positive, additional measures should take place to minimize the risk of spreading HIV to other patients and clinic staff. HIV test result shall be included in the data set. Clinic staff are also required to share the information because careless contacts with the patient must be avoided. Also, a recipient haemodialysis clinic must provide a haemodialysis machine which is dedicated to HIV patients. Before referral, a physician can contact the recipient hospital for the availability of HIV haemodialysis. Since an HIV dedicated machine is equipped in limited numbers, it is imperative to check that schedules are not overlapping and that those machines are ready for the next scheduled haemodialysis.

5.4 Haemodialytic patient with refractory heart failure

Heart failure and renal dysfunction can coexist. The number of comorbidities increases over time. A critical issue with this combination is that each disease can exacerbate another. ESRD can lead to retention of water which can cause cardiac dysfunction. ESRD can also exacerbate heart failure progressively and finally make it refractory to treatment. A possible treatment approach can be to address the treatable cause of valvular heart disease. Physicians should deal with the volume of excess fluid. Patients are requested to reduce the dietary intake of salt during the transfer process. The data set which is related to heart failure shall be transmitted before transfer process. The recipient clinic shall make a decision on whether they accept the transfer. Medication should be also monitored and approximated because renal dysfunction commonly limits the use of medication on cardiac problems.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

6 Data sets for transfer of haemodialysis

6.1 General

In this document, data set types are categorized as required, conditionally-required and optional. Data sets are also categorized according to when they are needed, so they are defined as needed before, during or after the haemodialysis procedure. Data sets are also categorized according to the importance of the dialysis procedure, as the focus is on the dialysis procedure rather than general health. Therefore, they may differ from general healthcare data sets. Table 2 shows the overall data tabulated and accompanied by the necessary metadata, including a glossary of terms and brief examples with code and value property. Also, more detailed exemplary data are provided in Annex B.

Table 2 — data elements for haemodialysis

Attribute name/Data element name	Coding system	Code value	Data value type	Qualifier[a]
Last dialysis date	Applicable terminology		DateTime	R
Haemodialysis interval	LOINC	LOINC 50951-9 Interdialytic time (ESRD)	DateTime	R
Dry weight	LOINC	LOINC 8341-0 Dry body weight measured	kg, lb	R
Haemodialysis machine model			Text	R
Haemodialysis blood flow rate	LOINC	LOINC 99711-4 Blood flow rate Renal replacement therapy circuit	ml/min	R
Haemodialysis access type	LOINC	LOINC 72050-8 Dialysis access	LOINC Answer List LL2169-2 ESRD-dialysis access type	R
Haemodialysis access site	LOINC	LOINC 99715-5 Dialysis access site	SNOMED Body Structure	R
Haemodialysis access status	LOINC	LOINC 99716-3 Dialysis access site appearance	LOINC Answer List LL2150-8 Access site apperance	R
Hepatitis type B antigen / antibody	LOINC	LOINC 5195-3 Hepatitis B virus surface Ag (presence) in serum	LOINC Answer List LL3865-4 True|-False|Unknown	R
Hepatitis C virus infection	LOINC	LOINC 11259-9 Hepatitis C virus RNA (presence) in serum or plasma	LOINC Answer List LL3865-4 True|-False|Unknown	R
Human immunodeficiency virus infection	LOINC	LOINC 7917-8 HIV 1 Ab (Presence) in serum	LOINC Answer List LL3865-4 True|-False|Unknown	R
Dialysis duration	LOINC	LOINC 68489-4 Dialysis hours per session	Hours	R
Dialysate	LOINC	LOINC 99732-0 Dialysate fluid renal replacement therapy circuit	Text	R
Initial dialysis date	Applicable terminology		DateTime	C (6.3.2.2)
Last laboratory test date	Applicable terminology		DateTime	C (6.3.3.2)

[a] R = required; C = conditionally required; O = optional.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

Table 2 *(continued)*

Attribute name/Data element name		Coding system	Code value	Data value type	Qualifier[a]
Blood type		LOINC	LOINC 882 ABO and Rh group (type) in blood	LOINC Answer List LL2972-9 ABORh	C (6.3.4.2)
Diagnosis		ICD 10/11, SNOMED Applicable Terminology	ICD10 M47 Spondylosis arthrosis or osteoarthritis of spine	Code, List of Codes	C (6.3.5.2)
Blood haemoglobin concentration		LOINC	LOINC 42810-2 Haemoglobin (mass/volume) in blood	g/dl	C (6.3.6.2)
Blood sodium concentration		LOINC	LOINC 2951-2 Sodium (moles/volume) in serum or plasma	mmol/l	C (6.3.7.2)
Blood potassium concentration		LOINC	LOINC 2823 Potassium (moles/volume) in serum or plasma	mmol/l	C (6.3.8.2)
Blood calcium concentration		LOINC	LOINC 2000-8 Calcium (moles/volume) in serum or plasma	mmol/l	C (6.3.9.2)
Blood phosphate concentration		LOINC	LOINC 2777-1 Phosphate (mass/volume) in serum or plasma	mg/dl	C (6.3.10.2)
Venereal disease search laboratory		LOINC	LOINC 5292-8 Reagin Ab in serum	LOINC Answer List LL3865-4 True|-False|Unknown	C (6.3.11.2)
Pre/post-dialysis vital sign	Systolic Blood pressure	LOINC	LOINC 8480-6 Systolic blood pressure	mmHg	C (6.3.12.2, 6.3.13.2)
	Diastolic blood pressure	LOINC	LOINC 8462-4 Diastolic blood pressure	mmHg	
	Heart rate	LOINC	LOINC 8867-4 Heart rate	/min	
	Respiration rate	LOINC	LOINC 9279-1 Respiratory rate	/min	
	Body temperature	LOINC	LOINC 8310-5 Body temperature	Degrees Celsius, Fahrenheit °C, °F	
Interdialytic weight gain		LOINC	LOINC-74006-8 Weight difference: pre-dialysis vs. post-dialysis	Numeric (kg, lb)	C (6.3.14.2)
Prescription		RxNorm, SNOMED	RxCUI 385629 Acetaminophen 5 mg/ml	Code/ List of Codes	C (6.3.15.2)
Remark note		Applicable terminology	Not determined	Text	C (6.3.16.2)
Heparin concentration		Applicable terminology	Not determined	Heparin unit/dose	O
Remnant kidney function		LOINC	LOINC 9188-4 urine output 1 hour	ml/h	O

[a] R = required; C = conditionally required; O = optional.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

Table 2 *(continued)*

Attribute name/Data element name	Coding system	Code value	Data value type	Qualifier[a]
Other laboratory result	LOINC / SNOMED applicable terminology	LOINC 789-8 erythrocytes in blood by automated count	Numeric /nominal /ordinal depending on element	O
Radiology report	LOINC	LOINC 18782-3 Radiology study observation	Text	O
Allergy	LOINC	LOINC 48765-2 Allergies and adverse reactions and document	Text/Code (SNOMED 703925004)	O
Presence of heart failure	LOINC	88020-3 Functional capacity NYHA	LOINC Answer List (LL4734-1, NYHA FA: Class I | Class II | Class III | Class IV)	O
Heart ejection fraction	LOINC	10230-1 left ventricular ejection fraction	Numeric (%)	O
Haemodialysis access construction date	LOINC		DateTime	O

[a] R = required; C = conditionally required; O = optional.

6.2 Fields of required data

6.2.1 General

The required data set is composed of information that helps clinicians making critical decisions and making sessions manageable by haemodialysis staff. For example, when medical staff are about to start a dialysis session, the data set helps them choose which machine to use and setting the parameters of a procedure. Infectious markers, such as HIV, HCV and HBV shall be included in the required data set. The information is important when a dialysis is prepared and during the initial stage of the procedure. When a patient is referred, all required data field shall be filled. The data elements are summarized in Table 2. The table covers terminology, data type and meta data.

6.2.2 Last dialysis date

6.2.2.1 Purpose

It is the date of most recent haemodialysis, which is required for deciding next dialysis schedule. Patients will suffer accumulation of toxins, fluid retention and electrolyte imbalance if they don't follow prescribed schedule of haemodialysis. Usually, patients are also instructed on the importance of haemodialysis schedule.

6.2.2.2 Business rules

a) This is a required element and shall be in the data set.

b) Unknown and known absence should be stated explicitly.

6.2.2.3 Missing data

This is a required element for conformance and shall not be empty; at least, a statement shall be given explaining the missing data.

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024

6.2.2.4 Example

“url”: “last-dialysis-date”

“performedDateTime”: “2022-08-22”

6.2.3 Haemodialysis interval

6.2.3.1 Purpose

It refers to currently prescribed interval between haemodialysis sessions. It provides information on when to start the next haemodialysis in receiving clinic.

6.2.3.2 Business Rules

a) This is a required element and shall be in the data set.

b) Unknown and known absence should be stated explicitly.

6.2.3.3 Missing Data

This is a required element for conformance and shall not be empty; at least, a statement shall be given explaining the missing data.

6.2.3.4 Example

“url”: “dialysis-interval”

“frequency”: 3

“periodUnit”: “wk”

6.2.4 Dry weight

6.2.4.1 Purpose

It helps clinicians determine how much excess fluid needs to be removed during each haemodialysis session, because it refers to optimal or target weigh that a patient should achieve during the treatment.

6.2.4.2 Business rules

a) This is a required element and shall be in the data set.

b) Unknown and known absence should be stated explicitly.

6.2.4.3 Missing data

This is a required element for conformance and shall not be empty.

6.2.4.4 Example

“url”: “dialysis-dryweight”

“value”: 47.5

“unit”: “kg”

© ISO 2024 – All rights reserved

Documents
Document Sharing Hub
Full document details: PD ISO/TS 9320:2024