



UL 61010-2-033

STANDARD FOR SAFETY

Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use – Part 2-033: Particular Requirements for Hand-Held Multimeters for Domestic and Professional Use, Capable of Measuring Mains Voltage

UL Standard for Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use – Part 2-033: Particular Requirements for Hand-Held Multimeters for Domestic and Professional Use, Capable of Measuring Mains Voltage, UL 61010-2-033

Second Edition, Dated January 10, 2020

Summary of Topics

This new edition of ANSI/UL 61010-2-033 dated January 10, 2020 is an Adoption of IEC 61010-2-033, Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use – Part 2-033: Particular Requirements for Hand-Held Multimeters for Domestic and Professional Use, Capable of Measuring Mains Voltage, UL 61010-2-033, (second edition, issued by IEC June 2019) as a new IEC-based UL standard with no US Differences.

The requirements are substantially in accordance with Proposal(s) on this subject dated August 30, 2019.

All rights reserved. No part of this publication may be reproduced, stored in a retrieval system, or transmitted in any form by any means, electronic, mechanical photocopying, recording, or otherwise without prior permission of UL.

UL provides this Standard "as is" without warranty of any kind, either expressed or implied, including but not limited to, the implied warranties of merchantability or fitness for any purpose.

In no event will UL be liable for any special, incidental, consequential, indirect or similar damages, including loss of profits, lost savings, loss of data, or any other damages arising out of the use of or the inability to use this Standard, even if UL or an authorized UL representative has been advised of the possibility of such damage. In no event shall UL's liability for any damage ever exceed the price paid for this Standard, regardless of the form of the claim.

Users of the electronic versions of UL's Standards for Safety agree to defend, indemnify, and hold UL harmless from and against any loss, expense, liability, damage, claim, or judgment (including reasonable attorney's fees) resulting from any error or deviation introduced while purchaser is storing an electronic Standard on the purchaser's computer system.

No Text on This Page

JANUARY 10, 2020



ANSI/UL 61010-2-033-2020

1

UL 61010-2-033

**Standard for Safety Requirements for Electrical Equipment for
Measurement, Control and Laboratory Use – Part 2-033: Particular
Requirements for Hand-Held Multimeters for Domestic and Professional
Use, Capable of Measuring Mains Voltage**

First Edition – August, 2014

Second Edition

January 10, 2020

This ANSI/UL Standard for Safety consists of the Second Edition.

The most recent designation of ANSI/UL 61010-2-033 as an American National Standard (ANSI) occurred on January 10, 2020. ANSI approval for a standard does not include the Cover Page, Transmittal Pages, Title Page, or Preface. The IEC Foreword is also excluded from the ANSI approval of IEC-based standards.

Comments or proposals for revisions on any part of the Standard may be submitted to UL at any time. Proposals should be submitted via a Proposal Request in UL's On-Line Collaborative Standards Development System (CSDS) at <https://csds.ul.com>.

UL's Standards for Safety are copyrighted by UL. Neither a printed nor electronic copy of a Standard should be altered in any way. All of UL's Standards and all copyrights, ownerships, and rights regarding those Standards shall remain the sole and exclusive property of UL.

COPYRIGHT © 2020 UNDERWRITERS LABORATORIES INC.

No Text on This Page

CONTENTS

PREFACE (UL)	7
FOREWORD	9
INTRODUCTION	13
1 Scope and object	15
1.1.1 Equipment included in scope	15
1.1.2 Equipment excluded from scope	15
1.2.1 Aspects included in scope	15
2 Normative references	15
3 Terms and definitions	16
3.5 Safety terms	16
4 Tests	16
4.3.2.5 MAINS supply	16
4.3.2.5 Power supply	16
4.3.2.6 Input and output voltages	17
5 Marking and documentation	17
5.1.2 Identification	17
5.1.5 TERMINALS, connections, and operating devices	17
5.2 Warning markings	18
5.4.1 General	19
6 Protection against electric shock	19
6.1 General	19
6.5.2 PROTECTIVE BONDING	21
6.5.2 Not used	21
6.5.5 Automatic disconnection of the supply	21
6.5.5 Note used	21
6.6 Connections to external circuits	21
6.7.1.3 CREEPAGE DISTANCES	22
6.7.1.5 Requirements for insulation according to type of circuit	22
6.8.3.1 The a.c. voltage test	23
6.9 Constructional requirements for protection against electric shock	23
7 Protection against mechanical HAZARDS	23
8 Resistance to mechanical stresses	23
9 Protection against the spread of fire	23
10 Equipment temperature limits and resistance to heat	24
11 Protection against HAZARDS from fluids and solid foreign objects	24
12 Protection against radiation, including laser sources, and against sonic and ultrasonic pressure	24
13 Protection against liberated gases and substances, explosion and implosion	24
14 Components and subassemblies	24
14.8 Circuits used to limit TRANSIENT OVERVOLTAGES	24
14.8 Not used	24
14.101 Probe assemblies and accessories	24
15 Protection by interlocks	24
16 HAZARDS resulting from application	24
17 RISK assessment	24
101 Measuring circuits	25
101.1 General	25
101.2 Current measuring circuits	25
101.3 Protection against mismatches of inputs and ranges	25
101.4 Protection against MAINS overvoltages	25



102	Indicating devices.....	29
102.1	General	29
102.2	Battery level.....	30
102.3	Over-range	30
102.4	Permanent overvoltages.....	31

Annexes

Annex K (normative) Insulation requirements not covered by 6.7

K.3	Insulation in circuits not addressed in 6.7, Clause K.1 or Clause K.2	33
K.3	Insulation for circuits not addressed in 6.7, Clauses K.1, K.2 or K.101	33
K.3.1	General.....	33
K.101	Insulation requirements for measuring circuits RATED for MEASUREMENT CATEGORIES III and IV	33
K.101.1	General	33
K.101.2	CLEARANCES.....	33
K.101.3	CREEPAGE DISTANCES	34
K.101.4	Solid insulation	34

Annex L (informative) Index of defined terms

Annex AA (normative) Measurement categories

AA.1	General	40
AA.2	MEASUREMENT CATEGORIES.....	40
AA.2.1	MEASUREMENT CATEGORY II	40
AA.2.2	MEASUREMENT CATEGORY III	40
AA.2.3	MEASUREMENT CATEGORY IV	40
AA.2.4	Measuring circuits without a MEASUREMENT CATEGORY RATING	41

Annex BB (informative) HAZARDS pertaining to measurements performed in certain environments

BB.1	General	42
BB.2	HAZARDS.....	42
BB.2.1	General	42
BB.2.2	Electric shock.....	42
BB.2.3	Arc blast	42
BB.2.4	Thermal burns.....	42
BB.3	MAINS.....	42
BB.4	Telecommunications networks	43
BB.5	Current measurements in inductive circuits.....	43
BB.6	Battery-driven circuits	43
BB.7	Measurements at higher frequencies.....	43

Annex CC (informative) 4-mm "banana" TERMINALS

CC.1	General	44
CC.2	Dimensions.....	44

Annex DD (informative) Flowchart for insulation according to the type of circuit

Bibliography

No Text on This Page

PREFACE (UL)

This UL Standard is based on IEC Publication 61010-2-033: 2.0 edition Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-033: Particular requirements for hand-held multimeters for domestic and professional use, capable of measuring mains voltage. IEC publication 61010-2-033 is copyrighted by the IEC.

This UL Standard 61010-2-033 Standard for Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use – Part 2-033: Particular Requirements for Hand-Held Multimeters and Other Meters, for Domestic and Professional Use, Capable of Measuring Mains Voltage, is to be used in conjunction with the latest edition of UL 61010-1. The requirements for hand-held multimeters for domestic and professional use, capable of measuring mains voltage are contained in this Part 2 Standard and UL 61010-1.

Requirements of this Part 2 Standard, where stated, amend the requirements of UL 61010-1.

Where a particular subclause of UL 61010-1 is not mentioned in UL 61010-2-033, the UL 61010-1 subclause applies.

These materials are subject to copyright claims of IEC and UL. No part of this publication may be reproduced in any form, including an electronic retrieval system, without the prior written permission of UL. All requests pertaining to the Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use – Part 2-033: Particular Requirements for Hand-Held Multimeters and Other Meters, for Domestic and Professional Use, Capable of Measuring Mains Voltage, 61010-2-033 Standard should be submitted to UL.

Note – Although the intended primary application of this Standard is stated in its Scope, it is important to note that it remains the responsibility of the users of the Standard to judge its suitability for their particular purpose.

No Text on This Page

FOREWORD

INTERNATIONAL ELECTROTECHNICAL COMMISSION

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE – Part 2-033: Particular requirements for hand-held multimeters for domestic and professional use, capable of measuring MAINS voltage

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61010-2-033 has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment.

This second edition cancels and replaces the first edition published in 2012. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) The scope has been reduced to hand-held multimeters. Voltmeters and clamp multimeters have been removed. They are addressed respectively by IEC 61010-2-030 and IEC 61010-2-032. The relevant definitions have been removed.

- b) Subclause 4.4.2.101 has been relocated into Clause 102.
- c) CLEARANCES and CREEPAGE DISTANCES for WET LOCATIONS and for measuring circuit TERMINALS exceeding 1 000 V a.c. or 1 414 V d.c. have been specified.
- d) Subclause 14.101 related to "Circuits or components used as TRANSIENT OVERVOLTAGE limiting devices in measuring circuits used to measure MAINS" has been removed.
- e) References to IEC 61010-031 for probe assemblies and IEC 61010-2-032 for current sensors have been added.
- f) Requirements for protection against MAINS overvoltage measuring circuits have been added.
- g) Clause 102 has been rewritten.
- h) Requirements for measuring circuits from 1 000 V to 3 000 V have been added.
- i) An informative Annex CC about dimensions of 4-mm banana TERMINALS has been added.
- j) A flowchart for insulation according to the type of circuit has been added in a new Annex DD.

The text of this standard is based on the following documents:

FDIS	Report on voting
66/692/FDIS	66/694/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all the parts in the IEC 61010 series, published under the general title *Safety requirements for electrical equipment for measurement, control, and laboratory use*, can be found on the IEC website.

This Part 2-033 is to be used in conjunction with the latest edition of IEC 61010-1. It was established on the basis of the third edition (2010) of IEC 61010-1 and its Amendment 1 (2016), hereinafter referred to as Part 1.

This Part 2-033 supplements or modifies the corresponding clauses in IEC 61010-1 so as to convert that publication into the IEC standard: *Particular requirements for hand-held multimeters for domestic and professional use, capable of measuring MAINS voltage*.

Where a particular subclause of Part 1 is not mentioned in this Part 2-033, that subclause applies as far as is reasonable. Where this Part 2-033 states "addition", "modification", "replacement", or "deletion", the relevant requirement, test specification or note in Part 1 should be adapted accordingly.

In this standard:

- a) the following print types are used:

- requirements: in roman type;
- NOTES: in small roman type;
- *conformity and tests: in italic type*;;
- terms used throughout this standard which have been defined in Clause 3: SMALL ROMAN CAPITALS;

b) subclauses, figures, tables and notes which are additional to those in Part 1 are numbered starting from 101. Additional annexes are lettered starting from AA and additional list items are lettered from aa).

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

No Text on This Page

INTRODUCTION

Part 2-030 specifies the safety requirements for equipment with testing and measuring circuits which are connected for test or measurement purposes to devices or circuits outside the measurement equipment itself. Requirements of Part 2-030 have been included in this Part 2-033. Equipment within the scopes of both Part 2-030 and Part 2-033 are considered to be covered by the requirements of this Part 2-033.

Part 2-032 specifies the safety requirements for hand-held and hand-manipulated current sensors. For equipment within the scope of Part 2-032 and Part 2-033, only Part 2-032 is applicable.

Part 2-034 specifies the safety requirements for measurement equipment for insulation resistance and test equipment for electric strength which are connected to units, lines or circuits for test or measurement purposes. For equipment within the scope of Part 2-033 and Part 2-034, only Part 2-034 is applicable.